


SharePoint

Darenhill Odessa - ?



KZN HEALTH

KZN Health Intranet

Search this site

[HOME](#) [CORPORATE INFORMATION](#) [COMPONENTS](#) [DIRECTORY](#) [DISTRICT OFFICES](#) [HEALTH FACILITIES](#)

KZN Health > Components > Supply Chain Management

AdvertQuote



Quotation Advert

Opening Date: 2022-08-12

Closing Date: 2022-08-19

Closing Time: 11:00

INSTITUTION DETAILS

Institution Name: St Andrews hospital

Province: KwaZulu-Natal

Department or Entity: Department of Health

Division or section: Central Supply Chain Management

Place where goods / services is required

Date Submitted 2022-08-12

ITEM CATEGORY AND DETAILS

Quotation Number: ZNQ:
SAH: 148/22/23

Item Category: Goods

Item Description: SURGICAL ITEMS

Quantity (if supplies)

COMPULSORY BRIEFING SESSION / SITE VISIT

Select Type: Not Applicable

Date :

Time:

Venue:

QUOTES CAN BE COLLECTED FROM: WEBSITE

QUOTES SHOULD BE DELIVERED TO: DEPOSIT IN THE TENDER BOX SITUATED IN THE BOTTOM MAIN GATE OR
EMAIL TO Standrews.Quotations@kznhealth.gov.za

ENQUIRIES REGARDING THE ADVERT MAY BE DIRECTED TO:


Name: MS B. MNOMIYA

Email: Standrews.Quotations@kznhealth.gov.za

Contact Number: 039-4331955

Finance Manager Name: MR N.I MDINGI

Finance Manager Signature:


No late quotes will be considered

ST ANDREWS HOSPITAL

Continuation Sheet-Surgical Item(s): SAH: 148/22/23

	Item Description	Qty	Unit Price	Total Price
1..	Dual thermometer	11 Units		
2.	Ambubag musk size 00	02 Units		
3.	Ambubag musk size 03	02 Units		
4.	Ambubag musk size 04	02Units		
5.	Disposable vaginal cusco large (Box /100)	20 Boxes		
6.	Disposable vaginal cusco small (Box/100)	10 Boxes		
7.	Disposable vaginal cusco medium(Box/100)	20 Boxes		
8.	Heavy duty gloves (long cuff) red	02 Pairs		
9.	Room thermometer	10 Units		
10.	Reservoir mask adult	06 Units		
11.	Reservoir mask child	06 Units		
12.	Reservoir mask neonate	06 Units		
13.	Laryngeal mask (size 3 or 4) adult size	02 Units		
14.	Manual BP machine with cuffs (sphygmomanometer)	14 Units		
15.	Outreach examination lamp portable – headlamp	02 Units		
16.	Peak flow meter	01 Unit		

Tenderer's Signature: _____

Date: _____

Confirmation

Compulsory to be completed by the bidder and returned with quotation:

I,, from	
<i>(Print Name)</i>	<i>(Name of Business)</i>
do hereby acknowledge that I have read and understand the specifications as laid out above and will ensure that the quotation price submitted will address all the requirements as stipulated.	
..... Signature / / Date

BIDDER'S DISCLOSURE

1. PURPOSE OF THE FORM

Any person (natural or juristic) may make an offer or offers in terms of this invitation to bid. In line with the principles of transparency, accountability, impartiality, and ethics as enshrined in the Constitution of the Republic of South Africa and further expressed in various pieces of legislation, it is required for the bidder to make this declaration in respect of the details required hereunder.

Where a person/s are listed in the Register for Tender Defaulters and / or the List of Restricted Suppliers, that person will automatically be disqualified from the bid process.

2. BIDDER'S DECLARATION

2.1. Is the bidder, or any of its directors / trustees / shareholders / members / partners or any person having a controlling interest¹ in the enterprise, employed by the state? YES/NO

2.1.1 If so, furnish particulars of the names, individual identity numbers, and, if applicable, state employee numbers of sole proprietor/ directors / trustees / shareholders / members/ partners or any person having a controlling interest in the enterprise, in table below.

Table with 3 columns: Full Name, Identity Number, Name of State Institution

2.2. Do you, or any person connected with the bidder, have a relationship with any person who is employed by the procuring institution? YES/NO

2.2.1. If so, furnish particulars:

2.3. Does the bidder or any of its directors / trustees / shareholders / members / partners or any person having a controlling interest in the enterprise have any interest in any other related enterprise whether or not they are bidding for this contract? YES/NO

2.3.1. If so, furnish particulars:

3. DECLARATION

I, the undersigned,(name)..... in submitting the accompanying bid, do hereby make the following statements that I certify to be true and complete in every respect:

- 3.1. I have read and I understand the contents of this disclosure;
3.2. I understand that the accompanying bid will be disqualified if this disclosure is found not to be true and complete in every respect;
3.3. The bidder has arrived at the accompanying bid independently from, and without consultation, communication, agreement or arrangement with any competitor. However, communication between partners in a joint venture or consortium² will not be construed as collusive bidding.
3.4. In addition, there have been no consultations, communications, agreements or arrangements with any competitor regarding the quality, quantity, specifications, prices, including methods, factors or formulas used to calculate prices, market allocation, the intention or decision to submit or not to submit the bid, bidding with the intention not to win the bid and conditions or delivery particulars of the products or services to which this bid invitation relates.
3.5. The terms of the accompanying bid have not been, and will not be, disclosed by the bidder, directly or indirectly, to any competitor, prior to the date and time of the official bid opening or of the awarding of the contract.
3.6. There have been no consultations, communications, agreements or arrangements made by the bidder with any official of the procuring institution in relation to this procurement process prior to and during the bidding process except to provide clarification on the bid submitted where so required by the institution; and the bidder was not involved in the drafting of the specifications or terms of reference for this bid.
3.7. I am aware that, in addition and without prejudice to any other remedy provided to combat any restrictive practices related to bids and contracts, bids that are suspicious will be reported to the Competition Commission for investigation and possible imposition of administrative penalties in terms of section 59 of the Competition Act No 89 of 1998 and or may be reported to the National Prosecuting Authority (NPA) for criminal investigation and or may be restricted from conducting business with the public sector for a period not exceeding ten (10) years in terms of the Prevention and Combating of Corrupt Activities Act No 12 of 2004 or any other applicable legislation.

I CERTIFY THAT THE INFORMATION FURNISHED IN PARAGRAPHS 1, 2 and 3 ABOVE IS CORRECT.
I ACCEPT THAT THE STATE MAY REJECT THE BID OR ACT AGAINST ME IN TERMS OF PARAGRAPH 6 OF PFMA SCM INSTRUCTION 03 OF 2021/22 ON PREVENTING AND COMBATING ABUSE IN THE SUPPLY CHAIN MANAGEMENT SYSTEM SHOULD THIS DECLARATION PROVE TO BE FALSE.

Name of Bidder Signature Position Date

1 the power, by one person or a group of persons holding the majority of the equity of an enterprise, alternatively, the person/s having the deciding vote or power to influence or to direct the course and decisions of the enterprise.

2 Joint venture or Consortium means an association of persons for the purpose of combining their expertise, property, capital, efforts, skill and knowledge in an activity for the execution of a contract.

SPECIAL CONTRACT CONDITIONS OF QUOTATIONS

1. AMENDMENT OF CONTRACT

- 1.1. Any amendment to or renunciation of the provisions of the contract shall at all times be done in writing and shall be signed by both parties.

2. CHANGE OF ADDRESS

- 2.1. Bidders must advise the Department of Health (institution where the offer was submitted) should their address (*domicilium citandi et executandi*) details change from the time of bidding to the expiry of the contract.

3. GENERAL CONDITIONS ATTACHED TO THIS QUOTATION

- 3.1. The Department is under no obligation to accept the lowest or any quote.
- 3.2. The Department reserves the right to communicate in writing with vendors in cases where information is incomplete or where there are obscurities regarding technical aspects of the offer, to obtain confirmation of prices or preference claims in cases where it is evident that a typing, written, transfer or unit error has been made, to investigate the vendor's standing and ability to complete the supply/service satisfactorily.
- 3.3. **ALL DECISIONS TAKEN BY THE DEPARTMENT ARE FINAL, INCLUDING THE AWARD OR CANCELLATION OF THIS QUOTATION.**
- 3.4. The price quoted must include VAT (if VAT vendor).
- 3.5. Should a bidder become a VAT vendor after award or during the implementation of a contract, they may not request the VAT percentage from the Department as the service provider made an offer during the period they were not registered as a VAT vendor. The Department is only liable for any VAT from registered VAT vendors as originally stated on the quotation document.
- 3.6. The bidder must ensure the correctness & validity of the quotation:
- (i) *that the price(s), rate(s) & preference quoted cover all for the work/item (s) & accept that any mistakes regarding the price (s) & calculations will be at the bidder's risk*
- (ii) *it is the responsibility of the bidder to confirm receipt of their quotation and to keep proof thereof.*
- 3.7. The bidder must accept full responsibility for the proper execution & fulfilment of all obligations conditions devolving on under this agreement, as the Principal (s) liable for the due fulfilment of this contract.
- 3.8. This quotation will be evaluated based on the 80/20 points system, specification, correctness of information and/or functionality criteria. All required documentation must be completed in full and submitted.
- 3.9. Offers must comply strictly with the specification.
- 3.10. Only offers that meet or are greater than the specification will be considered.
- 3.11. Late offers will not be considered.
- 3.12. Expired product/s will not be accepted. All products supplied must be valid for a minimum period of six months.
- 3.13. Used/ second-hand products will not be accepted.
- 3.14. A bidder not registered on the Central Suppliers Database or whose verification has failed will not be considered.
- 3.15. All delivery costs must be included in the quoted price for delivery at the prescribed destination.
- 3.16. Only firm prices will be accepted. Such prices must remain firm for the contract period. Non-firm prices (including rates of exchange variations) will not be considered.
- 3.17. In cases where different delivery points influence the pricing, a separate pricing schedule must be submitted for each delivery point.
- 3.18. In the event of a bidder having multiple quotes, only the cheapest according to specification will be considered.
- 3.19. Verification will be conducted to identify if bidders have multiple companies and are cover-quoting for this bid.
- 3.20. In such instances, the Department reserves the right to immediately disqualify such bidders as cover-quoting is an offence that represents both corruption and acquisition fraud.

4. SPECIAL INSTRUCTIONS AND NOTICES TO BIDDERS REGARDING THE COMPLETION OF THIS QUOTATION.

- 4.1. Unless inconsistent with or expressly indicated otherwise by the context, the singular shall include the plural and vice versa and with words importing the masculine gender shall include the feminine and the neuter.
- 4.2. Under no circumstances whatsoever may the quotation/bid forms be retyped or redrafted. Photocopies of the original bid documentation may be used, but an original signature must appear on such photocopies.
- 4.3. The bidder is advised to check the number of pages and to satisfy himself that none are missing or duplicated.
- 4.4. Quotations submitted must be complete in all respects. However, where it is identified that information in a bidder's response, which does not affect the preference points or price, is incomplete in any respect, the said supplier meets all specification requirements and scores the highest points in terms of preference points and price, the Department reserves the right to request the bidder to complete/submit such information.
- 4.5. Any alteration made by the bidder must be initialled; failure to do so may render the response invalid.
- 4.6. Use of correcting fluid is prohibited and may render the response invalid.
- 4.7. Quotations will be opened in public as soon as practicable after the closing time of quotation.
- 4.8. Where practical, prices are made public at the time of opening quotations.
- 4.9. If it is desired to make more than one offer against any individual item, such offers should be given on a photocopy of the page in question. Clear indication thereof must be stated on the schedules attached.

4.10. The Department is under no obligation to pay suppliers in part for work done if the supplier can no longer fulfil their obligation.

5. SPECIAL INSTRUCTIONS REGARDING HAND DELIVERED QUOTATIONS

- 5.1. Quotation shall be lodged at the address indicated not later than the closing time specified for their receipt, and in accordance with the directives in the quotation documents.
- 5.2. Each quotation shall be addressed in accordance with the directives in the quotation documents and shall be lodged in a separate sealed envelope, with the name and address of the bidder, the quotation number and closing date indicated on the envelope. The envelope shall not contain documents relating to any quotation other than that shown on the envelope. If this provision is not complied with, such quotations/bids may be rejected as being invalid.
- 5.3. All quotations received in sealed envelopes with the relevant quotation numbers on the envelopes are kept unopened in safe custody until the closing time of the quotation/bids. Where, however, a quotation is received open, it shall be sealed. If it is received without a quotation/bid number on the envelope, it shall be opened, the quotation number ascertained, the envelope sealed and the quotation number written on the envelope.
- 5.4. A specific box is provided for the receipt of quotations, and no quotation found in any other box or elsewhere subsequent to the closing date and time of quotation will be considered.
- 5.5. No quotation/bid sent through the post will be considered if it is received after the closing date and time stipulated in the quotation documentation, and proof of posting will not be accepted as proof of delivery.
- 5.6. Quotation documents must not be included in packages containing samples. Such quotations may be rejected as being invalid.

6. SAMPLES

- 6.1. In the case of the quote document stipulating that samples are required, the supplier will be informed in due course when samples should be provided to the institution. (This decreases the time of safety and storage risk that may be incurred by the respective institution). The bidders sample will be retained if such bidder wins the contract.
 - (i) If a company/s who has not won the quote requires their samples, they must advise the institution in writing of such.
 - (ii) If samples are not collected within three months of close of quote the institution reserves the right to dispose of them at their discretion.
- 6.2. **Samples must be made available when requested in writing or if stipulated on the document.**
 - (i) If a Bidder fails to provide a sample of their product on offer for scrutiny against the set specification when requested, their offer will be rejected. All testing will be for the account of the bidder.

7. COMPULSORY SITE INSPECTION / BRIEFING SESSION

- 7.1. Bidders who fail to attend the compulsory meeting will be disqualified from the evaluation process.
 - (i) The institution has determined that a compulsory site meeting take place.
 - (ii) Date ____/____/____ Time ____:____ Place _____

Institution Stamp:	Institution Site Inspection / briefing session Official Full Name: Signature: Date:
--------------------	--

8. STATEMENT OF SUPPLIES AND SERVICES

8.1. The contractor shall, when requested to do so, furnish particulars of supplies delivered or services executed. If he/she fails to do so, the Department may, without prejudice to any other rights which it may have, institute inquiries at the expense of the contractor to obtain the required particulars.

9. SUBMISSION AND COMPLETION OF SBD 6.1

9.1. Should a bidder wish to qualify for preference points they must complete a SBD 6.1 document. Failure by a bidder to provide all relevant information required, will result in such a bidder not being considered for preference point's allocation. The preferences applicable on the closing date will be utilized. Any changes after the closing date will not be considered for that particular quote.

10. TAX COMPLIANCE REQUIREMENTS

- 10.1. In the event that the tax compliance status has failed on CSD, **it is the suppliers' responsibility to provide a SARS pin in order for the institution to validate the tax compliance status of the supplier.**
- 10.2. In the event that the institution cannot validate the suppliers' tax clearance on SARS as well as the Central Suppliers Database, **the quote will not be considered and passed over as non-compliant according to National Treasury Instruction Note 4 (a) 2016/17.**

TAX INVOICE

10.3. A tax invoice shall be in the currency of the Republic of South Africa and shall contain the following particulars:

- | | |
|--|--|
| (i) the name, address and registration number of the supplier; | (iv) a description and quantity or volume of the goods or services supplied; |
| (ii) the name and address of the recipient; | (v) the official department order number issued to the supplier; |
| (iii) an individual serialized number and the date upon which the tax invoice is issued; | (vi) the value of the supply, the amount of tax charged; |
| | (vii) the words tax invoice in a prominent place. |

11. PATENT RIGHTS

The supplier shall indemnify the **KZN Department of Health** (hereafter known as the purchaser) against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the goods or any part thereof by the purchaser.

12. PENALTIES

- 12.1. If at any time during the contract period, the service provider is unable to perform in a timely manner, the service provider must notify the institution in writing/email of the cause of and the duration of the delay. Upon receipt of the notification, the institution should evaluate the circumstances and, if deemed necessary, the institution may extend the service provider's time for performance.
- 12.2. In the event of delayed performance that extends beyond the delivery period, the institution is entitled to purchase commodities of a similar quantity and quality as a substitution for the outstanding commodities, without terminating the contract, as well as return commodities delivered at a later stage at the service provider's expense.
- 12.3. Alternatively, the institution may elect to terminate the contract and procure the necessary commodities in order to complete the contract. In the event that the contract is terminated the institution may claim damages from the service provider in the form of a penalty. The service provider's performance should be captured on the service provider database in order to determine whether or not the service provider should be awarded any contracts in the future.
- 12.4. If the supplier fails to deliver any or all of the goods or to perform the services within the period(s) specified in the contract, the purchaser shall, without prejudice to its other remedies under the contract, deduct from the contract price, as a penalty, a sum calculated on the delivered price of the delayed goods or unperformed services using the current prime interest rate calculated for each day of the delay until actual delivery or performance.

13. TERMINATION FOR DEFAULT

- 13.1. The purchaser, without prejudice to any other remedy for breach of contract, by written notice of default sent to the supplier, may terminate this contract in whole or in part:
 - (i) if the supplier fails to deliver any or all of the goods within the period(s) specified in the contract,
 - (ii) if the supplier fails to perform any other obligation(s) under the contract; or
 - (iii) if the supplier, in the judgment of the purchaser, has engaged in corrupt or fraudulent practices in competing for or in executing the contract.
- 13.2. In the event the purchaser terminates the contract in whole or in part, the purchaser may procure, upon such terms and in such manner as it deems appropriate, goods, works or services similar to those undelivered, and the supplier shall be liable to the purchaser for any excess costs for such similar goods, works or services.
- 13.3. Where the purchaser terminates the contract in whole or in part, the purchaser may decide to impose a restriction penalty on the supplier by prohibiting such supplier from doing business with the public sector for a period not exceeding 10 years.

14. THE DEPARTMENT RESERVES THE RIGHT TO PASS OVER ANY QUOTATION WHICH FAILS TO COMPLY WITH THE ABOVE.

3. POINTS AWARDED FOR PRICE

3.1 THE 80/20 PREFERENCE POINT SYSTEMS

A maximum of 80 points is allocated for price on the following basis:

$$P_s = 80 \left(1 - \frac{P_t - P_{\min}}{P_{\min}} \right) \text{ Where}$$

- P_s = Points scored for price of bid under consideration
- P_t = Price of bid under consideration
- P_{min} = price of lowest acceptable bid

4. POINTS AWARDED FOR B-BBEE STATUS LEVEL OF CONTRIBUTOR

4.1 In terms of Regulation 6 (2) and 7 (2) of the Preferential Procurement Regulations, preference points must be awarded to a bidder for attaining the B-BBEE status level of contribution in accordance with the table below:

B-BBEE Status Level of Contributor	Number of points (80/20 system)
1	20
2	18
3	14
4	12
5	8
6	6
7	4
8	2
Non-compliant contributor	0

5. BID DECLARATION

5.1 Bidders who claim points in respect of B-BBEE Status Level of Contribution must complete the following:

6. B-BBEE STATUS LEVEL OF CONTRIBUTOR CLAIMED IN TERMS OF PARAGRAPHS 1.4 AND 4.1

6.1 B-BBEE Status Level of Contributor: =(maximum of 20 points)

(Points claimed in respect of paragraph 7.1 must be in accordance with the table reflected in paragraph 4.1 and must be substantiated by relevant proof of B-BBEE status level of contributor.

7. SUB-CONTRACTING applicable box)

(Tick

YES		NO	
-----	--	----	--

7.1 Will any portion of the contract be sub-contracted?

7.1.1 If yes, indicate:

- i) What percentage of the contract will be subcontracted.....%
- ii) The name of the sub-contractor.....
- iii) The B-BBEE status level of the sub-contractor.....

8. Whether the sub-contractor is an EME or QSE

(Tick applicable box)

iv) Specify, by ticking the appropriate box, if subcontracting with an enterprise in terms of Preferential Procurement Regulations,2017:

YES		NO	
-----	--	----	--

Designated Group: An EME or QSE which is at last 51% owned by:	EME √	QSE √
Black people		
Black people who are youth		
Black people who are women		
Black people with disabilities		
Black people living in rural or underdeveloped areas or townships		
Cooperative owned by black people		
Black people who are military veterans		
OR		
Any EME		
Any QSE		

9. DECLARATION WITH REGARD TO COMPANY/FIRM

9.1 Name of company/firm:.....

9.2 VAT registration number:.....

9.3 Company registration number:.....

9.4 TYPE OF COMPANY/ FIRM [TICK APPLICABLE BOX]

- Partnership/Joint Venture / Consortium
- One person business/sole propriety
- Close corporation
- Company
- (Pty) Limited

9.5 DESCRIBE PRINCIPAL BUSINESS ACTIVITIES

.....
.....

9.6 COMPANY CLASSIFICATION [TICK APPLICABLE BOX]

- Manufacturer
- Supplier
- Professional service provider
- Other service providers, e.g. transporter, etc.

9.7 Total number of years the company/firm has been in business:.....

9.8 I/we, the undersigned, who is / are duly authorised to do so on behalf of the company/firm, certify that the points claimed, based on the B-BBE status level of contributor indicated in paragraphs 1.4 and 6.1 of the foregoing certificate, qualifies the company/ firm for the preference(s) shown and I / we acknowledge that:

- i) The information furnished is true and correct;
- ii) The preference points claimed are in accordance with the General Conditions as indicated in paragraph 1 of this form;
- iii) In the event of a contract being awarded as a result of points claimed as shown in paragraphs 1.4 and 6.1, the contractor may be required to furnish documentary proof to the satisfaction of the purchaser that the claims are correct;
- iv) If the B-BBEE status level of contributor has been claimed or obtained on a fraudulent basis or any of the conditions of contract have not been fulfilled, the purchaser may, in addition to any other remedy it may have –
 - (a) disqualify the person from the bidding process;
 - (b) recover costs, losses or damages it has incurred or suffered as a result of that person's conduct;
 - (c) cancel the contract and claim any damages which it has suffered as a result of having to make less favourable arrangements due to such cancellation;
 - (d) recommend that the bidder or contractor, its shareholders and directors, or only the shareholders and directors who acted on a fraudulent basis, be restricted by the National Treasury from obtaining business from any organ of state for a period not exceeding 10 years, after the *audi alteram partem* (hear the other side) rule has been applied; and
 - (e) forward the matter for criminal prosecution.

<p>WITNESSES</p> <p>1.</p> <p>2.</p>
--

<p>.....</p> <p>SIGNATURE(S) OF BIDDERS(S)</p> <p>DATE:</p> <p>ADDRESS.....</p> <p>.....</p> <p>.....</p>

SPECIFICATION

SAH: 121/22/23

1. MASK, OXYGEN, REBREATHING, UNDER THE CHIN TYPE, SOFT STRONG VINYL BREATHING BAG, FACE PIECE OF SOFT ODOURLESS POLYVINYL CHORIDE, TUBING ATTACHED, MINIMUM LENGTH 2M, PAEDIATRIC SIZE.


(NON – REBREATHING OR PARTIAL REBREATHING : MUST HAVE A VALVE TO PREVENT EXPIRED AIR FROM ENTRERING THE BAG.)

2. MASK, OXYGEN, REBREATHING, UNDER THE CHIN TYPE, SOFT STRONG VINYL BREATHING BAG, FACE PIECE OF SOFT ODOURLESS POLYVINYL CHORIDE, TUBING ATTACHED, MINIMUM LENGTH 2M, ADULT SIZE.

(NON – REBREATHING OR PARTIAL REBREATHING : MUST HAVE A VALVE TO PREVENT EXPIRED AIR FROM ENTRERING THE BAG.)

Examination light: headlamp

Catalogue Number	Item Name	Available on Transversal Contract	Contract Number or Code
PHC-E-098	Examination light: headlamp	No	


Description	<p>A lightweight battery-operated light, worn on the forehead.</p>	
Use	<p>Useful for hands-free illumination of an area of interest, during a clinical examination or simple procedure, or when mains power is unavailable.</p>	
Clinical or User Specifications	<p>Note: this is not a surgical headlamp – the light output is significantly lower.</p> <p>The lightweight lamp is mounted on an adjustable headband. The lamp is adjustable for tilt.</p> <p>Easy-access on-off switch on the lamp unit.</p> <p>Easy-access focusing ring on the lamp, to narrow or widen the beam of light.</p> <p>Belt-worn rechargeable Lithium-ion battery pack.</p>	
Accessories or Related Consumables	<p>Charger, charge cable.</p>	
Care	<p>Washable with standard detergents – do not immerse.</p> <p>Keep charged when not in use.</p> <p>Caution: This type of light maintains a constant light output until the battery level is drained to a certain point, then switches off suddenly to protect the battery. The light does not dim before switching off. Be aware of this possibility when item is being used as the sole source of light.</p>	
Supplier Warranty	<p>1 year (excluding deliberate or accidental damage).</p>	

ST ANDREWS HOSPITAL
SPECIFICATION MEETING

2022-08-08

APPROVED/NOT APPROVED:.....

CHAIRPERSON: *A. J. Bell*




- Should be chemical resistant
- Cover whole arm to the elbow
- Made out of rubber
- SABS approved
- Withstand heat up to 95 degrees
- Anti-slip grip
- Needle / any sharp object resistant
- Should have iso 9001 certification

Diagnostic

Blood pressure measurement: aneroid type

TECHNICAL SPECIFICATIONS
page
542

Catalogue Number	Item Name	Available on Transversal Contract	Contract Number or Code
PHC-E-115	Sphygmomanometer: aneroid	No	

Description	A unit which measures blood pressure.	
Use	Suitable for blood pressure measurement by a clinician, using the auscultatory method. This is the preferable type if the patient has a disturbance of heart rhythm, e.g. Atrial fibrillation.	
Clinical or User Specifications	The user inflates and deflates an upper-arm cuff, by squeezing a bulb and closing or releasing the exit valve. Blood pressure is displayed on a large, clear, high-contrast analogue gauge, with a dial graduated in mm.Hg. Not dependent on an external power source. The unit can be mounted to a desktop, or to an adjacent wall or pole with special brackets.	
Accessories or Related Consumables	Large adult, adult and paediatric upper-arm cuffs. Special mounting brackets must be ordered separately, if required.	
Care	Clinicians must be adequately trained on the correct identification of the Korotkoff sounds. The cuffs (outer cover and inflatable bladder) can be washed. They must be replaced if worn or damaged, or if the Velcro is no longer adhering.	
Maintenance	The unit must be calibrated at least annually, by technical personnel of the department or the supplier.	
Supplier Warranty	2 years on cuffs and covers (including bladders, tubes, outer covers, inflation bulbs, air inlet valves, release valves, filters, and cuff-fastening mechanism – velcro or other). Instrument itself guaranteed for 10 years, including an annual calibration guarantee. Guarantee excludes deliberate or accidental damage.	

REVISED: 15/07/2016

PROVINCE OF KWAZULU-NATAL

DEPARTMENT OF HEALTH

**HEALTH TECHNOLOGY SERVICES
(H.T.S)**

SPECIFICATION FOR:

UMDNS: 16156

**SPHYGMOMANOMETERS ANEROID – UNIVERSAL DESK MOUNT AND
WALL MOUNT**

SPECIFICATION: H.T.S. NO. S79 (SURGICAL)

NB: GENERAL CLAUSES THAT DO NOT APPLY TO THE EQUIPMENT OFFERED MUST BE ANSWERED "COMPLIES", "DOES NOT COMPLY" OR ANSWER THE QUESTION UNDER BIDDERS COMMENTS.

NO	SPECIFICATION	BIDDERS COMMENTS: STATE "COMPLIES" OR "DOES NOT COMPLY" OR ANSWER THE QUESTION.
Clause G1.1	<p>The space provided under "Bidder's Comments" for each clause must be used for this purpose. Bidders who neglect to provide answers to every Clause in this Bid Specification will be disqualified. Bidders must note that abbreviated answers e.g. N/A etc. will not be accepted.</p> <p>Bidders must also note that no part of any clause/s in this Bid Specification may be altered. Where there are traces of alterations found to any clauses in this Bid Specification during Adjudication, the Adjudication Committee will reserve the right to disqualify the bidder.</p> <p>The Bidder must clearly indicate if their offered product complies with the stated requirements, by indicating, "Complies" or "Does not comply" or answer the question next to the corresponding clause.</p>	
Clause G2	All responses must be clear and legible.	
Clause G3	GUARANTEE:	
Clause G3.1	<p>All Equipment, Materials and Workmanship provided under this Contract must be Guaranteed for a minimum period of twenty four (24) Months. The successful bidder must arrange with the respective Hospital / Institution and the Health Technology Services before Commissioning the Equipment at the respective Hospital / Institution.</p> <p>The bidder to note that the Guarantee period must only take effect upon successful Commissioning at the respective Hospital / Institution and successful test and acceptance by the Health Technology Services.</p>	
Clause G3.2	State percentage guaranteed up time of machine (Should be at least 99%).	
Clause G3.3	The recommended number of services, per annum, by the manufacturer, must be included during and up until the end of the guarantee period and all costs related to the provision of such service/s will be for the bidders account.	
Clause G3.4	The bidder must state the number of services that will be provided during and up to the end of the guarantee period.	
Clause G3.5	Any breakdown during the guarantee period must include all cost (spares, labour, travelling and sundries) for any prescribed maintenance services (major and minor) as well as any QA testing that is required by Department Health's Radiation Control Board during the guarantee period.	
Clause G3.6	Travelling and Travelling Time costs must be included during the Guarantee Period?	
Clause G3.7	Spares that may be required during the Guarantee Period will be supplied at the expense of the bidder.	
Clause G3.8	Downtime during the Guarantee Period must extend the Guarantee time on a Day-to-Day basis.	
Clause G3.9	Any repetition (twice or more) of the same type of fault that first occurred during the guarantee period must be considered as a repair under guarantee if it occurs within the first year after the expiry of the guarantee period.	

NO	SPECIFICATION	BIDDERS COMMENTS: STATE "COMPLIES" OR "DOES NOT COMPLY" OR ANSWER THE QUESTION.
Clause G3.10	The same guarantee conditions must apply to replacement units.	
Clause G4	The successful bidder must Supply, Deliver, Commission and install the Equipment and will be required to demonstrate the product to the applicable Staff at the Institution and costs for the abovementioned must be included in the final bid price.	
Clause G5	Bidders must offer the Health Technology Service's In House Technicians a demonstration of the product, which will enable the Health Technology Service's In House Technicians to become acquainted with the equipment during the Test and Acceptance phase.	
Clause G6	Preference may be given to a make and model that has been technically and clinically evaluated by a Government Institution within the R.S.A. (Attach proof of evaluation where applicable).	
Clause G7	The successful bidder must provide the Health Technology Service's in house Technicians, full training in the calibration, maintenance, service and repair of the product down to PCB Level. N.B. The quality and level of the training must be equivalent to the manufacturer's original factory training and any costs incurred to provide this training will be for the bidders account. A Certificate of Competency must be issued on completion of the training. The Training must be provided by the successful bidder to the Health Technology Services within three months from date of initial supply and delivery of the equipment to the end user.	
Clause G8	SERVICING:	
Clause G8.1	The bidder must have a well established service and repair facility in KwaZulu-Natal, to service, repair and calibrate the equipment offered. (The Health Technology Services reserves the right to inspect the premises).	
Clause G8.2	If the service is subcontracted to a local service agent, a signed copy of the letter of appointment by the bidder and acceptance by the subcontractor must be submitted with this bid / quotation. (The Health Technology Services reserves the right to inspect the premises).	
Clause G8.3	State Number of other medical equipment "Repair & Service" Agencies (excluding your Agency) represented by the subcontractor.	
Clause G8.4	<p>Supply the Name, Address and Telephone Number/s of the Local Service Department within KwaZulu-Natal.</p> <p>Please supply details as follows:</p> <p>Company name : _____</p> <p>Physical Address : _____</p> <p>_____</p> <p>Telephone Number/s : _____</p> <p>Fax number : _____</p> <p>_____</p> <p><i>(The Health Technology Services reserves the right to inspect the</i></p>	

NO	SPECIFICATION	BIDDERS COMMENTS: STATE "COMPLIES" OR "DOES NOT COMPLY" OR ANSWER THE QUESTION.
	<i>premises).</i>	
Clause G8.5	State if the Technician(s) are in the direct employ of the bidder or a subcontractor.	
Clause G8.6	The bidder must supply information on the number of Technicians permanently working in KwaZulu-Natal and their names and contact Telephone Number/s must be listed (Directly employed or subcontracted) in an annexure to the bid document.	
Clause G8.7	The Technician(s) must be original equipment manufacturer trained to deal with the service, repair and calibration of the equipment quoted on. N.B. Proof of original equipment manufacturer training must be submitted with this bid / quotation offer.	
Clause G8.8	The Institution's requirement is that a technician is available within a reasonable time (24 hours) to attend to malfunctioning equipment. The Bidder to state the technician per install base e.g. equipment ratio to technician ratio, e.g. 1 technician per 10 pieces of equipment.	
Clause G9	The bidder must Guarantee that no additional equipment will be required for the successful operation of the equipment bided for on delivery and commissioning at the customers site. A starter pack of all essential accessories and disposables must be supplied so that the unit can be put into immediate operation. The cost of the starter pack must be included in the final bid price.	
Clause G10	Optional accessories must be offered for separately on the Schedule of optional accessories found at the end of this Technical specification, indicating catalogue numbers, correct descriptions and Prices inclusive of V.A.T.	
Clause G11	Bidder must state the period of time for delivery of Spare parts following the receipt of an official order as follows: 0 to 10 days; 0 to 20 days; 0 to 30 days; 0 to 60 days; 0 to 90 days; more than 90 days.	
Clause G11.1	The Bidder must supply with this offer a list together with the quantities of spares held locally in stock in the KwaZulu-Natal Province on the offered product. The Health Technology Services reserves the right to inspect the premises to verify the spares stock held.	
Clause G12	The bidder must include a firm commitment in writing, which must be attached with this bid that they would supply spares, components, upgrades, complete original service / repair manual, technical support and ongoing training support for technical staff of the Health Technology Services and the end users Department of Health, KwaZulu-Natal throughout the life cycle of the equipment offered.	
Clause G13	Spares must be available for 10 (Ten) years from the original equipment manufacturer for the product offered.	
Clause G14	The successful bidder must include in their offer at no extra cost to the final bid price:	

NO	SPECIFICATION	BIDDERS COMMENTS: STATE "COMPLIES" OR "DOES NOT COMPLY" OR ANSWER THE QUESTION.
Clause G14.1	Complete user Operation / Maintenance Manual x 2 (two) Book / File; CD; DVD copies in English Language.	
Clause G14.2	Complete ORIGINAL Service / Repair Manual x 2 (two) Book / File; CD; DVD copies in English Language which MUST include the following information: Fault Finding Guide, Circuit Diagrams / Schematics, Circuit Descriptions, and PCB Layouts, Calibration Guide, Part Numbers and exploded diagram of Mechanical Parts / Panels.	
Clause G14.3	All the above Manuals must be properly bound in either a Book, File or CD form.	
Clause G14.4	The Bidder must supply all software (including software-keys and / or passwords) to allow for trouble shooting (faultfinding), maintenance, calibrations, repairs and services at no additional cost.	
Clause G15	Does your Company have an after hour service back up facility.	
Clause G16	If the equipment is taken away for repairs, a loan set must be made available on request to the end user by the Institution until the Institution's unit is returned. All costs incurred for providing the loan unit must be for the bidders account.	
Clause G17	Bidder must bid on the latest model and Technology that fully complies with this Technical Specification.	
Clause G17.1	The Bidder must state how long this technology has been commercially available (state when the model offered was launched).	
Clause G17.2	The bidder must state if there are any near future updates expected.	
Clause G18	The successful bidder must maintain a system for notifying and providing users with Updates, Modifications, new Software Releases and Recalls.	
Clause G19	The successful bidders must arrange for an acceptance test of the equipment with the Manager of the Health Technology Services and the Hospital Manager. A copy of the original answered Specification, copy of the invoice order and relevant paperwork (PH form) from the receiving Hospital must be submitted with the equipment when the ACCEPTANCE TEST is to be undertaken.	
Clause G20	Where equipment bided for, operates off 220 Volt, 50Hz a.c. supply, bidder must ensure that the product being quoted for is fitted with a 15 Amp approved mains plug top, which is held together by two screws.	
Clause G21	The unit must comply with an acceptable International Electrical Safety Standard such as IEC 60601-1 and 60601-1-2 for Medical Equipment where the quoted equipment operates off an electrical supply.	
Clause G22	All equipment, the installation and any alteration / additions must comply with:	
Clause G22.1	The Occupational Health and Safety Act (1993);	
Clause G22.2	The wiring code S.A.N.S. 0142.	
Clause G23	Units being quoted for must be CE Certified. (Attach a copy of certification). The make and the model offered must be reflected on the certificate.	

NO	SPECIFICATION	BIDDERS COMMENTS: STATE "COMPLIES" OR "DOES NOT COMPLY" OR ANSWER THE QUESTION.
Clause G24	The Mains Cable of the unit being quoted for must be the Hospital Grade Type and it must be a minimum length of (3) three metres. N.B. The mains cable of the unit being quoted for must be S.A.N.S. colour coded.	
Clause G25	The equipment being quoted for must be protected against Electro magnetic Interference.	
Clause G26	Only new equipment must be quoted for. Refurbished and reconditioned equipment being quoted on will not be accepted.	
Clause G27	Bidders must note that dedicated test equipment, spare parts and any special tooling required for the upkeep and maintenance of the equipment quoted on must be available to the Health Technology Services to procure if requested.	
Clause G28	All the necessary calibration and maintenance software, where applicable, required to maintain and calibrate the equipment, must be supplied with the equipment to the Health Technology Services at no extra cost to the final bid price.	
Clause G29	NB. HAZARDOUS SUBSTANCE ACT:	
Clause G29.1	If this type of equipment / apparatus appears on the schedule of Hazardous Substances issued by the Directorate: Health Technology of the Department of Health, a license in terms of the Act on Hazardous Substances (Act. 15/1973) must be submitted with this bid document. The license must be registered under the bidders name or a letter of joint venture must be submitted by the license holder where the license is not in the name of the bidder. Bidders that neglect to submit a license will not be considered.	
Clause G29.2	Bidder must state the Radiation Control licence number of the make and model of equipment offered.	License No:
Clause G29.3	Where it has been established by the bidder that the equipment offered does not require Radiation Control licence, proof from the Radiation Control authority must be submitted with this bid document.	
Clause G30	The system offered must comply fully with or exceed all of the minimum specification requirements per the Technical Clauses.	
Clause G31	The offer submitted must be supported by descriptive literature, colour pamphlets, colour brochures and technical data sheets applicable to the offer (i.e. supporting information for all components of the system) must accompany the bid, failing which the bid will not be considered.	
Clause G32	The equipment and any accessories ordered from the successful bidder will be delivered, installed, tested, calibrated, demonstrated (including specified training) and commissioned in the specific Hospital at the expense of the successful Bidder, prior to full payment being made.	
Clause G33	All prices are to include V.A.T. and must be quoted in the South African currency. The price must be valid for a period of 180 days from closing date of bid.	
Clause G34	If the product offered is unknown to the Department, the Department reserves the right to have the unit evaluated by a team of Technical and Clinical <u>experts</u> with regards to its functionality, performance and quality. The decision of this committee will be used as a motivation for the evaluation and recommendation of the bid. For this reason a demonstration unit must be readily available, or the bidder must take arrange for demonstration with	

NO	SPECIFICATION	BIDDERS COMMENTS: STATE "COMPLIES" OR "DOES NOT COMPLY" OR ANSWER THE QUESTION.
	representatives of the Department for the equipment offered at a site within South Africa where a same make and model of unit is installed and is in full clinical operation. The cost of this site visit is for the account of the bidder and it must therefore not place any obligation on the Department to procure from the bidder.	
Clause G35	The Institution requesting the unit reserves the right to clinically trial and evaluate the unit in order to ensure that the unit meets the clinical requirements of the Department before adjudication of the bid.	
Clause G36	UPGRADEABILITY WHERE APPLICABLE:	
Clause G36.1	Bidders are to state the policy with regard to future software updates and the costs that will be involved.	
Clause G36.2	The Bidder to state what hardware and software will be available, with costs and projected dates.	
Clause G37	UPGRADE POLICY:	
Clause G37.1	All future upgrades (hardware and software) involving <u>patient safety</u> must be offered at no additional cost.	
Clause G37.2	All future upgrades removing software viruses from existing software must be supplied at no cost.	
Clause G37.3	Any upgrade before or after installation of the equipment involving additional cost must be brought to the attention of the Manager, Health Technology Services.	
Clause G38	The Bidder must indicate the expected life of their offered unit and software in years.	

TECHNICAL SPECIFICATION.

Clause T1

The sphygmomanometer offered must be a universal type of aneroid sphygmomanometer, which must be suitable for use, both as a desk mount and wall mount type.

BIDDER'S COMMENTS:

Clause T1.1

The unit offered must have both a calibration and manufacturing warranty of a lifetime and a written confirmation from the manufacturer must be submitted with this bid / quotation offer.

N.B. Bidders that do not submit a written confirmation from the manufacturer, stating that the unit offered is provided with a lifetime calibration and manufacturing warranty will not be considered.

BIDDER'S COMMENTS:

Clause T2

The unit offered must be fitted with latex-free extendable tubing of at least 2 metre length.

BIDDER'S COMMENTS:

Clause T3

THE SPECIFICATION OF THE GAUGE MUST COMPLY AS FOLLOWS:

- a. The gauge case must be manufactured from a durable non-corrosion material and it must be dust proof.
- b. The gauge must be easy to read and it must be protected by a transparent glass or suitable plastic cover.
- c. The gauge must be provided with an accurate observation pointer for reading of the scale.
- d. When no pressure is applied the pointer must not leave the zero mark.
- e. The motion of the pointer must be smooth and unobstructed, when pressure is applied.
- f. When pressure is completely released, the pointer must return to the zero mark.
- g. The gauge must provide precision movement and must be accurate within one (1) percent of scale $\pm 3\text{mmHg}$.
- h. The gauge must have large clear numerals, which are easy to read and it must provide a measuring range from 0 to 300mmHg.
- i. Must have a large distal face of at least 15cm width or diameter and brightly coloured for easy reading.
- j. The gauge must have a swivel base so that staff (operator) can read from various angles $\pm 360^\circ$, 40° swivel.
- k. Unit with laser-engraved dial face will be given preference.

BIDDER'S COMMENTS:

Clause T4

SPECIFICATION FOR THE INFALATION CUFFS OFFERED MUST CONFORM AS FOLLOWS:

- T4.1 The cuffs offered must be of a nylon durable latex free material.
- T4.2 The latex free nylon durable cuff must not make use of an internal bladder.

BIDDER'S COMMENTS:

Clause T5

THE SPECIFICATION FOR INFLATION BULB ON THE UNIT OFFERED MUST BE AS FOLLOWS:

- T5.1 The inflation bulb offered must be of a latex free durable type with an internal steel non-return valve.
- T5.2 The inflation valve on the inflation bulb offered must be of a durable steel material chromium plated and operated from 260mmHg to 15mmHg max deflation time 10 sec.
- T5.3 The inflation valve on the inflation bulb offered must have a chromium plated air tap with release regulation by means of a reliable screw valve.

BIDDER'S COMMENTS:

Clause T6

One adult, one large adult, one obese adult and one paediatric latex free durable nylon cuff must be included with the unit offered at no extra cost to the final bid price.

BIDDER'S COMMENTS:

Clause T7

A facility for storage of tubing and cuff etc. must be supplied with the unit offered.

BIDDER'S COMMENTS:

Clause T8

State other options of latex free nylon cuffs that are available and the cost inclusive of V.A.T. of these together with the catalogue number must be stated on the schedule of optional accessories.

BIDDER'S COMMENTS:

Clause T9

UPGRADABILITY

All future upgrades (hardware and software), where applicable, involving patient safety must be supplied at no additional cost.

All future upgrades removing software viruses from existing software, where applicable, must be supplied at no additional cost.

Any software upgrade, where applicable, before or after installation of the equipment must be brought to the attention of the CEO of the Institution and the Manager, Health Technology Services.

BIDDER'S COMMENTS:

Clause T10

MANUALS

The bidder must include in their offer at no extra cost to the final bid price:

- (a) Complete user Operation/Maintenance Manual x 2 (two) Book/File; CD; DVD copies in English Language
- (b) Complete ORIGINAL Service/Repair Manual x 2 (two) Book/File; CD; DVD copies in English Language which MUST include the following information:
 - (i) Fault Finding Guide
 - (ii) Circuit Diagrams/Schematics
 - (iii) Circuit Descriptions
 - (iv) PCB Layouts
 - (v) Calibration Guide
 - (vi) Part numbers and exploded diagram of mechanical parts/panels.

The offer submitted must be supported by descriptive literature, colour pamphlets, colour brochures and technical data sheets applicable to the offer.

FAILURE TO SUBMIT THE ABOVE WILL RESULT IN THE BID BEING DISQUALIFIED.

BIDDER'S COMMENTS:

Clause T11

RADIATION CONTROL LICENCE

Bidders must state the Radiation Control Licence number of the make and model of the equipment offered. If this type of equipment/apparatus appears on the schedule of Hazardous Substances, issued by the Directorate: Radiation Control of the

Department of Health, a licence in terms of the Act on Hazardous Substances (Act 15/1973) must be submitted with the bid document. The licence must be registered under the bidders name or the letter of Joint Venture must be submitted by the Licence holder where the licence is not in the name of the bidder.

BIDDERS THAT NEGLECT TO SUBMIT A LICENCE WILL BE DISQUALIFIED

BIDDER'S COMMENTS:

Clause T12

GUARANTEE / WARRANTY

The bidder must provide a minimum of 24-month warranty / guarantee period for the unit offered.

NB: The bidder must state the duration of warranty/guarantee period for the Transducers supplied with the main monitor unit at no extra cost to the Department of Health Institution.

BIDDER'S COMMENTS:

Clause T13

MAINTENANCE AND SERVICE AGREEMENT

Upon termination of the guarantee / warranty period the DOH Health Technology Services Component Technicians will be responsible for the maintenance, repair and service of the unit offered.

BIDDER'S COMMENTS:

DETAILED TECHNICAL SPECIFICATION

GENERAL INFORMATION REQUIRED

FAILURE TO COMPLETE THIS PART WILL DISQUALIFY THE BIDDER

Make: _____

Model Number / Part Number for: _____

Country of Origin _____

Delivery Period _____

R S A Import Permit Holder (License No) _____

BIDDER _____

SIGNATURE _____ DATE _____

ADDRESS _____

TELEPHONE NO _____ FAX NO. _____


CONTACT PERSON _____
(PLEASE PRINT)

SAH SPECIFICATION NO.: BT 01/1920
ROOM THERMOMETER

SECTION A – General

Clause	Clause Description	Bidders Remarks
G1	Bidders must quote on the supply and delivery of the item.	
G2	SABS/SANS proof of certification or any other relevant certification documentation must be provided with the quotation. Bidder must forward CSD forms with completed quotation and spec forms.	
G3	Bidders must provide a sample/picture in colour of the item quoted for.	
G4	Bidder to state the warrantee /guarantee period. If applicable	_____ months

SECTION B – Technical Specification

Clause	Clause Description	Bidders Remarks
T1	The required indoor room thermometer must have a large LCD screen that shows you indoor temperature and humidity.	
T2	Must have a sampling rate at least every hour. Preferably the day of the week or date/time can be set.	
T3	Must have a reset button for daily readings.	
T4	Must be a wall mounted unit. Must be detachable from wall for battery change.	
T5	Batteries for the required item must be included and ready for use.	
T6	<p>The item must be equivalent to the diagram as below:</p> 	

SECTION C – Confirmation

Compulsory to be completed by the bidder and returned with quotation:

I,,		from
<i>(Print Name)</i>		<i>(Name of Business)</i>
acknowledge that I have read and understand the specifications as laid out above and will ensure that the quotation price submitted will address all the requirements as stipulated.		
.....	 / /
Signature		Date

SECTION D – End User Confirmation

NAME	DESIGNATION	DEPARTMENT	SIGNATURE	DATE

Reference: ZNQ

SAH SPECIFICATION: S02/1920
SPECULUM VAGINAL CUSCO

SECTION A - GENERAL

Clause	Clause Description	Bidders Remarks
G1	Bidders must quote on the supply and delivery of the item.	
G2	CE/ISO13485 proof of certification or any other relevant certification documentation must be provided with the quotation.	
G3	Bidders must provide a sample/picture in colour or catalogue of the item quoted for.	

SECTION B – Technical Specification

Clause	Clause Description	Bidders Remarks
T1	USE: To open the genital tract for easy access to the cervix.	
T2	2.1 Vaginal Speculum for single use 2.2 Types: lateral screw type, push type, middle screw type etc., 2.3 Sizes: Large, Medium, Small ; 2.4 Sterilized by EO gas, Nontoxic. 2.5 Medical grade PS, non-toxic 2.6 Standards: CE, ISO13485, FDA Proved 2.7 Customized color & Logo are available 2.8 Private design & package are welcome	
T3	Material: 3.1 It is made by injection moulding of PS plastic, non-toxic, non-irritant. 3.2 It can adjust at will, smooth edges from plastic resin-less, apprehension from patients. 3.3 Single use only, 3.4 Sterile unless package is opened or damaged, sterilized by ethylene oxide. 3.5 Size: Large, Medium, Small	
T4	4.1 Vaginal Speculum is made by injection moulding of PS plastic, non-toxic, non-irritant. 4.2. Can adjust at will, smooth edges from plastic resin-less, apprehension from patients. 4.3. For single use only, sterile unless package is opened or damaged, sterilized by ethylene oxide 4.4 Sterile, do not use if the packing is damaged or open 4.5 For single use only, Prohibited to re-use 4.6. Store under shady, cool, dry, ventilated and clean condition	

T5	The packet/box must include the following details: 5.1 Lot No. 5.2 Ref. No. 5.3 Batch No. 5.4 Name and contact details of the company.	
----	--	--

SECTION C – Confirmation

Compulsory to be completed by the bidder and returned with quotation:


I,	From	do hereby
(Print Name)		(Name of Business)
Acknowledge that I have read and understand the specifications as laid out above and will ensure that the quotation price submitted will address all the requirements as stipulated.		
.....	/...../.....
Signature		Date

SECTION D – End User Confirmation

NAME	DESIGNATION	DEPARTMENT	SIGNATURE	DATE


Thermometer: refrigerator

Catalogue Number	Item Name	Available on Transversal Contract	Contract Number or Code
PHC-E-153	Thermometer: refrigerator	No	

Description	Simple thermometer for use inside a standard refrigerator.	
Use	To determine that the thermostat setting delivers the desired temperature range.	
Clinical or User Specifications	Indicates temperature inside the refrigerator at any given time. Does not record maximum and minimum temperatures. Analogue-type. No batteries required. Stainless steel.	
Accessories or Related Consumables		
Care	Do not drop.	
Supplier Warranty	2 years (excluding deliberate or accidental damage).	

Peak flow meter

Catalogue Number	Item Name	Available on Transversal Contract	Contract Number or Code
PHC-E-267	Peak flow meter	No	RT 252-01-163 ME

Description	<p>A tubular instrument with a mouthpiece and a scale on which the peak flow is read.</p>	
Use	<p>Used to measure the peak expiratory flow (in l/min), which is a proxy for the narrowing of the airways in acute or chronic airways disease. The patient is encouraged to exhale through the instrument as forcefully as possible, and the position of a sliding pointer is read off a scale.</p>	
Clinical or User Specifications	<p>Disposable paper mouthpiece inserts. Direct-reading scale on side of instrument.</p>	
Accessories/related consumables	<p>Disposable mouthpiece PHC-C-341.</p>	
Care	<p>Standard cleaning and disinfection.</p>	
Supplier Warranty	<p>1 year.</p>	