

BID NUMBER:	ZNB 9044/2024-H	
BID DESCRIPTION:	SUPPLY, DELIVERY, INSTALLATION, DEMONSTRATION AND COMMISSIONING 128 SLICE LARGE BORE COMPUTERISED TOMOGRAPHY SCANNER SYSTEM AT ADDINGTON HOSPITAL.	
PERIOD	ONCE-OFF	
COMPULSORY SITE INSPECTION:	ADDINGTON HOSPITAL	
DATE, TIME AND VENUE	PHYSICAL ADDRESS :	
	ERSKINE TERRACE SOUTH BEACH DURBAN 4000	
	03 DECEMBER 2024 @10:00AM	
	40 1401114 707 0005	
Closing Date:	13 JANUARY 2025	
Closing Date: Closing Time:	13 JANUARY 2025 11:00	
Closing Time:	11:00	
Closing Time: Physical Address for Collection or	11:00 KZN Department of Health	
Closing Time:	11:00 KZN Department of Health Central Supply Chain Management Unit	
Closing Time: Physical Address for Collection or	11:00 KZN Department of Health Central Supply Chain Management Unit Old Boys School	
Closing Time: Physical Address for Collection or	11:00 KZN Department of Health Central Supply Chain Management Unit Old Boys School 310 Jabu Ndlovu Street	
Closing Time: Physical Address for Collection or	11:00 KZN Department of Health Central Supply Chain Management Unit Old Boys School 310 Jabu Ndlovu Street	
Closing Time: Physical Address for Collection or Delivery of Bid Documents	11:00 KZN Department of Health Central Supply Chain Management Unit Old Boys School 310 Jabu Ndlovu Street	

KWAZULU-NATAL PROVINCIAL GOVERNMENT BIDDING FORMS

SECTIONS	DESCRIPTION	PAGE NO.
PART A	INVITATION TO BID (SBD 1)	3
PART B	TERMS AND CONDITIONS FOR BIDDING (SBD 1)	4
SECTION A	SPECIAL INSTRUCTIONS REGARDING COMPLETION OF BID	5
SECTION B	REGISTRATION ON CENTRAL SUPPLIERS DATABASE	6
SECTION C	DECLARATION THAT INFORMATION ON CENTRAL SUPPLIER DATABASE IS CORRECT AND UP TO DATE	7
SECTION D	OFFICIAL BRIEFING SESSION FORM	8
SECTION E	BIDDER'S DISCLOSURE (SBD 4)	10
SECTION F	THE NATIONAL INDUSTRIAL PARTICIPATION PROGRAMME (SBD 5)	12
SECTION G	PREFERENCE POINTS CLAIM FORM (SBD 6.1)	15
SECTION H	GENERAL CONDITIONS OF CONTRACT (GCC)	21
SECTION I	SPECIAL CONDITIONS OF CONTRACT (SCC)	22
SECTION J	EVALUATION CRITERIA	35
SECTION K	AUTHORITY TO SIGN THE BID	37
SECTION L	TERMS OF REFERENCE (SPECIFICATIONS)	38
SECTION M	PRICING SCHEDULE (SBD 3.2)	79



PART A INVITATION TO BID

YOU ARE HEREBY INVITED TO BID FOR REQUIREMENTS OF THE (NAME OF DEPARTMENT/ PUBLIC ENTITY)								
	NB 9044/2024-H							
		PLY, DELIVERY, INSTALLATION, DEMONSTRATION AND COMMISSIONING 128 SLICE LARGE BORE COMPUTERISED TOMOGRAPHY						
		ADDINGTON HOSPITAL. ON		AT (OTDEET AD	0050	.01		
		OSITED IN THE BID BOX SIDIRECTORATE (OLD BOYS)					DIETEDMADITZRI IDG	3200
OLIVINAL OUT LI	OTAIN MANAGEMENT	SINCOTONATE (OLD BOTO		BOILDING), 510	סאסט	MDLOVO OTKLLT, T	ILILIMANIZBONO	3200
BIDDING PROCEDU	IRE ENQUIRIES MAY BE	DIRECTED TO	TECHN	IICAL ENQUIRIE	S MAY	BE DIRECTED TO:		
CONTACT PERSON	Demand Managem	ent	CONTA	ACT PERSON		Mr T. Ngidi		
TELEPHONE NUMBER	033 815 8361/8386/	8357	TELEP	HONE NUMBER		031 714 3711		
FACSIMILE NUMBE	R		FACSIN	MILE NUMBER				
E-MAIL ADDRESS	Scm.demandmana	gement@kznhealth.gov.za				Thami.ngidi@kzn	nhealth.gov.za	
SUPPLIER INFORM	ATION							
NAME OF BIDDER								
POSTAL ADDRESS								
STREET ADDRESS								
TELEPHONE NUMBER	CODE			NUMBER				
CELLPHONE NUMBER								
FACSIMILE NUMBE	R CODE			NUMBER				
E-MAIL ADDRESS								
VAT REGISTRATION NUMBER								
SUPPLIER	TAX COMPLIANCE			CENTRAL				
COMPLIANCE STATUS	SYSTEM PIN:		OR	SUPPLIER DATABASE				
SIAIUS				No:	MAA	AA		
B-BBEE STATUS	TICK AP	PLICABLE BOX]	B-BBE	E STATUS LEVEL			APPLICABLE BOX]	
LEVEL			SWOR	N AFFIDAVIT				
VERIFICATION		□ Na				□ v ₌		
CERTIFICATE	☐ Yes	☐ No				☐ Ye	es 🗌 No	
[A B-BBEE STATUS	S LEVEL VERIFICATION	I CERTIFICATE/ SWORN AF	FIDAVIT	(FOR EMES &	QSEs)	MUST BE SUBMITT	ED IN ORDER TO Q	UALIFY FOR
PREFERENCE POIN	NTS FOR B-BBEE]		ı					
ARE YOU THE ACCREDITED			ΔRF Y(OU A FOREIGN				
REPRESENTATIVE				SUPPLIER FOR		□Yes		□No
IN SOUTH AFRICA	□Yes	□No	THE G	OODS /SERVICE		_		_
FOR THE GOODS	_		/WORK	S OFFERED?		[IF YES, ANSWER TH	E QUESTIONNAIRE F	BELOW]
/SERVICES /WORKS OFFERED?	S [IF YES ENCLOSE	PROOF]						
QUESTIONNAIRE T	O BIDDING FOREIGN SU	JPPLIERS						
IS THE ENTITY A RESIDENT OF THE REPUBLIC OF SOUTH AFRICA (RSA)?								
DOES THE ENTITY HAVE A BRANCH IN THE RSA?								
DOES THE ENTITY	HAVE A PERMANENT E	STABLISHMENT IN THE RSA	.?			☐ YES ☐] NO	
DOES THE ENTITY	HAVE ANY SOURCE OF	INCOME IN THE RSA?				☐ YES ☐] NO	
IF THE ANSWER IS	IS THE ENTITY LIABLE IN THE RSA FOR ANY FORM OF TAXATION? IF THE ANSWER IS "NO" TO ALL OF THE ABOVE, THEN IT IS NOT A REQUIREMENT TO REGISTER FOR A TAX COMPLIANCE STATUS SYSTEM PIN CODE FROM THE SOUTH AFRICAN REVENUE SERVICE (SARS) AND IF NOT REGISTER AS PER 2.3 BELOW.			EM PIN CODE				

PART B TERMS AND CONDITIONS FOR BIDDING

BID SUBMISSION:

- a. BIDS MUST BE DELIVERED BY THE STIPULATED TIME TO THE CORRECT ADDRESS. LATE BIDS WILL NOT BE ACCEPTED FOR CONSIDERATION.
- b. ALL BIDS MUST BE SUBMITTED ON THE OFFICIAL FORMS PROVIDED-(NOT TO BE RE-TYPED) OR IN THE MANNER PRESCRIBED IN THE BID DOCUMENT.
- c. THIS BID IS SUBJECT TO THE PREFERENTIAL PROCUREMENT POLICY FRAMEWORK ACT, 2000 AND THE PREFERENTIAL PROCUREMENT REGULATIONS, 2022, THE GENERAL CONDITIONS OF CONTRACT (GCC) AND, IF APPLICABLE, ANY OTHER SPECIAL CONDITIONS OF CONTRACT.
- d. THE SUCCESSFUL BIDDER WILL BE REQUIRED TO FILL IN AND SIGN A WRITTEN CONTRACT FORM (SBD7).

. TAX COMPLIANCE REQUIREMENTS

- (a) BIDDERS MUST ENSURE COMPLIANCE WITH THEIR TAX OBLIGATIONS.
- (b) BIDDERS ARE REQUIRED TO SUBMIT THEIR UNIQUE PERSONAL IDENTIFICATION NUMBER (PIN) ISSUED BY SARS TO ENABLE THE ORGAN OF STATE TO VERIFY THE TAXPAYER'S PROFILE AND TAX STATUS.
- (c) APPLICATION FOR TAX COMPLIANCE STATUS (TCS) PIN MAY BE MADE VIA E-FILING THROUGH THE SARS WEBSITE WWW.SARS.GOV.ZA.
- (d) BIDDERS MAY ALSO SUBMIT A PRINTED TCS CERTIFICATE TOGETHER WITH THE BID.
- (e) IN BIDS WHERE CONSORTIA / JOINT VENTURES / SUB-CONTRACTORS ARE INVOLVED, EACH PARTY MUST SUBMIT A SEPARATE TCS CERTIFICATE / PIN / CSD NUM§BER.
- (f) WHERE NO TCS PIN IS AVAILABLE BUT THE BIDDER IS REGISTERED ON THE CENTRAL SUPPLIER DATABASE (CSD), A CSD NUMBER MUST BE PROVIDED.
- (g) NO BIDS WILL BE CONSIDERED FROM PERSONS IN THE SERVICE OF THE STATE, COMPANIES WITH DIRECTORS WHO ARE PERSONS IN THE SERVICE OF THE STATE, OR CLOSE CORPORATIONS WITH MEMBERS PERSONS IN THE SERVICE OF THE STATE."

NB: FAILURE TO PROVIDE / OR COMPLY WITH ANY OF THE ABOVE PA	ARTICULARS MAY RENDER THE BID INVALID
SIGNATURE OF BIDDER:	
CAPACITY UNDER WHICH THIS BID IS SIGNED: (Proof of authority must be submitted e.g. company resolution)	
DATE:	

SECTION A

SPECIAL INSTRUCTIONS AND NOTICES TO BIDDERS REGARDING THE COMPLETION OF BIDDING FORMS

PLEASE NOTE THAT THIS BID IS SUBJECT TO TREASURY REGULATIONS 16A ISSUED IN TERMS OF THE PUBLIC FINANCE MANAGEMENT ACT, 1999, THE KWAZULU-NATAL SUPPLY CHAIN MANAGEMENT POLICY FRAMEWORK.

- 1. Unless inconsistent with or expressly indicated otherwise by the context, the singular shall include the plural and visa versa and with words importing the masculine gender shall include the feminine and the neuter.
- 2. Under no circumstances whatsoever may the bid forms be retyped or redrafted. Photocopies of the original bid documentation may be used, but an original signature must appear on such photocopies.
- 3. The bidder is advised to check the number of pages and to satisfy himself that none are missing or duplicated.
- 4. Bids submitted must be complete in all respects.
- 5. Bids shall be lodged at the address indicated not later than the closing time specified for their receipt, and in accordance with the directives in the bid documents.
- 6. Each bid shall be addressed in accordance with the directives in the bid documents and shall be lodged in a separate sealed envelope, with the name and address of the bidder, the bid number and closing date indicated on the envelope. The envelope shall not contain documents relating to any bid other than that shown on the envelope. If this provision is not complied with, such bids may be rejected as being invalid.
- 7. All bids received in sealed envelopes with the relevant bid numbers on the envelopes are kept unopened in safe custody until the closing time of the bids. Where, however, a bid is received open, it shall be sealed. If it is received without a bid number on the envelope, it shall be opened, the bid number ascertained, the envelope sealed and the bid number written on the envelope.
- 8. A specific box is provided for the receipt of bids, and no bid found in any other box or elsewhere subsequent to the closing date and time of bid will be considered.
- 9. No bid sent through the post will be considered if it is received after the closing date and time stipulated in the bid documentation, and proof of posting will not be accepted as proof of delivery.
- 10. No bid submitted by telefax, telegraphic or other electronic means will be considered.
- 11. Bidding documents must not be included in packages containing samples. Such bids may be rejected as being invalid.
- 12. Any alteration made by the bidder must be initialed.
- 13. Use of correcting fluid is prohibited
- 14. Bids will be opened in public as soon as practicable after the closing time of bid.
- 15. Where practical, prices are made public at the time of opening bids.
- 16. If it is desired to make more than one offer against any individual item, such offers should be given on a photocopy of the page in guestion. Clear indication thereof must be stated on the schedules attached.
- 17. Bidder must initial each and every page of the bid document.

SECTION B

REGISTRATION ON THE CENTRAL SUPPLIERS DATABASE

- 1. In terms of the National Treasury Instruction Note, all suppliers of goods and services to the State are required to register on the Central Suppliers Database.
- 2. Prospective suppliers should self-register on the CSD website www.csd.gov.za
- 3. If a business is registered on the Database and it is found subsequently that false or incorrect information has been supplied, then the Department may, without prejudice to any other legal rights or remedies it may have;
 - 3.1 cancel a bid or a contract awarded to such supplier, and the supplier would become liable for any damages if a less favourable bid is accepted or less favourable arrangements are made.
- 4. The same principles as set out in paragraph 3 above are applicable should the supplier fail to request updating of its information on the Central Suppliers Database, relating to changed particulars or circumstances.
- 5. IF THE SUPPLIER IS NOT REGISTERED AT THE CLOSING TIME OF BID, THE SUPPLIER WILL BE DISQUALIFIED AT THE BID EVALUATION PROCESS.

SECTION C

DECLARATION THAT INFORMATION ON CENTRAL SUPPLIER DATABASE IS CORRECT AND UP TO DATE (To be completed by bidder)

THIS IS TO CERTIFY THAT I (name of bidder/authorized representative)	, WHO
REPRESENTS (state name of bidder)	CSD Registration
Number	
AM AWARE OF THE CONTENTS OF THE CENTRAL SUPPLIER DATABASE WITH RESPECT TO THE BID REGISTRATION INFORMATION, AND THAT THE SAID INFORMATION IS CORRECT AND UP TO DATE λ SUBMITTING THIS BID.	
AND I AM AWARE THAT INCORRECT OR OUTDATED INFORMATION MAY BE A CAUSE FOR DISQUA BID FROM THE BIDDING PROCESS, AND/OR POSSIBLE CANCELLATION OF THE CONTRACT THAT MATHE BASIS OF THIS BID.	
SIGNATURE OF BIDDER OR AUTHORISED REPRESENTATIVE	
DATE:	

SECTION D

OFFICIALSITE INSPECTION CERTIFICATE

ZNB9044/2024-H: SUPPLY, DELIVERY, INSTALLATION, DEMONSTRATION AND COMMISSIONING 128 SLICE LARGE BORE COMPUTERISED TOMOGRAPHY SCANNER SYSTEM AT ADDINGTON HOSPITAL. ONCE-OFF.

N. B.: THIS FORM IS ONLY TO BE COMPLETED WHEN APPLICABLE TO THE BID.

Site/Building/Institution Involved: ADDINGTON HOSPITAL

PHYSICAL ADDRESS: ERSKINE TERRACE SOUTH BEACH DURBAN 4000
Bid Reference No: ZNB9044/2024-H
Goods/Service/Work: SUPPLY, DELIVERY, DE-COMMISSIONING AND INSTALL 128 SLICE LARGE BORE COMPUTERISED TOMOGRAPHY SCANNER SYSTEM AT ADDINGTON HOSPITAL: ONCE-OFF.

This is to certify that (bidder's representative name)
On behalf of (company name)
Visited and inspected the site on (date) 03 DECEMBER 2024 @10:00AM and is therefore familiar with the circumstances and the scope of the service to be rendered.
Signature of Bidder or Authorized Representative (PRINT NAME)
DATE://
Name of Departmental or Public Entity Representative (PRINT NAME)
Departmental Stamp With Signature

BIDDER NAME				
	I FGISI ATIOI	N ON DISCLOSURE	OF INTEREST	_
herself to perform re	Act 103 of 1994 indicates in	n section 30(1) that " or her employment i	No employee shall perform or en the relevant department, excep	
any organ of state o	or be a director of a public or pi	rivate company condu	(c), "An employee shall not conducting business with an organ of s nedule 2 and 3 of the Public Finar	tate, unless such
close family member contract to be award	er, partner or associate of sucl	h official or other role player must-(a) disclo	n management official or other ro player, has any private or busine se that interest; and (b) withdraw	ss interest in any
		TY ON HOW TO DIS		
the entire KZN Department use other Com	artment of Health, even if that puter Assisted Techniques to	t person is not emplo verify possible intere	close a relationship with any pers yed by the procuring institution. T est, should you be found to have t ted as non-responsive and disqua	he Department failed to disclose
Manguzi Hospital, a interest. Therefore to person who is employed	as long as that official is emplo the question is, do you, or any	oyed by the Departme y person connected w	cal, yet the person with interest is ent of Health, the bidder is require with the bidder, have a relationship ase furnish particulars on Bidders	ed to disclose o with any
	arity on disclosure of interest a		ose as directed, should I fail to dis	sclose correctly, I
BIDDER SURNAM	E AND INITIALS	SIGNATURE	DATE	
This document mu	ust be signed and submitted to	ogether with your bid		

SBD 4

SECTION E

BIDDER'S DISCLOSURE

1. PURPOSE OF THE FORM

Any person (natural or juristic) may make an offer or offers in terms of this invitation to bid. In line with the principles of transparency, accountability, impartiality, and ethics as enshrined in the Constitution of the Republic of South Africa and further expressed in various pieces of legislation, it is required for the bidder to make this declaration in respect of the details required hereunder.

Where a person/s are listed in the Register for Tender Defaulters and / or the List of Restricted Suppliers, that person will automatically be disqualified from the bid process.

2. Bidder's declaration

2.1 Is the bidder, or any of its directors / trustees / shareholders / members / partners or any person having a controlling interest1 in the enterprise, employed by the state?

YES/NO

2.1.1. If so, furnish particulars of the names, individual identity numbers, and, if applicable, state employee numbers of sole proprietor/ directors / trustees / shareholders / members/ partners or any person having a controlling interest in the enterprise, in the table below.

Full Name	Identity Number	Name of State institution

2.2.	Do you, or any person connected with the bidder, have a relationship with any person who is employed by the procuring institution? YES/NO
	2.2.1.If so, furnish particulars:
2.3.	Does the bidder or any of its directors / trustees / shareholders / members / partners or any person having a controlling interest in the enterprise have any interest in any other related enterprise whether or not they are bidding for this contract? YES/NO
	2.3.1 If so, furnish particulars:

ZNB9044/2024-H

¹ the power, by one person or a group of persons holding the majority of the equity of an enterprise, alternatively, the person/s having the deciding vote or power to influence or to direct the course and decisions of the enterprise.

3. DECLARATION

- 3.1 I have read and I understand the contents of this disclosure;
- 3.2 I understand that the accompanying bid will be disqualified if this disclosure is found not to be true and complete in every respect;
- 3.3 The bidder has arrived at the accompanying bid independently from, and without consultation, communication, agreement or arrangement with any competitor. However, communication between partners in a joint venture or consortium will not be construed as collusive bidding.
- 3.4 In addition, there have been no consultations, communications, agreements or arrangements with any competitor regarding the quality, quantity, specifications, prices, including methods, factors or formulas used to calculate prices, market allocation, the intention or decision to submit or not to submit the bid, bidding with the intention not to win the bid and conditions or delivery particulars of the products or services to which this bid invitation relates.
- 3.5 The terms of the accompanying bid have not been, and will not be, disclosed by the bidder, directly or indirectly, to any competitor, prior to the date and time of the official bid opening or of the awarding of the contract.
- 3.6 There have been no consultations, communications, agreements or arrangements made by the bidder with any official of the procuring institution in relation to this procurement process prior to and during the bidding process except to provide clarification on the bid submitted where so required by the institution; and the bidder was not involved in the drafting of the specifications or terms of reference for this bid.
- 3.7 I am aware that, in addition and without prejudice to any other remedy provided to combat any restrictive practices related to bids and contracts, bids that are suspicious will be reported to the Competition Commission for investigation and possible imposition of administrative penalties in terms of section 59 of the Competition Act No 89 of 1998 and or may be reported to the National Prosecuting Authority (NPA) for criminal investigation and or may be restricted from conducting business with the public sector for a period not exceeding ten (10) years in terms of the Prevention and Combating of Corrupt Activities Act No 12 of 2004 or any other applicable legislation.

I CERTIFY THAT THE INFORMATION FURNISHED IN PARAGRAPHS 1, 2 and 3 ABOVE IS CORRECT.

I ACCEPT THAT THE STATE MAY REJECT THE BID OR ACT AGAINST ME IN TERMS OF PARAGRAPH 6 OF PFMA SCM INSTRUCTION 03 OF 2021/22 ON PREVENTING AND COMBATING ABUSE IN THE SUPPLY CHAIN MANAGEMENT SYSTEM SHOULD THIS DECLARATION PROVE TO BE FALSE.

Signature	
	Date
Position	Name of bidder

	SBD 5
This document must be signed and submitted together with your bid	

11

² Joint venture or Consortium means an association of persons for the purpose of combining their expertise, property, capital, efforts, skill and knowledge in an activity for the execution of a contract.

SECTION F:

THE NATIONAL INDUSTRIAL PARTICIPATION PROGRAMME

INTRODUCTION

The National Industrial Participation (NIP) Programme, which is applicable to all government procurement contracts that have an imported content, became effective on the 1 September 1996. The NIP policy and guidelines were fully endorsed by Cabinet on 30 April 1997. In terms of the Cabinet decision, all state and parastatal purchases / lease contracts (for goods, works and services) entered into after this date, are subject to the NIP requirements. NIP is obligatory and therefore must be complied with. The Industrial Participation Secretariat (IPS) of the Department of Trade and Industry (DTI) is charged with the responsibility of administering the programme.

1. PILLARS OF THE PROGRAMME

- 1.1. The NIP obligation is benchmarked on the imported content of the contract. Any contract having an imported content equal to or exceeding US\$ 10 million or other currency equivalent to US\$ 10 million will have a NIP obligation. This threshold of US\$ 10 million can be reached as follows:
 - i) Any single contract with imported content exceeding US\$10 million.

or

ii) Multiple contracts for the same goods, works or services each with imported content exceeding US\$3 million awarded to one seller over a 2 year period which in total exceeds US\$10 million.

or

iii) A contract with a renewable option clause, where should the option be exercised the total value of the imported content will exceed US\$10 million.

OI

- iv) Multiple suppliers of the same goods, works or services under the same contract, where the value of the imported content of each allocation is equal to or exceeds US\$ 3 million worth of goods, works or services to the same government institution, which in total over a two (2) year period exceeds US\$10 million.
- 1.2. The NIP obligation applicable to suppliers in respect of sub-paragraphs 1.1 (a) to 1.1 (c) above will amount to 30 % of the imported content whilst suppliers in respect of paragraph 1.1 (d) shall incur 30% of the total NIP obligation on a *pro-rata* basis.
- 1.3. To satisfy the NIP obligation, the DTI would negotiate and conclude agreements such as investments, joint ventures, sub-contracting, licensee production, export promotion, sourcing arrangements and research and development (R&D) with partners or suppliers.
- 1.4. A period of seven years has been identified as the time frame within which to discharge the obligation

2. REQUIREMENTS OF THE DEPARTMENT OF TRADE AND INDUSTRY

- 2.1. In order to ensure effective implementation of the programme, successful bidders (contractors) are required to, immediately after the award of a contract that is in excess of R10 million (ten million Rands), submit details of such a contract to the DTI for reporting purposes.
- 2.2. The purpose for reporting details of contracts in excess of the amount of R10 million (ten million Rands) is to cater for multiple contracts for the same goods, works or services; renewable contracts and multiple suppliers for the same goods, works or services under the same contract as provided for in paragraphs 1.1.(b) to 1.1. (d) above.

3. BID SUBMISSION AND CONTRACT REPORTING REQUIREMENTS OF BIDDERS AND SUCCESSFUL BIDDERS (CONTRACTORS)

- 3.1. Bidders are required to sign and submit this Standard Bidding Document (SBD 5) together with the bid on the closing date and time.
- 3.2. In order to accommodate multiple contracts for the same goods, works or services; renewable contracts and multiple suppliers for the same goods, works or services under the same contract as indicated in subparagraphs 1.1 (b) to 1.1 (d) above and to enable the DTI in determining the NIP obligation, successful bidders (contractors) are required, immediately after being officially notified about any successful bid with a value in excess of R10 million (ten million Rands), to contact and furnish the DTI with the following information:
 - I. Bid / contract number.
 - II. Description of the goods, works or services.
 - III. Date on which the contract was accepted.
 - IV. Name, address and contact details of the government institution.
 - V. Value of the contract.
 - VI. Imported content of the contract, if possible.
- 3.3. Private Bag X 84, Pretoria, 0001 for the attention of Mr Elias Malapane within five (5) working days after award of the contract. Mr Malapane may be contacted on telephone (012) 394 1401, facsimile (012) 394 2401 or email at Elias@thedti.gov.za for further details about the programme.

4. PROCESS TO SATISFY THE NIP OBLIGATION

- 4.1. Once the successful bidder (contractor) has made contact with and furnished the DTI with the information required, the following steps will be followed:
 - i) the contractor and the DTI will determine the NIP obligation;
 - ii) the contractor and the DTI will sign the NIP obligation agreement;
 - iii) the contractor will submit a performance guarantee to the DTI;
 - iv) the contractor will submit a business concept for consideration and approval by the DTI:
 - v) upon approval of the business concept by the DTI, the contractor will submit detailed business plans outlining the business concepts;
 - vi) the contractor will implement the business plans; and
 - vii) the contractor will submit bi-annual progress reports on approved plans to the DTI.
- 4.2. The NIP obligation agreement is between the DTI and the successful bidder (contractor) and, therefore, does not involve the purchasing institution.

Bid number: ZNB 9044/2024-H
Name of bidder
Postal address
Signature Name (in print)
Date

SECTION G

PREFERENCE POINTS CLAIM FORM IN TERMS OF THE PREFERENTIAL PROCUREMENT REGULATIONS 2022:

This preference form must form part of all Bids invited. It contains general information and serves as a claim form for preference points for specific goals.

NB: BEFORE COMPLETING THIS FORM, BIDERS MUST STUDY THE GENERAL CONDITIONS, DEFINITIONS AND DIRECTIVES APPLICABLE IN RESPECT OF THE BID AND PREFERENTIAL PROCUREMENT REGULATIONS, 2023

1. GENERAL CONDITIONS

- 1.1 The following preference point systems are applicable to invitations to Bid:
 - the 80/20 system for requirements with a Rand value of up to R50 000 000 (all applicable taxes included); and
 - the 90/10 system for requirements with a Rand value above R50 000 000 (all applicable taxes included).

1.2 To be completed by the organ of state

- a. The applicable preference point system for this Bid is the 80/20preference point system.
- b. The 80/20preference point system will be applicable in this Bid. The lowest/ highest acceptable Bid will be used to determine the accurate system once Bids are received.
- 1.3 Points for this Bid (even in the case of a Bid for income-generating contracts) shall be awarded for:
 - (a) Price; and
 - (b) Specific Goals.

1.4 To be completed by the organ of state:

The maximum points for this Bid are allocated as follows:

	POINTS
PRICE	80
SPECIFIC GOALS	20
Total points for Price and SPECIFIC GOALS	100

- 1.5 Failure on the part of a Bidder to submit proof or documentation required in terms of this Bid to claim points for specific goals with the Bid, will be interpreted to mean that preference points for specific goals are not claimed.
- 1.6 The organ of state reserves the right to require of a Bidder, either before a Bid is adjudicated or at any time subsequently, to substantiate any claim in regard to preferences, in any manner required by the organ of state.

2. **DEFINITIONS**

- (a) "Bid" means a written offer in the form determined by an organ of state in response to an invitation to provide goods or services through price quotations, competitive Bidding process or any other method envisaged in legislation;
- (b) "price" means an amount of money Bided for goods or services, and includes all applicable taxes less all unconditional discounts;
- (c) "rand value" means the total estimated value of a contract in Rand, calculated at the time of bid invitation, and includes all applicable taxes;
- (d) "Bid for income-generating contracts" means a written offer in the form determined by an organ of state in response to an invitation for the origination of income-generating contracts through any method envisaged in legislation that will result

in a legal agreement between the organ of state and a third party that produces revenue for the organ of state, and includes, but is not limited to, leasing and disposal of assets and concession contracts, excluding direct sales and disposal of assets through public auctions: and

(e) "the Act" means the Preferential Procurement Policy Framework Act, 2000 (Act No. 5 of 2000).

3. FORMULAE FOR PROCUREMENT OF GOODS AND SERVICES

3.1. **POINTS AWARDED FOR PRICE**

3.1.1 THE 80/20 OR 90/10 PREFERENCE POINT SYSTEMS

A maximum of 80 or 90 points is allocated for price on the following basis:

80/20 or 90/10

$$Ps = 80\left(1 - rac{Pt - P\,min}{P\,min}
ight)$$
 or $Ps = 90\left(1 - rac{Pt - P\,min}{P\,min}
ight)$

Where

Ps = Points scored for price of Bid under consideration

Pt = Price of Bid under consideration

Pmin = Price of lowest acceptable Bid

3.2. FORMULAE FOR DISPOSAL OR LEASING OF STATE ASSETS AND INCOME GENERATING PROCUREMENT

3.2.1. POINTS AWARDED FOR PRICE

A maximum of 80 or 90 points is allocated for price on the following basis:

80/20 or 90/10

$$Ps = 80\left(1 + \frac{Pt - P max}{P max}\right)$$
 or $Ps = 90\left(1 + \frac{Pt - P max}{Pmax}\right)$

Where

Ps = Points scored for price of Bid under consideration

Pt = Price of Bid under consideration

Pmax = Price of highest acceptable Bid

4. POINTS AWARDED FOR SPECIFIC GOALS

- 4.1. In terms of Regulation 4(2); 5(2); 6(2) and 7(2) of the Preferential Procurement Regulations, preference points must be awarded for specific goals stated in the Bid. For the purposes of this Bid the Bider will be allocated points based on the goals stated in table 1 below as may be supported by proof/ documentation stated in the conditions of this Bid:
- 4.2. In cases where organs of state intend to use Regulation 3(2) of the Regulations, which states that, if it is unclear whether the 80/20 or 90/10 preference point system applies, an organ of state must, in the Bid documents, stipulate in the case of—
 - (a) an invitation for Bid for income-generating contracts, that either the 80/20 or 90/10 preference point system will apply and that the highest acceptable Bid will be used to determine the applicable preference point system; or
 - (b) any other invitation for Bid, that either the 80/20 or 90/10 preference point system will apply and that the lowest

acceptable Bid will be used to determine the applicable preference point system,

then the organ of state must indicate the points allocated for specific goals for both the 90/10 and 80/20 preference point system.

Table 1: Specific goals for the Bid and points claimed are indicated per the table below.

(Note to organs of state: Where either the 90/10 or 80/20 preference point system is applicable, corresponding points must also be indicated as such.

Note to Bidders: The Bidder must indicate how they claim points for each preference point system.)

The specific goals allocated points in terms of this Bid	Number of points allocated (80/20 System) (To be completed by the organ of state)	Number of points claimed (80/20) (To be completed by the Bidder)
In terms of Departmental Preferential Procurement Regulation Policy 2024, 10 full points are allocated to companies who are at least 51% owned by Black People in addition 10 full points are allocated to companies who are at least 100% owned by Black Africans or 10 full points are allocated to companies who are at least 100% owned by Black African Woman or 10 full points are allocated to companies who are at least 51% owned by Youth or 10 full points are allocated to companies who are at least 51% owned by black people or persons living with disabilities	20 points (To be allocated for specific goals)	

DECLARATION WITH REGARD TO COMPANY/FIRM

4.3.	Name of company/firm

- 4.4. Company registration number:
- 4.5. TYPE OF COMPANY/ FIRM
 - Y Partnership/Joint Venture / Consortium
 - Y One-person business/sole propriety
 - Υ Close corporation
 - Y Public Company
 - Y Personal Liability Company
 - Υ (Pty) Limited
 - Y Non-Profit Company
 - Y State Owned Company

[TICK APPLICABLE BOX]

- 4.6. I, the undersigned, who is duly authorised to do so on behalf of the company/firm, certify that the points claimed, based on the specific goals as advised in the Bid, qualifies the company/ firm for the preference(s) shown and I acknowledge that:
 - i) The information furnished is true and correct;
 - ii) The preference points claimed are in accordance with the General Conditions as indicated in paragraph 1 of this form;
 - iii) In the event of a contract being awarded as a result of points claimed as shown in paragraphs 1.4 and 4.2, the contractor may be required to furnish documentary proof to the satisfaction of the organ of state that the claims are

correct:

- iv) If the specific goals have been claimed or obtained on a fraudulent basis or any of the conditions of contract have not been fulfilled, the organ of state may, in addition to any other remedy it may have
 - (a) disqualify the person from the Biding process;
 - (b) recover costs, losses or damages it has incurred or suffered as a result of that person's conduct;
 - (c) cancel the contract and claim any damages which it has suffered as a result of having to make less favourable arrangements due to such cancellation;
 - (d) recommend that the Bidder or contractor, its shareholders and directors, or only the shareholders and directors who acted on a fraudulent basis, be restricted from obtaining business from any organ of state for a period not exceeding 10 years, after the *audi alteram partem* (hear the other side) rule has been applied; and
 - (e) forward the matter for criminal prosecution, if deemed necessary.

SIGNATURE(S) OF TEND	DERER(S)
SURNAME AND NAME:	
DATE:	
ADDRESS:	

EME'S AND QSE'S MUST COMPLETE THE FOLLOWING APPLICABLE AFFIDAVIT FORM TO CLAIM PREFERENCE POINTS

SWORN AFFIDAVIT – B-BBEE EXEMPTED MICRO ENTERPRISE

I, the undersigned,

Full name & Surname	
Identity number	

Hereby declare under oath as follows:

- a) The contents of this statement are to the best of my knowledge a true reflection of the facts.
- b) I am a member / director / owner of the following enterprise and am duly authorised to act on its behalf:

Enterprise Name				
Trading Name (If				
Applicable):				
Registration Number				
Enterprise Physical Address:				
Type of Entity (CC, (Pty)				
Ltd, Sole Prop etc.):				
Nature of Business:				
Definition of "Black	As per the Broad-Based Black Economic Empowerment Act 53 of 2003 as			
People"	Amended by Act No 46 of 2013 "Black People" is a generic term which			
	means Africans, Coloureds and Indians –			
	a) who are citizens of the Republic of South Africa by birth or descent;			
	or			
	b) who became citizens of the Republic of South Africa by naturalisationi-			
	i) before 27 April 1994; or			
	ii) on or after 27 April 1994 and who would have been entitled to			
	acquire citizenship by naturalization prior to that date;"			
	acquire citizeriship by hataranzation prior to that date,			
Definition of "Black	"Black Designated Groups means:			
Designated Groups"	a) unemployed black people not attending and not required by law to			
	attend an educational institution and not awaiting admission to an			
	educational institution:			
	,			
	b) Black people who are youth as defined in the National Youth			
	Commission Act of 1996;			
	c) Black people who are persons with disabilities as defined in the Code			
	of Good Practice on employment of people with disabilities issued			
	under the Employment Equity Act;			
	d) Black people living in rural and under developed areas;			
	e) Black military veterans who qualifies to be called a military veteran in			
	terms of the Military Veterans Act 18 of 2011;"			

c) I he	reby declare	under Oath that:	
6.		rprise is% Black Owned as per Amended Code Series 100 of the amended Good Practice issued under section 9 (1) of B-BBEE Act No 53 of 2003 as amended by Act No 3.	0
7.	The Enterp Amended	rprise is% Black Female Owned as per Amended Code Series 100 of the Codes of Good Practice issued under section 9 (1) of B-BBEE Act No 53 of 2003 as Amended 46 of 2013,	d
8. 9.	The Enterpof the Ame Amended I	rprise is% Black Designated Group Owned as per Amended Code Series 10 lended Codes of Good Practice issued under section 9 (1) of B-BBEE Act No 53 of 2003 as by Act No 46 of 2013, signated Group Owned % Breakdown as per the definition stated above:)0
0.	Black Bool	i) Black Youth % =%	
		ii) Black Disabled % =%	
		iii) Black Unemployed % =%	
		iv) Black People living in Rural areas % =%	
		v) Black Military Veterans % =%	
10.	Based on t	the Financial Statements/Management Accounts and other information available on the latest	1
	financial ye	year-end of, the annual Total Revenue was R10,000,000.00 (Ten Million Rand	ds)
	or less		
11.	Please Co	onfirm on the below table the B-BBEE Level Contributor, by ticking the applicable box.	
1000/ DIa	ale Owned	Level One (1259/ D. DDFF procurement recognition	
100% Bla	ick Owned	Level One (135% B-BBEE procurement recognition level)	
At least 5 Owned	1% Black	Level Two (125% B-BBEE procurement recognition level)	
	1 51% Black	Level Four (100% B-BBEE procurement recognition	
Owned		level)	
the	oath binding	erstand the contents of this affidavit and I have no objection to take the prescribed oath and cons on my conscience and on the Owners of the Enterprise, which I represent in this matter. avit will be valid for a period of 12 months from the date signed by commissioner.	ider
Deponent :	Signature:		
Date:	1 1		
Stamp			
Signature	of Commission	ioner of Oaths	

SECTION H:

GENERAL CONDITIONS OF CONTRACT (GCC)

In terms of Treasury Regulation 16A6.3(a)(i) "The accounting officer must ensure that bid documentation and the general conditions of a contract are in accordance with the instructions of the National Treasury."

Bidders are expected to be familiar with the general conditions applicable to government bids, contracts and orders; and rights and obligations of all parties involved in doing business with government.

Bidders are therefore required to initial each page of the attached **Annexure A** for General Conditions of Contract (GCC) and return with the bid document.

I hereby confirm that I have read the General Conditions of Contract (GCC) as published by the National Treasury and I confirm that I fully understands its contents and conditions. I also confirm that I am willfully committing to abiding by its contents.			
Name:		Signature:	
Title/ Role:		Date:	

Note: Should you fail to submit <u>initialed</u> Annexure A for General Conditions of Contract (GCC) and return with the bid document as well as to sign this schedule, your bid may be disqualified.

SECTION I:

SPECIAL CONDITIONS OF CONTRACT (SCC)

1.1. ADDITIONAL DEFINITIONS

In addition to the definitions contained in paragraph 1 of the GCC, the following terms shall be interpreted as indicated:

"Accounting Officer"	means a person described in Section 36 of the Public Finance Management Act, Act No. 1 of 1999		
	(As amended by Act 29 of 1999).		
"Contract Duration"	means the period between the commencement and termination of the contract.		
"Confidential Information"	means but is not limited to contents of the contract, or any provision thereof, or any specification,		
	plan, know-how, drawing, pattern, sample, or information furnished by or on behalf of the		
	Department in connection therewith, to any person other than a person employed by contractor or		
	service provider in the performance of the contract.		
"Department"	means the KwaZulu-Natal Department of Health.		
"Head of Department"	means the Head of Department for KwaZulu-Natal Department of Health as defined in Schedule 2		
·	Column 1 and 2 of the Public Service Act 1994 (Proclamation 103 of 3 June 1994, as amended).		
"Health Facilities"	means Head Office, District Offices, Hospitals, Community Health Centres, Specialized Centres		
	and Clinics under the auspices of the Department of Health in the Province		
"ISO Standards"	means standards recognized by International Standard Organisation		
"Parties"	means the KwaZulu-Natal Department of Health and Contractor or Service provider		
"Province"	means the Province of KwaZulu-Natal.		
"ROE"	means the Rate of Exchange.		
"SABS"	means the South African Bureau of Standards		
"SANS"	means the South African National Standards.		
"Vendor"	means Contracted Supplier or Service Provider		

1.2. INTERPRETATIONS

In amplification of the provisions of paragraph 2 of the GCC, unless inconsistent with the context, an expression which denotes:

- 1.2.1 Any gender includes the other genders.
- 1.2.2 A natural person includes a juristic person and vice versa.
- 1.2.3 The singular includes the plural and vice versa.
- 1.2.4 When any number of days is prescribed in this Contract, the same shall be reckoned exclusively of the first and inclusively of the last day unless the last day falls on a Saturday, Sunday or proclaimed public holiday in the Republic of South Africa, in which event the last day shall be the next succeeding day which is not a Saturday, Sunday or public holiday.
- 1.2.5 Figures are referred to in numerals and in words, if there is any conflict between the two, the words shall prevail.
- 1.2.6 Any reference in this contract to "goods" includes works and/or services.
- 1.2.7 The written and signed contract represents the final agreement between the parties and it super cedes any prior oral agreements or discussions of the Contract.
- 1.2.8 All annexures and appendices shall form part of the contract.
- 1.2.9 The headings used throughout the Contract do not have any special significance save to ensure the easy reading of the contract.
- 1.2.10 Words and phrases defined in this Contract shall bear the meaning assigned to them throughout this Contract.
- 1.2.11 Words and phrases used in this Contract which are defined or used in any statute or regulation which applies to the subject matter, professional person.
- 1.2.12 The bid is issued in accordance with Section 217 of the Constitution, The Public Finance Management Act, Treasury Regulations 16A and National Treasury regulations and guidelines.

1.3. LEGISLATIVE AND REGULATORY FRAMEWORK

- 1.3.1 This bid and all contracts emanating there from will be subject to General Conditions of Contract issued in accordance with Treasury Regulation 16A6.3, published in terms of the Public Finance Management Act,1999 (Act 1 of 1999) (PFMA) as well as the Preferential Procurement Policy Framework Act 2000 (PPPFA), the Preferential Procurement Regulations 2022 (PPR 2022) and KZN Department Preferential Procurement Regulation Policy 2023.
- 1.3.2 The Special Conditions of Contract (SCC) are supplementary to that of General Conditions of Contract (GCC). However, where the Special Conditions of Contract conflict with the General Conditions of Contract, the Special Conditions of Contract prevail.

2.1. ACCEPTANCE OF A BID

- 2.1.1 This Bid will be evaluated and adjudicated in terms of Kwazulu-Natal Department of Health SCM Policy and Delegations.

 The Department of Health Bid Adjudication Committee (DBAC) is under no obligation to accept any bid.
- 2.1.2 The financial standing of a bidder and its ability to render services may be examined before the bid is considered for acceptance.

2.2. CERTIFICATE OF COMPLIANCE

- 2.2.1 If the bidder submits offers for items that make reference to South African National Standards (SANS) or South African Bureau of Standards (SABS) or International Organisation for Standardisation (ISO) specifications, a Certificate of Compliance must be submitted with the bid document at the time of closing of the bid. SABS/SANS can be contacted for testing and conformity services at Tel: 031 203 2900/ Fax: 031 203 2907. SANS, SABS AND CKS specifications will be for the account of the prospective bidder. Failure to submit the certificate, where applicable, will result in the bid being disqualified. The Department reserves its rights to contact SABS/SANS/CKS for testing and conformity services.
- 2.2.2 Prior to an award of the bid being made and/or during the evaluation process, the Department of Health reserves the right to conduct inspections of the premises of the most acceptable bidder. Therefore, premises of the bidder shall be open, at reasonable hours, for inspection by a representative of the Department or organization acting on its behalf. Any specification/s and conformity testing will be for the account of the prospective bidder.
- 2.2.3 Bidders must state the Radiation Control License number of the make and model of the Equipment offered. If this type of equipment/apparatus appears on the schedule of Hazardous Substances, issued by the Directorate: Radiation Control of the Department of Health, a license in terms of the Act on Hazardous Substances (Act 15/1973) must be submitted with the bid document. The license must be registered under the bidder's name or the letter of Joint Venture must be submitted by the License holder where the license is not in the name of the bidder.
- 2.2.4 If more than one item of equipment is offered, bidders must submit the Radiation Control License for each item of equipment that is offered in the bid. The make, model and license number of the various items offered in the bid must be highlighted on the Radiation Control License.
- 2.2.5 The Technician(s) must be the original equipment manufacturer trained to deal with the service, repair and calibration of the equipment offered in the bid. NB: Proof of original equipment manufacturer training must be submitted with the bid offer.

2.3. COMPLIANCE WITH SPECIFICATION

- 2.3.1 Offers must comply strictly with the specification, offers exceeding specification requirements will be deemed to comply with the specification.
- 2.3.2 The quality of services must not be less than what is specified.

2.4. EQUAL BIDS

- 2.4.1 If two or more tenderers score an equal total number of points, the contract must be awarded to the tenderer that scored the highest points for Specific Goals.
- 2.4.2 If capacity to deliver is part of the evaluation process and two or more tenderers score equal total points and equal preference points, the contract must be awarded to the tenderer that scored the highest points for functionality.
- 2.4.3 If two or more tenderers score equal total points in all respects, the award must be decided by the drawing of lots.

2.5. LATE BIDS

- 2.5.1 Bids are permissible to be submitted prior to closing date and time this is to avoid unfortunate or unplanned circumstances that could prevent the bidder from arriving on time during the closing date. If the bidder fails to arrive on time the department will not be held liable, to accept late bids.
- 2.5.2 Bids are late if they are received at the address indicated in the bid documents after the closing date and time.

2.6 MORE THAN ONE OFFER/ COUNTEROFFERS

- 2.6.1 Should the bidder make more than one offer, where applicable, against any individual item, such offer/s must be detailed in the Schedule of Additional Offer/s. The Department reserves its rights in and to the consideration of any additional offer/s subject to compliance with specification and the bidding conditions.
- 2.6.2 Bidders' attention is drawn to the fact that counter offers with regard to any of the abovementioned Special Conditions of Contract will invalidate such bids.

2.7. ONLY ONE OFFER RECEIVED

- 2.7.1 Where only 1 offer is received, the Department of Health will determine whether the price is fair and reasonable. Proof of reasonableness will be determined as follows:
 - (i) Comparison with prices, after discounts, to the bidder's other normal clients and the relative discount that the State enjoys;
 - (ii) Where this is not possible, profit before tax based on a full statement of relevant costs; and
 - (iii) In all cases, comparison with previous bid prices where these are available.

2.8 AWARD OF BID (S)

- 2.8.1 The Department of Health Bid Adjudication Committee reserves the right to award the bid to more than one bidder, provided the respective bidder's offer complies with the specification and meets all the conditions attached to the bid.
- 2.8.2 Once the evaluation process is complete there will be a recommendation report by the Department of Health Bid Evaluation Committee (DBEC) to the Department of Health Bid Adjudication Committee (DBAC) who has the authority to either (approve) or (not approve) the recommendation/s and appointment/s.
- 2.8.4 Notification of the intention to award the bid shall be in the same media that the bid was advertised, unless there is another directive from National Treasury to publish on other platforms.

2.8.5 "A bidder aggrieved by a decision of the Departmental Bid Adjudication Committee or Accounting Officer or delegated official may appeal to the BID APPEAL TRIBUNAL (BAT).

BAT finds its establishment in the Treasury Regulation 16A9.3 and Section 18(1) of the KwaZulu-Natal Supply Chain Management Policy Framework. Treasury Regulation 16A9.3 empowers National and Provincial Treasury to establish a mechanism to consider complaints and make recommendations for remedial actions to be taken for the non-compliance with the norms and standards. Section 18(1) of the KZN SCM Policy Framework empowers the MEC for Finance to establish an independent and impartial Bid Appeals Tribunal. In line with Paragraph 19 of the KZN SCM Policy Framework of 2006 the following procedure must be followed to lodge an appeal:

The bidder must, within five working days of receipt of the notification of an award, deliver written notification of an intention to appeal.

The bidder may, together with the notification of intention to appeal under paragraph (2) of the KZN SCM Policy Framework, deliver a request for written reasons for the award of the said bid.

The address provided for the lodging of appeals is:

Email: Batsecretariat@kzntreasury.gov.za

The Chairperson

Bid Appeals Tribunal

Private Bag X9082

Pietermaritzburg, 3200

2.9 EMPLOYEES TRADING WITH THE ORGANS OF THE STATE

- 2.9.1 The Public Service Act 103 of 1994 indicates in section 30(1) that "No employee shall perform or engage himself or herself to perform remunerative work outside his or her employment in the relevant department, except with the written permission of the executive authority of the department."
- 2.9.2 Furthermore, in terms of the Public Service Regulations paragraph 13I, "An employee shall not conduct business with any organ of state or be a director of a public or private company conducting business with an organ of state, unless such employee is in an official capacity as a director of a company listed in schedule 2 and 3 of the Public Finance Management Act"
- 2.9.3 If a bidder is found to be employed by the state, through the verification via acceptable means such as CSD, DPSA verification etc, the bid will be immediately disqualified.
- 2.9.4 If it is discovered through other Computer Assisted Audit Techniques (CAATS), that the bidder is employed by the state, the award will be withdrawn or contract may be terminated without notice.

2.10 TRUST, CONSORTIUM OR JOINT VENTURE

- 2.10.1 In terms of the Preferential Procurement Policy Framework Act and Regulations, as amended, a Trust, Consortium or Joint Venture must submit a consolidated Status Level Verification Certificate for every separate bid.
- 2.10.2 A separate B-BBEE Certificate must be submitted by each company participating in the Trust, Consortium or Joint Venture.
- 2.10.3 The non-submission of a B-BBEE Certificate by a Trust, Consortium or Joint Venture will result in zero (0) preference points being allocated for evaluation purposes (where applicable).

- 2.10.4 Should this bid be submitted by a Joint Venture, the Joint Venture agreement must accompany the bid document.
- 2.10.5 The Joint Venture agreement must clearly specify the percentage of the contract to be undertaken by each company participating therein.
- 2.10.6 The Joint Venture/Consortium must submit a formal agreement that outlines the roles and responsibilities of each member of the Joint Venture/ Consortium, nomination of an authorised person to represent the Joint Venture or Consortium in all matters relating to this bid and the details of the bank account for payments to be affected.
- 2.10.7 No award will be made to a Trust/ Joint Venture/ Consortium that is not tax compliant at the finalisation of the award.
- 2.10.8 For verification purposes, each party must submit separate proof of TCS/ PIN / CSD number.

2.11 VALIDITY PERIOD OF BID AND EXTENSION THEREOF

2.11.1 The validity (binding) period for the bid will be 180 days from close of bid. However, circumstances may arise whereby the department may request bidders to extend the validity (binding) period. Should this occur, the department will request bidders to extend the validity (binding) period under the same terms and conditions as originally offered for by bidders? This request will be done before the expiry of the original validity (binding) period. Should the Department forward a formal request for extension of validity period and the bidder opts not to respond, the department will assume that the extension of the validity period is accepted without any conditions.

2.12. CHANGE OF ADDRESS

2.12.1 Bidders must advise the Department of Health's Central Supply Chain Management Unit, Contract Section, should their ownership and/or address (domicilium citandi et executandi) details change from the time of bidding to the expiry of the contract.

2.13. INVOICES AND PAYMENTS

- 2.13.1 All invoices must be submitted in the original format.
- 2.13.2 All invoices submitted by the Contractor must contain the word "INVOICE" for non-VAT vendors or "TAX INVOICE" for VAT vendors only. VAT number must be reflected for VAT vendors.
- 2.13.3 A tax invoice shall be in the currency of the republic of South Africa and shall contain the following particulars:
 - (a) The name, address and registration number of the supplier;
 - (b) The name and address of the recipient:
 - (c) An individual serialized number and the date upon which the tax invoice is issued;
 - (d) A description of the goods or services supplied;
 - (e) The quantity or volume of the goods or services supplied
 - (f) The value of the supply, the amount of tax charged and the consideration for the supply; or
 - (g) Where the amount of tax charged is calculated by applying the tax fraction to the consideration, the consideration for the supply and either the amount of the tax charged, or a statement that it includes a charge in respect of the tax and the rate at which the tax was charged.
- 2.13.4 A Contractor shall be paid by the institution concerned, in accordance with supplies delivered and services rendered. The goods must be accepted and signed off by the relevant delegated official.
- 2.13.5 Should a Contractor indicate a special discount on his/her account provided payment is made within a certain time, every effort shall be made to take advantage of such discount. Where discounts or rebates received by the Department, the Contractor to provide credit note.

- 2.13.6 Any query concerning the non-payment of accounts must be directed to the institution concerned. The following protocol will apply if accounts are queried:
 - (i) Contact must be made with the officer-in-charge of Logistics and Accounts Payable;
 - (ii) If there is no response from Logistics and Accounts Payable, the Director Logistics and the Director: Expenditure Management of the institution must be contacted.
 - (iii) Failing all of the above, the Contractor must contact the Chief Director: Accounting Services supplying the following details:
 - 1. Name/s of person/s contacted at the Institution and dates; and
 - 2. Details of outstanding account.
 - 3. The Chief Director: Accounting Services will then take the appropriate action.
- 2.13.7 The Institutions shall not be responsible for payment of any statutory increases in tariffs or imports or any fluctuations in foreign exchange rate for any item required Contractor, to realise its obligations in terms of this Contract. The rate of exchange, as agreed upon in this Contract is subject to review if stipulated within this contract and as agreed consented by both Parties.

2.14 VALUE ADDED TAX (VAT)

- 2.14.1 All bid prices must be inclusive of all applicable taxes.
- 2.14.2 Bidders who make taxable supplies in excess of R1 million in any 12-month consecutive period are liable for compulsory VAT registration, but an entity may also choose to register voluntarily provided that the minimum threshold of R50 000 (as of 1 March 2010) has been exceeded in the past 12 month period. Bidders who meet the above requirement must register as VAT vendors, if successful, as soon as possible to avoid penalties from SARS.
- 2.14.3 <u>VAT will not be included</u> after an award of the bid or during contract management period. It is the responsibility of every bidder to correctly forecast whether they will require to register for VAT during the life of this contract based on the proposed bid amount.

2.15 COMPLIANCE WITH TAX REQUIREMENTS

- 2.15.1 It is a condition of this bid that the tax matters of the successful bidder(s) are in order, or that satisfactory arrangements have been made with South African Revenue Service (SARS) to meet the bidder's tax obligations.
- 2.15.2 The successful bidder(s) tax matters are expected to be in order during the tenure of the contract, should the bidder fail to comply with tax obligations, the orders may not be issued or the contract may be terminated.
- 2.15.3 The Tax Compliance status requirements are also applicable to potential foreign bidders / individuals who wish to submit a bid.
- 2.15.4 It is a requirement that bidders grant a written confirmation when submitting this bid response that SARS may on an ongoing basis during the tenure of the periodic contract disclose the bidder's tax compliance status and by submitting this bid such confirmation is deemed to have been granted.
- 2.15.5 Bidders are required to be registered on the CSD and National Treasury shall verify the bidder's tax compliance status through the CSD or through SARS.

2.15.6 Where Consortia / Joint Ventures / Sub-Contractors are to be involved, each party must be registered on the CSD, and their tax compliance status will be verified through the CSD or through SARS.

2.16 ENTERING OF HOSPITAL/CLINIC STORES

2.16.1 No representative from a company shall be permitted to enter the hospital/clinic premises, buildings or containers where stores are kept unless he/she is accompanied by the responsible official in charge of stores. Before entering the hospital/clinic premises, buildings or containers where stores are kept, the company representative must in writing, motivate why entry is necessary and written authority must be obtained to enter from the Head of the Institution or delegated official.

2.17 DEPARTMENTAL PROPERTY IN POSSESSION OF A CONTRACTOR

- 2.17.1 The Department's property supplied to a Contractor for the execution of a contract remains the property of the Department and shall at all times be available for inspection by the Department or its representatives. Any such property in the possession of the Contractor on the completion of the contract shall, at the Contractor's expense, be returned to the Department forthwith.
- 2.17.2 The Contractor shall be responsible at all times for any loss or damages to the Department's property in his possession and, if required, he shall furnish such security for the payment of any such loss or damages as the Department may require.

2.18 IRREGULARITIES

2.18.1 Companies are encouraged to advise the Department of Health timeously of any possible irregularities which might come to their notice in connection with this or other contracts.

2.19 UNSATISFACTORY PERFORMANCE

- 2.19.1 In amplification of paragraph 21; 22 and 23 of the GCC, unsatisfactory performance occurs when performance is not in accordance with the contract conditions.
 - (i). The institution shall warn the Contractor by registered/certified mail or email that action will be taken in accordance with the contract conditions unless the Contractor complies with the contract conditions and delivers satisfactory supplies or services within a specified reasonable time (7 days minimum). If the Contractor does not perform satisfactorily despite the warning the institution will:

Take necessary and appropriate action such as termination of contract in terms of its delegated powers.

(ii) When correspondence is addressed to the Contractor, reference will be made to the contract number/item number/s and an explanation of the complaint.

2.20 RESTRICTION OF BIDDING

2.20.1 The Accounting Officer or his/her delegate must:

Notify the supplier and any other person of the intention to restrict it doing business with Department by registered mail or email. The letter of restriction must provide for:

The grounds for restriction:

- i. The period of restriction which must not exceed 10 years; of 14 calendar days for the supplier to provide reasons why the restriction should not be imposed.
- ii. The identity number of individuals and the registration number of the entity; and
- iii. The period of restriction.

- b) National Treasury will load the details on the Database of Prohibited Vendors.
- c) The restriction period applicable will be based on the value of award/s made to the supplier over a financial year. The table below illustrates the restriction period that will be applicable per the award threshold:

2.21 CONTRACTOR'S LIABILITY

- 2.21.1 In the event of the contract being cancelled by the Department in the exercise of its rights in terms of these conditions, the Contractor shall be liable to pay to the Department any losses sustained and/or additional costs or expenditure incurred as a result of such cancellation, and the Department shall have the right to recover such losses, damages or additional costs by means of set-off from moneys due or which may become due in terms of the contract or any other contract or from guarantee provided for the due fulfilment of the contract and, until such time as the amount of such losses, damages or additional costs have been determined, to retain such moneys or guarantee or any deposit as security for any loss which the Department may suffer or may have suffered.
- 2.21.2 The Contractor may be held responsible for any consequential damages and loss sustained which may be caused by any defect, latent or otherwise, in supply or service rendered or if the goods or service as a result of such defect, latent or otherwise, does not conform to any condition or requirement of the contract.

2.22 RIGHTS TO PROCURE OUTSIDE THE CONTRACT

- 2.22.1 The Department reserves the right to procure goods outside the contract in cases of urgency or emergency or if the quantities are too small to justify delivery costs, or if the goods are obtainable from another organ of State or if the Contractor's point of supply is not situated at or near the place where the goods are required or if the Contractor's goods are not readily available.
- 2.22.2 No provision in a contract shall be deemed to prohibit the obtaining of goods or services from a Department or local authority.
- 2.22.3 If contracted item/s become available from National Treasury transversal contract, the Department reserve a right to cancel the contract with a winning bidder by giving thirty (30) days' notice. If it in the advantage and interest of the department to participate on transversal contract.

2.23 PATENTS

2.23.1 The Contractor shall pay all royalties and expenses and be liable for all claims in respect of the use of patent rights, trademarks or other protected rights, and hereby indemnifies the Department against any claims arising there from.

2.24 WAIVER

- 2.24.1 The granting by any party of any indulgence or postponement shall not be a waiver of its rights arising from this contract to demand full and specific performance of the contract.
- 2.24.2 No favour, delay or relaxation or indulgence on the part of any party in exercising any power or right conferred on each party in terms of this contract shall operate as a waiver of such power or right nor preclude any other or further exercises thereof or the exercise of any other power or right under this contract.

2.25 SUSPENSION

2.25.1 The Department may temporarily suspend whole or part of the supplied goods by providing no less than 5 days written notice to the Contractor, who shall on receipt of such written notice immediately cease the supply the goods. The Department will indicate the date on which the contract will be resumed in the aforementioned notice. No suspension shall exceed a total of 90 days unless otherwise agreed to by the parties in writing.

- 2.25.2 When the supply of the goods is suspended, the Contractor shall be entitled to pro-rata payment for the goods already delivered and reimbursement of all costs incidental to the prompt and orderly suspension of the contract.
- 2.25.3 Suspension of the contract shall not prejudice or affect the accrued rights and liabilities of the parties as at the date of suspension.

2.26 BREACH

- 2.26.1 Any termination notice referred to in GCC paragraph 23.1 shall be preceded by written notice requiring the defaulting party to remedy a breach of this contract within 14 days of the date of receipt of the notice.
- 2.26.2 If the defaulting party fails to remedy the breach within the 14 days, the aggrieved party shall be entitled without notice, in addition to any other remedy available to them at law or under this contract:
- 2.26.3 To claim specific performance of any whether or not the due date for performance has arrived; or
- 2.26.4 To terminate this contract in accordance with paragraph 23.1 of the GCC, against the defaulting party, in either event without prejudice to the aggrieved party's rights to claim damages.
- 2.26.5 The Contractor shall immediately advise the Department of the same, upon which the Department shall, in its sole and absolute discretion, decide whether to proceed with this contract or to terminate forthwith. Failure by the Contractor to advise the Department of a conflict of interest shall amount to a material breach of this contract.
- 2.26.6 A Party shall be deemed to be in breach of this Contract should the Party fail to comply with any material provisions of this Contract.
- 2.26.7 The aggrieved Party shall be obliged to first attempt to settle the matter by way of consultation with the defaulting Party. If the consultation fails, then the aggrieved Party shall promptly give the defaulting Party fourteen (14) days written notice to remedy the breach. If the defaulting Party fails to comply with such notice, the aggrieved Party may, without prejudice to any other's right at law:
- 2.26.7.1 Cancel this Contract in the event the defaulting Party committed a material breach.
- 2.26.7.2 Claim specific performance by the defaulting Party if such is a competent remedy in the circumstance.
- 2.26.7.3 Claim damages suffered, as limited under this Contract.

2.27 PREFERENCES

- 2.27.1 Should the Contractor apply for preferences in the submission of his bid, and it is found at a later stage that these applications were incorrect or made under false pretenses, the Department may, at its own right:
 - a) Recover from the Contractor all costs, losses or damages incurred or sustained by the Department as a result of the award of the Contract; and/or
 - a) Cancel the contract and claim any damages which the Department may suffer by having to make less favourable arrangements after such cancellation.

b) The Department may impose penalties, however, only if provision therefore is made in the Special Conditions of Contract and Bid.

2.28 SEVERABILITY

2.28.1 The finding of any invalidity to any provision of the contract shall not render the whole contract a nullity. A court of law or arbitrator may sever the invalid provision and the remainder of the contract shall remain enforceable.

2.29 EXPORT LICENSES

- 2.29.1 When orders are placed for goods in respect of which an export licence from the country of origin of supplies is required, Contractor shall:
- 2.29.2 Not incur any direct or indirect costs in connection with the supply or dispatch of such supplies before they have obtained such license:
- 2.29.3 If the government of the country from which the supplies are to be exported refuses, or fails to grant such license within three months of the placing of the order, the order shall be considered to be cancelled and no liability will be accepted for any loss or expenses irrespective of the nature thereof, including loss or expenditure suffered or incurred by Contractor or any other person in respect of the production, supply, transportation or delivery of such supplies.

2.30 INSURANCE

- 2.30.1 The goods supplied under the contract shall be fully insured in a freely convertible currency against loss or damage incidental to manufacture or acquisition, transportation, storage and delivery.
- 2.30.2 Any insurance policies taken out by Contractor to cover goods delivered for a contract must be taken out with a company registered in South Africa in terms of relevant insurance and companies acts.
- 2.30.3 The Contractor must ensure that the insurance remains in force throughout the contract period.
- 2.30.4 In the event that the Department requests for such Certificate of Insurance, the Contractor shall submit such Certificate within 5 days if this was not a mandatory requirement.

2.31 GENERAL QUANTITIES AND ORDERS

- 2.31.1 No quantities are reflected in this bid as orders will be placed based on an 'as and when required" and no guarantee is given or implied as to the actual quantity/quantities which will be procured during the periodic contract.
- 2.31.2 Orders will be placed for each institution by delegated officials, the order details will reflect the facility that will be responsible for the payment to Suppliers for the office automation solutions delivered and/or services rendered.
- 2.31.3 Suppliers should note that the order(s) will be placed as and when required during the periodic contract period and delivery points will be specified by the relevant delegated officials. The instructions appearing on the official purchase order form regarding the supply, dispatch and submission of invoices must be strictly adhered to and under no circumstances should the Supplier deviate from the purchase orders issued by the delegated officials.
- 2.31.4 The Department is under no obligation to accept any quantity/quantities which is more than the ordered quantity/quantities.
- 2.31.5 The Department officials will only be allowed to order a complete printing solution, services and accessories contracted

for.

2.32 EXPANSION, VARIATION AND EXTENSION

2.32.1. Consideration for expansion, variation or extension of contract will be in line with National Treasury Instruction notes and the KZN Department of Health Policy and delegations.

2.33 CESSION OF CONTRACTOR

- 2.33.1 The Contract will be personal to the winning bidder, who shall not sub-let, assign, cede or make over the Contract or any part thereof, or any share of interest therein, to any other person without the written consent of the Department, and on such conditions as it may approve.
- 2.33.2 This sub-clause shall not apply to sub-contracts given to regular suppliers of winning bidder for materials and minor components relating to the services supplied. The Department reserves the right to require winning bidder to submit, for noting, the names of such sub-contractors to ascertain their registration on the Central Suppliers Database and they must be legal entities.

2.34 CONTRACT AMENDMENTS / VARIATIONS

- 2.34.1 In amplification of paragraph 18 of the GCC, any amendments/variations, of the Contract shall come into effect in terms of the conditions contained in on "Contract Amendments/Variations Register". This register must be signed by the duly authorized signatories of winning bidder and the Head of Department: Health or his/her delegated official.
- 2.34.2 Contracted winning bidder shall not, in performing its obligation, vary from the terms and conditions stated in this Contract whether by way of addition thereto or by way of omission therefrom, without the prior written consent from the Department (Accounting Officer/delegated official), and no claim on the part of winning bidder for any extra payments on the grounds of any alterations or extra work will be entertained.
- 2.34.3 If, after the commencement of the contract, the cost or duration of the services is altered as a result of changes in, or in additions to, any statute, regulation or by-law, or the requirements of any authority having jurisdiction over any matter in respect of the contract, then the contract price and time for completion shall be adjusted in order to reflect the impact of those changes, provided that, within 14 days of first having become aware of the change, winning bidder shall furnish the Department with a detailed justification for the adjustment to the contract price.

2.35 INTELLECTUAL PROPERTY

2.35.1 In amplification of paragraph 6 of the GCC, the intellectual property discovered or created as the direct or indirect result of this contract shall remain the property of the Department.

2.36 INSOLVENCY

- 2.36.1 In the event to winning bidder institutes insolvency proceedings or has insolvency proceedings involuntarily instituted against it, the Department may terminate this Contract immediately.
- 2.36.2 In the event of assets and monies issued to winning bidder in terms of this Contract, such assets and monies shall be excluded from the estate of winning bidder and shall be returned immediately upon clause 40.1 coming into effect.

2.37 DISPUTE RESOLUTION

2.37.1 If any dispute arises between the Department and Contractor, in connection with the Specification and deliverables, either party may give the other notice in writing of the existence of such dispute, and the same shall thereupon be referred to arbitration in South Africa by a person mutually agreed upon by both parties. The submission shall be deemed to be submission to arbitration within the meaning of the terms of the arbitration laws in force in the Republic of South Africa

2.38 DOMICILLIA CITANDI ET EXECUTANDI

For the purpose of this contract, the parties choose their respective domicillia citandi et executandi as follows:

The Department Physical and Postal Address:

Department Name	The KwaZulu-Natal Department of Health
Physical Address	Natalia Building, 330 Langalibalele Street, Pietermaritzburg, 3201
Postal Address:	Private Bag X9051, Pietermaritzburg, 3200
Telephone numbers	033 – 395 2111
Telefax:	Nil

The Contractor or Bidder Physical and Postal Address:

Bidder/ Contractor Name	
Physical Address	
Postal Address:	
Telephone numbers	
Telefax:	
Email Address	

- 2.38.1 The parties hereby choose <u>domicilium citandi et executandi</u> for all notices and processes to be given and served in pursuance hereof at their respective addresses given on the first page of this Contract. Any notice of any change in such address shall be given in writing by the parties concerned and delivered by hand or sent by registered mail to the other party, upon notification of which address so notified shall serve as the new <u>citandi et executandi</u>.
- 2.38.2 A party may at any time change that party's domicilium by notice in writing, provided that the new domicilium is in the Republic of South Africa and consists of, or includes, a physical address at which the process can be served.
- 2.38.3 Any notice to a party:

- 2.38.4 Sent by prepaid registered post in a correctly addressed envelope, to it, shall be deemed to have been received on the 7th (seventh) day after posting unless the contrary is proved);
- 2.38.5 Delivered by hand to a responsible person during ordinary business hours at the physical address chosen as its domicilium, shall be deemed to have been received on the day of delivery; or
- 2.38.6 Sent by telefax or email to its chosen telefax or email number, shall be deemed to have been received on the date of dispatch (unless the contrary is proved).

2.39 DURATION OF CONTRACT

2.39.1 Once-off

2.40. SAMPLES

- 2.40.1. Samples will not be accepted with the closing of the bid document.
- 2.40.2. A sample meeting will be arranged with selected companies whereby the companies will be invited to forward their samples on a specified date and time.
- 2.40.3. Samples must be made available for the sample meeting, failure to provide a sample will reject their bid offer.
- 2.40.4. Samples shall be supplied by the bidder at his/her own cost/risk. Samples must be packaged as per the specification. Failure to do so will render the bid invalid.
- 2.40.5. Representative samples will not be accepted.
- 2.40.6. The Department reserves the right not to return such samples and to dispose of them at its discretion.
- 2.40.7. Samples must be clearly marked: Item number:
 - i. Brand Name
 - ii. Name of the Company
 - iii. Bid number
 - iv. Name of the manufacturer/supplier
 - v. Description of item
 - vi. Date of manufacture
- 2.40.8 The award of this bid will be based on the sample submitted from a manufacturer based on a letter of undertaking, which is compliant to specification. If, during the contract, the awarded supplier wishes to change the item being supplied, the service provider shall apply to the Department in writing, giving reasons why they want to change the product being supplied, which the Department shall consider. This process will be subject to the sample being submitted to the technical committee for evaluation and if in order, to the adjudication committee for approval. This will be done via the contract management unit of the Department. If there is a change in the product being supplied, and no prior approval has been granted, the Department reserves its right to cancel the contract.
- N.B Failure to clearly mark the samples submitted shall result in the samples not being evaluated and eliminated from further consideration.

2.41 PRICE NEGOTIATION PRIOR TO THE AWARD OF BID

Should the bid price exceed reasonable and market related prices, the Department reserves a right to request cost breakdown and negotiate prices with responsive bidder/s before the award is published.

SECTION J: EVALUATION CRITERIA

This bid will be evaluated based on the Three (3) phases, should the bidder fail to comply with the requirements of this evaluation criteria it will not progress to the next or last phase of the evaluation.

Phase 1 – Administrative, Compulsory and Mandatory Requirements

No.	Document Name	Included in the	To be returned	
		published bid	by	
		document?	bidder?	
		(Yes/No)	(Yes/No)	
	Administrative and Compulsory Require	ements		
1.	Part A: Invitation To Bid (SBD 1)	Yes	Yes	
2.	Part B: Terms And Conditions For Bidding (SBD 1)	Yes	Yes	
3.	Section A: Special Instructions Regarding Completion Of Bid	Yes	Yes	
4.	Section B: Registration On Central Suppliers Database (CSD)	Yes	Yes	
5.	Section C: Declaration That Information On Central Suppliers	Yes	Yes	
6.	Section D:Official Briefing Session Form (Not Applicable)	Yes	Yes	
7.	Section E: Bidder's Disclosure (SBD 4)	Yes	Yes	
8.	Section F: The National Industrial Participation Programme (SBD 5)	Yes	Yes	
9.	Section G: Preference Points Claim Form (SBD 6.1)	Yes	Yes	
10	Section H: General Conditions of Contract (GCC)	Yes	Yes	
11	Section I: Special Conditions of Contract (SCC)	Yes	Yes	
12	Section J: Evaluation Criteria	Yes	Yes	
13	Section K: Authority To Sign A Bid	Yes	Yes	
14	Section L: Specifications	Yes	Yes	
	Mandatory Requirements	l .		
15	Consortium/ Joint Venture/ Partnership Agreement, If Applicable.	No	Yes	
	Consortium John Venture/ Farthership Agreement, if Applicable.		(If Applicable)	
16	B-BBEE certificate indicating the B-BBEE status level of contributor. The B-BBEE certificate must be issued by a SANAS accredited verification agency;		Yes	
17	Descriptive literature, colour pamphlets, colour brochures and technical data sheets applicable to the offer (refer to clause G31)		Yes, Phase 2	
18	Letter of undertaking if not the manufacturer of the Equipment, for each item as per specification	No	Yes	
19	Certified Copy of the Radiation Control License relevant to the equipment offered in terms of this bid.	No	Yes	
20	SAHPRA certification	No	Yes	

Note: Should the bidder fail to comply with the above administrative, compulsory and mandatory requirements the bidder will be disqualified

Phase 2: Technical Evaluation Criteria

The system offered must comply fully with or exceed all of the minimum specification requirements as per the Clauses as contained in the Specification. The prospective bidder is required to provide descriptive literature, colour pamphlets, colour brochures and technical data sheets applicable to the offer (i.e. supporting information for all components of the system) for the Technical Evaluation.

If the product offered is unknown to the Department, the Department reserves the right to have the unit evaluated by a team of Technical and Clinical experts with regards to its functionality, performance and quality. The decision of this committee will be used as a motivation for the evaluation and recommendation of the bid. For this reason a demonstration unit should be readily available within 14 working days, or the bidder must make arrangements for demonstration with representatives of the Department for the equipment offered at a site within South Africa where a same make and model of unit is installed and is in full clinical operation. The cost of this site visit is for the account of the bidder and it must therefore not place any obligation on the Department to procure from the bidder.

Phase 3: Price and Preference Points

The value of this bid is estimated not to exceed R 50 000 000 (all applicable taxes included) and therefore the 80/20 preference point system shall be applicable.

Points for this bid shall be awarded for:

Price; and Specific Goals

The maximum points for this bid are allocated as follows:

CATEGORY	POINTS
PRICE	80
SPECIFIC GOALS	20
Total points for Price and Specific Goals must not exceed	100

Please Note:

- 1) HDP& RDP GOALS: 10 full points are allocated to companies who are at least 51% owned by Black People in addition
 - 10 full points are allocated to companies who are at least 100% owned by Black Africans or
 - 10 full points are allocated to companies who are at least 100% owned by Black African woman or
 - 10 full points are allocated to companies who are at least 51% owned by Youth or
 - 10 full points are allocated to companies who are at least 51% owned by black people or persons living with disabilities

2) Proof to claim Specific Goals or required returnable documents are as follows:

- 1. Ownership Certificate issued by the Companies and Intellectual Property Commission (CIPC).
- 2. CSD Report Ownership information
- 3. The medical certificate or SASSA disability card or physical assessment by department's specialist doctors will serve as verification
- 4. The Department reserve a right to verify authenticity of information provided by using other Computer Assisted Technics.

3) False Declaration

The procuring institution reserve a right to verify information submitted, should the bidder submit false or fraudulent proof to claim points for specific goals, the bidder may be immediately disqualified or contract may be terminated.

4) Scoring of points

Should the responsive bidder fail to submit proof to claim points for specific goals, the bid will not be disqualified but the offer will not score points for specific goals (zero points).

SECTION K

AUTHORITY TO SIGN A BID

The bidder must indicate the enterprise status by signing the appropriate box hereunder.

(1)	(II)	(III)	(IV)	(V)	(VI)	
CLOSE CORPORATION	COMPANIES	SOLE PROPRIETOR	PARTNERSHIP	CO- OPERATIVE	JOINT VENTURE CONSORTIUM	1
					Incorporated	
					Unincorporated	

I/We, the undersigned, being the Member(s) of Cooperative/ Sole Owner (Sole Proprietor)/ Close Corporation/ Partners (Partnership
Company (Representative) or Lead Partner (Joint Venture / Consortium), in the enterprise trading as:
hereby authorise Mr/Mrs/Ms
acting in the capacity of
whose signature is
to sign all documents in connection with this bid and any contract resulting therefrom on behalf of the enterprise.

NAME	ADDRESS	SIGNATURE	DATE

(if the space provided is not enough please list all the director in the resolution letter)

Note:

The following document must be attached to this form according to the status of the enterprise, in the form of a resolution authorising the signatory to sign all documents in connection with this bid and any contract resulting therefrom on behalf of the enterprise, and <u>such resolution shall include a specimen signature of the signatory.</u>

Co-operative: Resolution letter from the directors
Close Corporation: Resolution letter from the directors
Company: Resolution letter from the director/s
Sole Proprietor: Resolution letter from the director
Partnership: Resolution letter from the director

Joint Venture / Consortium: Resolution/agreement passed/reached' signed by the authorised representatives of the

enterprises

Note: Director/s may appoint themselves if they will be the one signing all documents in connection with this bid and any contract resulting therefrom on behalf of the enterprise.

<u>Failure to complete, sign and date this form or failure to provide the certificate(s) in the form of a resolution as described above shall result in the tender being considered non-responsive and rejected.</u>

SECTION L: SPECIFICATION

PROVINCE OF KWAZULU-NATAL

DEPARTMENT OF HEALTH

HEALTH TECHNOLOGY SERVICES (H.T.S. – RADIOLOGY SERVICES)

SPECIFICATION FOR:

SUPPLY, DELIVERY, INSTALLATION, DEMONSTRATION AND COMMISSIONING 128 SLICE LARGE BORE COMPUTERISED TOMOGRAPHY SCANNER SYSTEM AT ADDINGTON HOSPITAL.

UMDNS:

SPECIFICATION: HTS No RAD 14A (ONCOLOGY)

NB: BIDDERS SHOULD NOTE THE FOLLOWING IMPORTANT INFORMATION;

BIDDERS MUST NOTE THAT THOSE GENERAL CLAUSES WHICH ARE SHADED OFF ARE COMPULSORY AND NOT OPEN FOR COMMENT.

		BIDDERS COMMENTS:
NO	SPECIFICATION	THE UNSHADED CLAUSES MUST BE COMPLETED BY THE BIDDER, FAILURE TO COMPLETE THESE CLAUSES WILL RENDER THE BID UNRESPONSIVE
Clause G1	The space provided under "Bidder's Comments" for each clause must be used for this purpose. Bidders who neglect to provide answers to every Clause in this Bid Specification will be disqualified. Bidders must note that abbreviated answers e.g. N/A etc. will not be accepted. Bidders must also note that no part of any clause/s in this Bid Specification may be altered. Where there are traces of alterations found to any clauses in this Bid Specification during Adjudication, the Adjudication Committee will reserve the right to disqualify the bidder. The Bidder must clearly indicate if their offered product complies with the stated requirements, by indicating, "Complies" or "Does not comply" or answer the question next to the corresponding clause.	
Clause G2	All responses must be clear and legible.	
Clause G3	GUARANTEE:	
Clause G3.1	All Equipment, Materials and Workmanship provided under this Contract must be Guaranteed for a minimum period of twenty four (24) Months. The successful bidder must arrange with the respective Hospital / Institution and the Health Technology Services before Commissioning the Equipment at the respective Hospital / Institution. The bidder to note that the Guarantee period must only take effect upon successful Commissioning at the respective Hospital / Institution and successful test and acceptance by the Health Technology Services.	
Clause G3.2	State percentage guaranteed up time of machine (Should be at least 99%).	
Clause G3.3	The recommended number of services, per annum, by the manufacturer, must be included during and up until the end of the guarantee period and all costs related to the provision of such service/s will be for the bidders account.	
Clause G3.4	The bidder must state the number of services that will be provided during and up to the end of the guarantee period.	
Clause G3.5	Any breakdown during the guarantee period must include all cost (spares, labour, travelling and sundries) for any prescribed maintenance services (major and minor) as well as any QA testing that is required by Department Health's Radiation Control Board during the guarantee period.	
Clause G3.6	Travelling and Travelling Time costs must be included during the Guarantee Period?	
Clause G3.7	Spares that may be required during the Guarantee Period will be supplied at the expense of the bidder.	
Clause G3.8	Downtime during the Guarantee Period must extend the Guarantee time on a Day-to-Day basis.	

NO	SPECIFICATION	BIDDERS COMMENTS: THE UNSHADED CLAUSES MUST BE COMPLETED BY THE BIDDER, FAILURE TO COMPLETE THESE CLAUSES WILL RENDER THE BID UNRESPONSIVE
Clause G3.9	Any repetition (twice or more) of the same type of fault that first occurred during the guarantee period must be considered as a repair under guarantee if it occurs within the first year after the expiry of the guarantee period.	
Clause G3.10	The same guarantee conditions must apply to replacement units.	
Clause G4	The successful bidder must Supply, Deliver, Commission and install the Equipment and will be required to demonstrate the product to the applicable Staff at the Institution and costs for the abovementioned must be included in the final bid price.	
Clause G5	Bidders must offer the Health Technology Service's In House Technicians a demonstration of the product, which will enable the Health Technology Service's In House Technicians to become acquainted with the equipment during the Test and Acceptance phase.	
Clause G6	Preference may be given to a make and model that has been technically and clinically evaluated by a Government Institution within the R.S.A. (Attach proof of evaluation where applicable).	
Clause G7	The successful bidder must provide the Health Technology Service's in house Technicians, full training in the calibration, maintenance, service and repair of the product down to PCB Level. N.B. The quality and level of the training must be equivalent to the manufacturer's original factory training and any costs incurred to provide this training will be for the bidders account. A Certificate of Competency must be issued on completion of the training. The Training must be provided by the successful bidder to the Health Technology Services within three months from date of initial supply and delivery of the equipment to the end user.	
Clause G8	SERVICING:	
Clause G8.1	The bidder must have a well established service and repair facility in KwaZulu-Natal, to service, repair and calibrate the equipment offered. (The Health Technology Services reserves the right to inspect the premises).	
Clause G8.2	If the service is subcontracted to a local service agent, a signed copy of the letter of appointment by the bidder and acceptance by the subcontractor must be submitted with this bid / quotation. (The Health Technology Services reserves the right to inspect the premises).	
Clause G8.3	State Number of other medical equipment "Repair & Service" Agencies (excluding your Agency) represented by the subcontractor.	
Clause G8.4	Supply the Name, Address and Telephone Number/s of the Local Service Department within KwaZulu-Natal.	

		BIDDERS COMMENTS:
NO	SPECIFICATION	THE UNSHADED CLAUSES MUST BE COMPLETED BY THE BIDDER, FAILURE TO COMPLETE THESE CLAUSES WILL RENDER THE BID UNRESPONSIVE
	Please supply details as follows:	
	Company name :	
	Physical Address :	
	Telephone Number/s :	
	Fax number :	
	(The Health Technology Services reserves the right to inspect the	
Clause G8.5	premises).	
Clause G8.5	State if the Technician(s) are in the direct employ of the bidder or a subcontractor.	
	Subcontractor.	
Clause G8.6	The bidder must supply information on the number of Technicians	
Clause Co.o	permanently working in KwaZulu-Natal and their names and contact	
	Telephone Number/s must be listed (Directly employed or subcontracted) in	
	an annexure to the bid document.	
Clause G8.7	The Technician(s) must be original equipment manufacturer trained to deal	
	with the service, repair and calibration of the equipment quoted on.	
	N.B. Proof of original equipment manufacturer training must be	
	submitted with this bid / quotation offer.	
Clause G8.8	The Institution's requirement is that a technician is available within a	
	reasonable time (24 hours) to attend to malfunctioning equipment. The	
	Bidder to state the technician per install base e.g. equipment ratio to	
	technician ratio, e.g. 1 technician per 10 pieces of equipment.	
Clause G9	The bidder must Guarantee that no additional equipment will be required	
	for the successful operation of the equipment bided for on delivery and	
	commissioning at the customers site. A starter pack of all essential accessories and disposables must be supplied so that the unit can be put	
	into immediate operation. The cost of the starter pack must be included in	
	the final bid price.	
Clause G10	Optional accessories must be offered for separately on the Schedule of	
	optional accessories found at the end of this Technical specification,	
	indicating catalogue numbers, correct descriptions and Prices inclusive of	
	V.A.T.	
Clause G11	Bidder must state the period of time for delivery of Spare parts following	
	the receipt of an official order as follows: 0 to 10 days; 0 to 20 days; 0 to 30	
	days; 0 to 60 days; 0 to 90 days; more than 90 days.	
Clause G11.1	The Bidder must supply with this offer a list together with the quantities of	
	spares held locally in stock in the KwaZulu-Natal Province on the offered	
	product. The Health Technology Services reserves the right to inspect the	
	premises to verify the spares stock held.	
Clause G12	The bidder must include a firm commitment in writing, which must be	
	attached with this bid that they would supply spares, components,	
	upgrades, complete original service / repair manual, technical support and	

NO	SPECIFICATION	BIDDERS COMMENTS: THE UNSHADED CLAUSES MUST BE COMPLETED BY THE BIDDER, FAILURE TO COMPLETE THESE CLAUSES WILL RENDER THE BID UNRESPONSIVE
	ongoing training support for technical staff of the Health Technology Services and the end users Department of Health, KwaZulu-Natal throughout the life cycle of the equipment offered.	
Clause G13	Spares must be available for 10 (Ten) years from the original equipment manufacturer for the product offered.	
Clause G14	The successful bidder must include in their offer at no extra cost to the final bid price:	
Clause G14.1	Complete user Operation / Maintenance Manual x 2 (two) Book / File; CD; DVD copies in English Language.	
Clause G14.2	Complete ORIGINAL Service / Repair Manual x 2 (two) Book / File; CD; DVD copies in English Language which MUST include the following information: Fault Finding Guide, Circuit Diagrams / Schematics, Circuit Descriptions, and PCB Layouts, Calibration Guide, Part Numbers and exploded diagram of Mechanical Parts / Panels.	
Clause G14.3	All the above Manuals must be properly bound in either a Book, File or CD form.	
Clause G14.4	The Bidder must supply all software (including software-keys and / or passwords) to allow for trouble shooting (faultfinding), maintenance, calibrations, repairs and services at no additional cost.	
Clause G15	Does your Company have an after hour service back up facility.	
Clause G16	If the equipment is taken away for repairs, a loan set must be made available on request to the end user by the Institution until the Institution's unit is returned. All costs incurred for providing the loan unit must be for the bidders account.	
Clause G17	Bidder must bid on the latest model and Technology that fully complies with this Technical Specification.	
Clause G17.1	The Bidder must state how long this technology has been commercially available (state when the model offered was launched).	
Clause G17.2	The bidder must state if there are any near future updates expected.	
Clause G18	The successful bidder must maintain a system for notifying and providing	
Clause G19	users with Updates, Modifications, new Software Releases and Recalls. The successful bidders must arrange for an acceptance test of the	
	equipment with the Manager of the Health Technology Services and the	
	Hospital Manager. A copy of the original answered Specification, copy of the invoice order and relevant paperwork (PH form) from the receiving	
	Hospital must be submitted with the equipment when the ACCEPTANCE	
	TEST is to be undertaken.	
Clause G20	Where equipment bided for, operates off 220 Volt, 50Hz a.c. supply, bidder must ensure that the product being quoted for is fitted with a 15	

		BIDDERS COMMENTS: THE UNSHADED
NO	SPECIFICATION	CLAUSES MUST BE COMPLETED BY THE BIDDER, FAILURE TO COMPLETE THESE CLAUSES WILL RENDER THE BID UNRESPONSIVE
	Amp approved mains plug top, which is held together by two screws.	
	The unit must comply with an acceptable International Electrical Safety Standard such as IEC 60601-1 and 60601-1-2 for Medical Equipment where the quoted equipment operates off an electrical supply.	
	All equipment, the installation and any alteration / additions must	
	comply with:	
Clause G22.1	The Occupational Health and Safety Act (1993);	
Clause G22.2	The wiring code S.A.N.S. 0142.	
	Units being quoted for must be CE Certified. (Attach a copy of certification). The make and the model offered must be reflected on the certificate.	
	The Mains Cable of the unit being quoted for must be the Hospital Grade Type and it must be a minimum length of (3) three metres. N.B. The mains cable of the unit being quoted for must be S.A.N.S. colour coded.	
	The equipment being quoted for must be protected against Electro magnetic	
	Interference.	
	Only new equipment must be quoted for. Refurbished and reconditioned equipment being quoted on will not be accepted.	
	Bidders must note that dedicated test equipment, spare parts and any special tooling required for the upkeep and maintenance of the equipment quoted on must be available to the Health Technology Services to procure if requested.	
	All the necessary calibration and maintenance software, where applicable, required to maintain and calibrate the equipment, must be supplied with the equipment to the Health Technology Services at no extra cost to the final bid price.	
	NB. HAZARDOUS SUBSTANCE ACT:	
	If this type of equipment / apparatus appears on the schedule of Hazardous Substances issued by the Directorate: Health Technology of the Department of Health, a license in terms of the Act on Hazardous Substances (Act. 15/1973) must be submitted with this bid document. The license must be registered under the bidders name or a letter of joint venture must be submitted by the license holder where the license is not in the name of the bidder. Bidders that neglect to submit a license will not be considered.	
	Bidder must state the Radiation Control licence number of the make and model of equipment offered.	License No:

NO	SPECIFICATION	BIDDERS COMMENTS: THE UNSHADED CLAUSES MUST BE COMPLETED BY THE BIDDER, FAILURE TO COMPLETE THESE CLAUSES WILL RENDER THE BID UNRESPONSIVE
Clause G29.3	Where it has been established by the bidder that the equipment offered does not require Radiation Control licence, proof from the Radiation Control authority must be submitted with this bid document.	
Clause G30	The system offered must comply fully with or exceed all of the minimum specification requirements per the Technical Clauses.	
Clause G31	The offer submitted must be supported by descriptive literature, colour pamphlets, colour brochures and technical data sheets applicable to the offer (i.e. supporting information for all components of the system) must accompany the bid, failing which the bid will not be considered.	
Clause G32	The equipment and any accessories ordered from the successful bidder will be delivered, installed, tested, calibrated, demonstrated (including specified training) and commissioned in the specific Hospital at the expense of the successful Bidder, prior to full payment being made.	
Clause G33	All prices are to include V.A.T. and must be quoted in the South African currency. The price must be valid for a period of 180 days from closing date of bid.	
Clause G34	If the product offered is unknown to the Department, the Department reserves the right to have the unit evaluated by a team of Technical and Clinical experts with regards to its functionality, performance and quality. The decision of this committee will be used as a motivation for the evaluation and recommendation of the bid. For this reason a demonstration unit must be readily available within14 days, or the bidder must take arrange for demonstration with representatives of the Department for the equipment offered at a site within South Africa where a same make and model of unit is installed and is in full clinical operation. The cost of this site visit is for the account of the bidder and it must therefore not place any obligation on the Department to procure from the bidder.	
Clause G35	The Institution requesting the unit reserves the right to clinically trial and evaluate the unit in order to ensure that the unit meets the clinical requirements of the Department before adjudication of the bid.	
Clause G36	UPGRADEABILITY WHERE APPLICABLE:	
Clause G36.1	Bidders are to state the policy with regard to future software updates and the costs that will be involved.	
Clause G36.2	The Bidder to state what hardware and software will be available, with costs and projected dates.	
Clause G37	UPGRADE POLICY:	
Clause G37.1	All future upgrades (hardware and software) involving <u>patient safety</u> must be offered at no additional cost.	
Clause G37.2	All future upgrades removing software viruses from existing software must be supplied at no cost.	
Clause G37.3	Any upgrade before or after installation of the equipment involving additional cost must be brought to the attention of the Manager, Health Technology Services.	
Clause G38	The Bidder must indicate the expected life of their offered unit and software in years.	

NO	SPECIFICATION	BIDDERS COMMENTS: THE UNSHADED CLAUSES MUST BE COMPLETED BY THE BIDDER, FAILURE TO COMPLETE THESE CLAUSES WILL RENDER THE BID UNRESPONSIVE
Clause G39	The Bidder must provide a detailed breakdown of the cost of ownership of their offered system for the life cycle including cost of services, disposables etc. The following formula must be used: Cost of Ownership = Unit Price + Installation / Commissioning costs + Training costs (End User & Technical) + Comprehensive Maintenance / QA checks per year (Nett Present Value) X Life expectance in years. The cost of Ownership may be used as part of the feasibility evaluation of bid.	
Clause G40	Registered product with SAHPRA (South African Health Products Regulatory Authority) at time of tender. Failure to submit confirmation will result to disqualification. Please state SAHPRA Licence number to distribute the product	License No:

<u>SPECIFICATIONS FOR A LARGE BORE 128 SLICE CT SCANNER</u>

Dedicated CT Simulator is required for Radiotherapy Department for advanced Simulation for planning, conventional, 3-D CRT, IMRT, IGRT Planning. The CT-Simulator is required for most accurate simulation, placement of treatment fields and marking of radiation field portals on the patient's skin for radiation therapy of cancer patients. The CT scanner should be a spiral, multi-slice should ensure easy, error free and total compatibility between the scanner and simulation workstation. If third party software is supplied, it will be sole responsibility of the vendor incorporating latest technology available in the market. The Simulation software should be user friendly and supplying the CT simulator to run the software. The system should be able to integrate the Virtual simulation Software, workstation to existing and future 2 D TPS, 3D TPS, and Linear accelerator machine of the department and this will be entirely and directly responsibility of the vendor.

CLAUSE 1: GANTRY

1.1 The system must be a 128-slice with a 128 x 0.6mm slices or more

BIDDER'S COMMENTS:

1.2 The unit must be low voltage slipring designed gantry technology.

BIDDER'S COMMENTS:

1.3 The unit must support whole body CT Scanning including whole body.

BIDDER'S COMMENTS:

1.4	The unit must have a minimum scan time of 0.35 second for sub-second scanning.
BIDDE	ER'S COMMENTS:
1.5	The unit must have a gantry aperture size of at least 80cm or larger.
BIDDE	ER'S COMMENTS:
1.6	Angulations of at least + 30 degrees and - 30 degrees.
BIDDE	ER'S COMMENTS:
1.7	The unit must have a minimum field of view at least 50cm and extended field of view of 70cm.
BIDDE	ER'S COMMENTS:
1.8	The unit must have dual control panels on either side (left and right) of table on the gantry to adjust gantry movements.
DIDDE	ER'S COMMENTS:
טוטטנ	LN 3 COMMINICIA 13.
1.9	The system must have three computer-controlled moving laser lights for marking the field reference points. The system must consist of a single overhead moving laser to project the sagittal plane, two moving lasers to project the coronal plane and two moving lasers to project the axial plane (Accuracy of 1mm).
BIDDE	ER'S COMMENTS:
1.11	Control of gantry movements from operator console must be possible.
BIDDE	ER'S COMMENTS:
1.11.1	The system shall have an intercom between gantry and operator console.
BIDDE	ER'S COMMENTS:
	46

1.11.	2 Electronic patient breathing instructions in multiple languages must be possible.
BIDD	PER'S COMMENTS:
1.12 BIDD	Emergency stop switches to be available on gantry as well as on the console. PER'S COMMENTS:
1.13 BIDD	An in-room monitor that duplicates console display information (parameters) must be supplied. PER'S COMMENTS:
CLA	JSE 2: PATIENT SUPPORT SYSTEM (PATIENT TABLE)
2.1 BIDD	Support patient weight of 180kg and above. The sag of the couch with weight should not be more than 5mm. PER'S COMMENTS:
2.2	Lowest table height should be 47.5cm or less
	PER'S COMMENTS:
2.3 BIDD	The scan range of the tabletop must be 150cm or better to facilitate whole body scanning applications. DER'S COMMENTS:
2.4 BIDD	Longitudinal tabletop travel must not be less than 150cm for whole body coverage. PER'S COMMENTS:
2.5 BIDD	Must have motorized patient support movements (longitudinal and vertical movements) PER'S COMMENTS:
2.6	Carbon fiber tabletop with indexing facilities (for all kinds of immobilization systems used in radiotherapy).

BIDDE	BIDDER'S COMMENTS:	
2.7	Control of table movements from operator console and gantry must be available.	
BIDDE	ER'S COMMENTS:	
2.8	Programming of motorized removal and re-indexing of tabletop without repositioning patient scanning for emergency access to patient must be available.	
BIDDE	ER'S COMMENTS:	
2.9	Patient positioning accessories required to perform the procedures specified, including headrest support, axial and coronal head rest support, table leg extender, security straps, infant immobilizer, arm support, knee support and immobilizing straps must be provided.	
BIDDE	ER'S COMMENTS:	
2.10	Emergency override to enable manual removal of patients in case of power failure or emergency must be included.	
BIDDE	ER'S COMMENTS:	
CI AII	SE 3: DETECTOR ACQUISITION SYSTEM	
3.1	Solid-state detector technology using low-dose and high-resolution acquisitions.	
	ER'S COMMENTS:	
וטטום	ER 5 COMMEN 15:	
3.2	Multi-slice detector technology for 128-slice acquisitions per 360-degree gantry rotation.	
BIDDE	ER'S COMMENTS:	

3.3 BIDDE	Multi-slice detector comprising at least 896 channels with 1536 views per second sampling rates or better. – ER'S COMMENTS:
3.4	The detector must be a true 64-row.
BIDDE	ER'S COMMENTS:
3.5 BIDDE	Coupled detector/ tube assembly on a slip-ring technology gantry design. ER'S COMMENTS:
3.6 3.6.1 3.6.2 3.6.3 3.6.4	All calibration and service phantoms to be included with wall brackets. Phantoms must include the following: Phantom Specifically Developed for CT Simulation and RTP with QC. Quality control accessories and phantom (CATPHAN, LASER alignment, diagnostic QC tools). Special quality control kits such as ACR phantom and AEC phantom for calibration of image detector. Special quality control software.
3.7	The detector must be able to produce a minimum slice thickness of 0.6mm or more. Bidders to state the minimum ess available. –
	ER'S COMMENTS:
CLAU	SE 4: X-RAY GENERATOR Modern high frequency generator allowing at least between 60kV to 135kV and 500mA for a 0.35 second scan time or better.
BIDDE	ER'S COMMENTS:
4.2 BIDDE	To support at least 200 scans in 300 seconds using a 0.35 second scan time or better during the spiral scan mode. ER'S COMMENTS:

4.3 BIDDE	The unit must be able to scan continuously a length of 170cm high resolution without cooling time. ER'S COMMENTS:
	The unit must support repeat scanning with minimal wait time, Bidder to state the waiting time. ER'S COMMENTS: it time – user dependant
-	
4.5 BIDDE	The unit must have at least 4 variable kV setting between 70kV (or less) and 135kV (or more). Bidder to state the variable settings.
4.6 BIDDE	The unit must have variable mA settings between 10mA to 600mA or better. ER'S COMMENTS:
4.7	State maximum mA allowed for each kV.
BIDDE	ER'S COMMENTS:
	IN 3 COMMINICIATS.
4.8	Generator Power output of at least 70kW or higher.
CLAU	SE 5: X-RAY TUBE
5.1	High speed rotating anode tube with dual focal spot technology. Bidder to state the dual focus spot size.
BIDDE	ER'S COMMENTS:
5.2 BIDDE	Anode heat dissipation rate not less than 1300kHU/minute. ER'S COMMENTS:

5.3	Cooling performance curves required.
	ER'S COMMENTS:
5.4	The unit must have a heat storage capacity of at least 7MHU with at least 80% full loading.
BIDD	ER'S COMMENTS:
<i>-</i> -	Every stand to be life. Otate in years and alies as assess
5.5	Expected tube life. State in years and slices or scans.
RIDD	ER'S COMMENTS:
5.6	Guaranteed tube life ≥ 200.000 rotation.
BIDD	ER'S COMMENTS:
CLAU	JSE 6: OPERATOR CONSOLE
6.1	Modern user interface with logical and intuitive operation.
BIDD	ER'S COMMENTS:
6.2	Operator console with at least 19-inch high-resolution flat panel color display monitors with 1024 x 1024 display
0.2	resolution or more.
BIDD	ER'S COMMENTS:
6.3	One operator console and two independent post-processing workstations must be included in the bid. The operator console to include a high-speed computer which will support multi-tasking operation to allow simultaneous scanning reconstruction, viewing, archiving, processing and filming operation.
BIDD	ER'S COMMENTS:

6.4	DICOM and PACS compatible storage of patient images.
BIDDE	R'S COMMENTS:
6.5	Intercom to gantry for patient instructions.
BIDDE	R'S COMMENTS:
6.6	Emergency stop switch for patient safety.
BIDDE	R'S COMMENTS:
6.7	The console must have comprehensive adult and pediatric anatomical programs/protocols.
BIDDE	R'S COMMENTS:
6.8	Must have at least 50 pre-programmed scan protocols required with option to add user defined protocols.
BIDDE	R'S COMMENTS:
6.9	Electronic patient breathing instructions in multiple languages.
BIDDE	R'S COMMENTS:
	IX O COMMENTO.
6.10	Image review system must be able to display, analyze, process and manage acquired imaging data such as: - Zoom
	- Grey scale (colour display) adjustment
	- Measurement of distance and angle- Rotation
	- Grid
	- Pan
	- Annotation
	- Cine display
	Facilities must be available to perform basic CT examination of all parts of the body, including reconstruction, review and printing.
	with britishing.

BIDDER'S COMMENTS:

6.11 BIDDE	Post processing software should include organ contouring, organ templates for reporting, contouring in 2D, 3D and oblique plane must be available on all workstations. ER'S COMMENTS:
6.12	Reconstruction algorithm must be specified.
BIDDE	ER'S COMMENTS:
6.13 BIDDE	Reconstruction Algorithm must be capable of Metal Artefact Reduction (Dynamic Range, Algorithm, Protocol) and Advanced artefact reduction. ER'S COMMENTS:
6.14 BIDDE	At least 512 x 512 reconstruction frame is required. ER'S COMMENTS:
6.15 BIDDE	A 15 frames / second or faster reconstruction time is required per 512 x 512 reconstructed images. ER'S COMMENTS:
6.16 BIDDE	Reconstruction fields of view must be at least 50cm. ER'S COMMENTS:
6.17	The image reconstruction time should be less than 1.5 second for any mode.
חחום	ER'S COMMENTS:

BIDDE	6.18 The CT number accuracy shall be better than 0 ± 4 HU for water and -1000 ± 10 HU for air. BIDDER'S COMMENTS:	
6.19 BIDD	Real-time reconstruction for scout view required. ER'S COMMENTS:	
6.20 BIDDE	Compatibility: To comply to DICOM 3 standards. Please provide DICOM 3 conformance statement. ER'S COMMENTS:	
6.21 BIDDE	Large image archive to store raw scan data as well as reconstructed data at least 3 Tera Bytes. ER'S COMMENTS:	
6.22 BIDDE	Archival storage by means of a DVD recording system must be provided. ER'S COMMENTS:	
6.23 BIDDE	X-Ray tube heat loading display on console is recommended. ER'S COMMENTS:	
6.24 BIDDE	Central point print facility to laser imager must be supported from all CT workstations. ER'S COMMENTS:	
6.25 BIDDE	Chairs: wheeled, height adjustable, swivel for the operators must be supplied. (Quantity 4) ER'S COMMENTS:	

6.26	Two level step (foot stool) for patients must be provided.
BIDDE	ER'S COMMENTS:
CLAU	SE 7: SOFTWARE
7.1	All high-end post processing software should include organ contouring, organ templates for reporting, contouring in 2D 3D and oblique plane must be on both the CT console and Doctors console. (The user must be able to either configuration the post processing software on CT console or Doctors console depending on workflow requirements.)
BIDDE	ER'S COMMENTS:
7.2	Software and hardware allowing full patient mark-up is required, including isocentre, field intersection. points and reference point marking.
BIDDE	ER'S COMMENTS:
7.3	The software must provide, at least Tumour and organ at risk (OAR) localisation and contouring Placement of isocentre and beams.
BIDDE	ER'S COMMENTS
7.4	Generation of Digitally Reconstructed radiographs (DRRs) must be real time.
BIDDE	ER'S COMMENTS
7.5	Virtual simulation (VSIM) plans must be exportable to the treatment planning system.
BIDDE	ER'S COMMENTS
7.6	Laser software must interface with CT simulators virtual simulation software. It must have an in-room monitor for moving lasers independently.
BIDDE	ER'S COMMENTS
7.5 BIDDE	Virtual simulation (VSIM) plans must be exportable to the treatment planning system. ER'S COMMENTS Laser software must interface with CT simulators virtual simulation software. It must have an in-room monitor for lasers independently.

	The CT Scanner must have a Respiratory Gating software with both prospective and retrospective respiratory gating module.
BIDDI	ER'S COMMENTS
CLAU	ISE 8: COMPUTER SYSTEM / RECONSTRUCTION
8.1	At least a 512 x 512 reconstruction matrix is required. Bidder to state the details of matrix sizes available.
BIDDI	ER'S COMMENTS:
8.2	Real time reconstruction for scout-view required.
BIDDI	ER'S COMMENTS:
8.3	A 0.1 second or faster reconstruction time is required per 512 x 512 reconstructed image.
BIDDI	ER'S COMMENTS:
8.4	3D re-construction must be possible.
BIDDI	ER'S COMMENTS:
CLAU	ISE 9: IMAGE STORAGE
9.1	Large image archive to store raw scan-data of at least 3TB as well as reconstructed data required. Bidder to state the size.
BIDDI	ER'S COMMENTS:
9.2 BIDDI	At least 3600 scan data sets are to be stored. Bidder to state storage size. ER'S COMMENTS:

9.3	The workstation must be able to burn images onto CD and DVD, (200 CDs must be supplied).
BIDDE	ER'S COMMENTS:
CLAU	SE 10: IMAGE VIEWING MODES
10.1	Comprehensive image acquisition, viewing, processing and filming modes required must be provided and supported by the systems.
BIDDE	ER'S COMMENTS:
10.2	3D surface and volume rendering, Maximum and Minimum Intensity projections, Curved and planar MPR reconstructions must be supported by the systems.
BIDDE	ER'S COMMENTS:
10.3	Noise reduction techniques must be available.
BIDDE	ER'S COMMENTS:
10.4	Window width.
	ER'S COMMENTS:
10.5	Window level.
BIDDE	ER'S COMMENTS:
10.6	Preset window.
BIDDE	ER'S COMMENTS:
10.7	Linear and non-linear window settings.
RINDF	ER'S COMMENTS:

10.8	Double window.
BIDDE	R'S COMMENTS:
-	
10.9	Non-linear user-programmable window.
BIDDE	R'S COMMENTS:
10.10	State CT number measurement range.
RIDDE	R'S COMMENTS:
	TO COMMENTO.
10.11	Inset scout-view display.
BIDDE	R'S COMMENTS:
10.12	ROI setting and processing.
BIDDE	R'S COMMENTS:
10.13	ROI shapes including point, rectangular, polygonal, elliptical and regular shapes.
BIDDE	R'S COMMENTS:
10.14	Mean, standard deviation, area and minimum 5 mega pixels.
שטטום	R'S COMMENTS:

CLAUSE 11: SCAN MODES

BIDDER'S COMMENTS:	
11.2 BIDDE	Single scan mode must be provided. ER'S COMMENTS:
11.3 BIDDE	Scanning with serial reconstruction must be provided. ER'S COMMENTS:
11.4	Scanning with parallel reconstruction must be provided.
BIDDE	ER'S COMMENTS:
CLAU	SE 12 : SPIRAL MODE PERFORMANCE
12.1	Spiral scan mode with 128-slice and other combination for thin 0.6mm slice acquisition being acquired per gantry rotation must be provided.
BIDDE	ER'S COMMENTS:
12.2	10 scan plans per scout view required.
BIDDE	ER'S COMMENTS:
12.3	The Unit must provide a variable pitch setting for spiral scanning mode, from 0.05 to 1.5 mm in 0.1mm increments. Bidder to state the pitch available.
BIDDE	ER'S COMMENTS:

Scout-view mode for scan planning must be provided. Scout-view scan field size should be more than 1500 mm long.

11.1

12.4	Single spiral sequence must be provided.
BIDDE	R'S COMMENTS:
12.5	Multiple spiral sequence must be provided.
BIDDE	R'S COMMENTS:
12.6	Combined conventional and spiral sequence must be provided.
RIDDE	R'S COMMENTS:
40.7	
12.7	Real time reconstruction and display of images during the spiral mode must be provided at a viewing rate of at 16 images per second or better.
BIDDE	R'S COMMENTS:
12.8	For accurate contrast examinations, a bolus-timing / tracking package must be included.
RIDDE	R'S COMMENTS:
12.9	leatronia valuma imagina ia raquirad. Stata vaval sizas usad far isatronia valuma imagina
	Isotropic volume imaging is required. State voxel sizes used for isotropic volume imaging.
BIDDE	R'S COMMENTS:
12.10	Standard mode spatial resolution of at least 15 lp/cm at cut-off or better is required. Bidder to state high contrast spatial resolution in line pairs per cm.
DID = -	
RINDF	R'S COMMENTS:
	90

12.11 state h	High-resolution mode spatial resolution of at least 15 lp/cm for axial and spiral scan at 0% MTF with full FOV. Bidder to ligh contrast spatial resolution in line pairs per cm.
BIDDE	ER'S COMMENTS:
12.12	At most 4.0mm at 0.3% low-contrast resolution must be possible at 90mAs, 10mm slice thickness, 120kV. Bidder to state resolution and scan technique used to obtain measurement.
BIDDE	ER'S COMMENTS:
12.13	International CATPHAN performance phantom must be provided for measurements. Bidder to state the details used for measurements.
BIDDE	ER'S COMMENTS:
CLAU	SE 13: DOSE REDUCTION TECHNIQUE
13.1	Applied dose for multi-slice CT applications is required to be kept to the absolute minimum for all applications protocols.
BIDDE	ER'S COMMENTS:
13.2 BIDDE	Bidders to detail all dose reduction techniques used during patient scanning to reduce the total applied patient dose. ER'S COMMENTS:
13.3 BIDDE	The applied patient dose must be quantified and displayed on the operator's console for the patient examination.
13.4	Bidders must include current dose reduction technology specifically to the following: - Gantry design.

- Choice of scan-times available.
- Detector design
- Slice profile used
- Real-time exposure control during scanning
- Variable pitch settings
- Beam collimation and X-ray tube filtration
- Adult and paediatric scanning exposure techniques/protocols.
- Carbon fibre patient support tabletop atténuation
- Accurate contrast timing.
- Real-time spiral scanning display
- Calculated dose display on operator's console monitor.
- Software image quality correction that allows reconstructed images to be improved without having to rescan the patient with additional scans or scans with a higher applied dose.
- For shoulders and pelvis with streak reduction techniques
- Posterior fossa with image fusion techniques.

BIDDE	- Abdomen with motion detection techniques. ER'S COMMENTS:
13.5 BIDDE	Any other techniques must be detailed. ER'S COMMENTS:
13.6 BIDDE	Supporting documentation indicating applied dose reduction must be included. ER'S COMMENTS:
CLAU	SE 14: DICOM 3 COMPATIBILITY
14.1	The CT Scanner must support image and data linkage via network including all additional workstations and laser camera. All DICOM software must be included in the cost. It must have the ability to be easily linked to any PACS and RIS system. It must have the ability to be linked to ADSL/ISDN line or any other tele-radiology system.
BIDDE	ER'S COMMENTS:
14.2 BIDDE	All hardware and software relating to DICOM compatibility, including licenses must be supplied. ER'S COMMENTS:

BIDDER'S COMMENTS: 14.4 Links to additional workstation and laser camera system is required through DICOM compatibility. BIDDER'S COMMENTS: 14.5 DICOM Conformance statements on the various service classes per modality to be provided. BIDDER'S COMMENTS: 14.6 The 3D workstation to support linkage to CT Scanner and laser camera. BIDDER'S COMMENTS: 14.7 DICOM storage commitment, query / retrieve / MWM and MPPS DICOM software functionality must be provided to ensure PACS/ RIS/HIS network compatibility. 15.1 A WINDOWS Operating computer system is required with at least 4 GB RAM, 800 GB hard drive, network card, suitable graphics card and at least a 48cm flat panel LCD monitor (supporting at least 3 Megapixel resolution) is required. BIDDER'S COMMENTS: 15.2 Provide modern user-interface with high speed-processing and storage facilities as per operator's console. BIDDER'S COMMENTS: 15.3 2D, 3D, surface rendering, volume rendering, image sculpting, Maximum Intensity Projections, Minimum Intensity Projections, MPR and curved MPR and other image processing tools must be supported by system. BIDDER'S COMMENTS:	14.3	The Laser camera to support DICOM compatible printing on 35 x 43 cm Laser film.
BIDDER'S COMMENTS: 14.5 DICOM Conformance statements on the various service classes per modality to be provided. BIDDER'S COMMENTS: 14.6 The 3D workstation to support linkage to CT Scanner and laser camera. BIDDER'S COMMENTS: 14.7 DICOM storage commitment, query / retrieve / MWM and MPPS DICOM software functionality must be provided to ensure PACS/ RIS/ HIS network compatibility. CLAUSE 15: THREE INDEPENDENT WORKSTATIONS (Two for scanning, one for virtual simulation) 15.1 A WINDOWS Operating computer system is required with at least 4 GB RAM, 800 GB hard drive, network card, suitable graphics card and at least a 48cm flat panel LCD monitor (supporting at least 3 Megapixel resolution) is required. BIDDER'S COMMENTS: 15.2 Provide modern user-interface with high speed-processing and storage facilities as per operator's console. BIDDER'S COMMENTS:	BIDDE	ER'S COMMENTS:
BIDDER'S COMMENTS: 14.5 DICOM Conformance statements on the various service classes per modality to be provided. BIDDER'S COMMENTS: 14.6 The 3D workstation to support linkage to CT Scanner and laser camera. BIDDER'S COMMENTS: 14.7 DICOM storage commitment, query / retrieve / MWM and MPPS DICOM software functionality must be provided to ensure PACS/ RIS/ HIS network compatibility. CLAUSE 15: THREE INDEPENDENT WORKSTATIONS (Two for scanning, one for virtual simulation) 15.1 A WINDOWS Operating computer system is required with at least 4 GB RAM, 800 GB hard drive, network card, suitable graphics card and at least a 48cm flat panel LCD monitor (supporting at least 3 Megapixel resolution) is required. BIDDER'S COMMENTS: 15.2 Provide modern user-interface with high speed-processing and storage facilities as per operator's console. BIDDER'S COMMENTS:		
14.5 DICOM Conformance statements on the various service classes per modality to be provided. BIDDER'S COMMENTS: 14.6 The 3D workstation to support linkage to CT Scanner and laser camera. BIDDER'S COMMENTS: 14.7 DICOM storage commitment, query / retrieve / MWM and MPPS DICOM software functionality must be provided to ensure PACS/ RIS/ HIS network compatibility. CLAUSE 15: THREE INDEPENDENT WORKSTATIONS (Two for scanning, one for virtual simulation) 15.1 A WINDOWS Operating computer system is required with at least 4 GB RAM, 800 GB hard drive, network card, suitable graphics card and at least a 48cm flat panel LCD monitor (supporting at least 3 Megapixel resolution) is required. BIDDER'S COMMENTS: 15.2 Provide modern user-interface with high speed-processing and storage facilities as per operator's console. BIDDER'S COMMENTS:	14.4	Links to additional workstation and laser camera system is required through DICOM compatibility.
BIDDER'S COMMENTS: 14.6 The 3D workstation to support linkage to CT Scanner and laser camera. BIDDER'S COMMENTS: 14.7 DICOM storage commitment, query / retrieve / MWM and MPPS DICOM software functionality must be provided to ensure PACS/ RIS/ HIS network compatibility. CLAUSE 15: THREE INDEPENDENT WORKSTATIONS (Two for scanning, one for virtual simulation) 15.1 A WINDOWS Operating computer system is required with at least 4 GB RAM, 800 GB hard drive, network card, suitable graphics card and at least a 48cm flat panel LCD monitor (supporting at least 3 Megapixel resolution) is required. BIDDER'S COMMENTS: 15.2 Provide modern user-interface with high speed-processing and storage facilities as per operator's console. BIDDER'S COMMENTS:	BIDDE	ER'S COMMENTS:
BIDDER'S COMMENTS: 14.6 The 3D workstation to support linkage to CT Scanner and laser camera. BIDDER'S COMMENTS: 14.7 DICOM storage commitment, query / retrieve / MWM and MPPS DICOM software functionality must be provided to ensure PACS/ RIS/ HIS network compatibility. CLAUSE 15: THREE INDEPENDENT WORKSTATIONS (Two for scanning, one for virtual simulation) 15.1 A WINDOWS Operating computer system is required with at least 4 GB RAM, 800 GB hard drive, network card, suitable graphics card and at least a 48cm flat panel LCD monitor (supporting at least 3 Megapixel resolution) is required. BIDDER'S COMMENTS: 15.2 Provide modern user-interface with high speed-processing and storage facilities as per operator's console. BIDDER'S COMMENTS:		
14.6 The 3D workstation to support linkage to CT Scanner and laser camera. BIDDER'S COMMENTS: 14.7 DICOM storage commitment, query / retrieve / MWM and MPPS DICOM software functionality must be provided to ensure PACS/ RIS/ HIS network compatibility. CLAUSE 15: THREE INDEPENDENT WORKSTATIONS (Two for scanning, one for virtual simulation) 15.1 A WINDOWS Operating computer system is required with at least 4 GB RAM, 800 GB hard drive, network card, suitable graphics card and at least a 48cm flat panel LCD monitor (supporting at least 3 Megapixel resolution) is required. BIDDER'S COMMENTS: 15.2 Provide modern user-interface with high speed-processing and storage facilities as per operator's console. BIDDER'S COMMENTS:	14.5	DICOM Conformance statements on the various service classes per modality to be provided.
BIDDER'S COMMENTS: 14.7 DICOM storage commitment, query / retrieve / MWM and MPPS DICOM software functionality must be provided to ensure PACS/ RIS/ HIS network compatibility. CLAUSE 15: THREE INDEPENDENT WORKSTATIONS (Two for scanning, one for virtual simulation) 15.1 A WINDOWS Operating computer system is required with at least 4 GB RAM, 800 GB hard drive, network card, suitable graphics card and at least a 48cm flat panel LCD monitor (supporting at least 3 Megapixel resolution) is required. BIDDER'S COMMENTS: 15.2 Provide modern user-interface with high speed-processing and storage facilities as per operator's console. BIDDER'S COMMENTS: 15.3 2D, 3D, surface rendering, volume rendering, image sculpting, Maximum Intensity Projections, Minimum Intensity Projections, MPR and curved MPR and other image processing tools must be supported by system.	BIDDE	ER'S COMMENTS:
BIDDER'S COMMENTS: 14.7 DICOM storage commitment, query / retrieve / MWM and MPPS DICOM software functionality must be provided to ensure PACS/ RIS/ HIS network compatibility. CLAUSE 15: THREE INDEPENDENT WORKSTATIONS (Two for scanning, one for virtual simulation) 15.1 A WINDOWS Operating computer system is required with at least 4 GB RAM, 800 GB hard drive, network card, suitable graphics card and at least a 48cm flat panel LCD monitor (supporting at least 3 Megapixel resolution) is required. BIDDER'S COMMENTS: 15.2 Provide modern user-interface with high speed-processing and storage facilities as per operator's console. BIDDER'S COMMENTS: 15.3 2D, 3D, surface rendering, volume rendering, image sculpting, Maximum Intensity Projections, Minimum Intensity Projections, MPR and curved MPR and other image processing tools must be supported by system.		
 14.7 DICOM storage commitment, query / retrieve / MWM and MPPS DICOM software functionality must be provided to ensure PACS/ RIS/ HIS network compatibility. CLAUSE 15: THREE INDEPENDENT WORKSTATIONS (Two for scanning, one for virtual simulation) 15.1 A WINDOWS Operating computer system is required with at least 4 GB RAM, 800 GB hard drive, network card, suitable graphics card and at least a 48cm flat panel LCD monitor (supporting at least 3 Megapixel resolution) is required. BIDDER'S COMMENTS: 15.2 Provide modern user-interface with high speed-processing and storage facilities as per operator's console. BIDDER'S COMMENTS: 2D, 3D, surface rendering, volume rendering, image sculpting, Maximum Intensity Projections, Minimum Intensity Projections, MPR and curved MPR and other image processing tools must be supported by system. 	14.6	The 3D workstation to support linkage to CT Scanner and laser camera.
ensure PACS/ RIS/ HIS network compatibility. CLAUSE 15: THREE INDEPENDENT WORKSTATIONS	BIDDE	ER'S COMMENTS:
ensure PACS/ RIS/ HIS network compatibility. CLAUSE 15: THREE INDEPENDENT WORKSTATIONS (Two for scanning, one for virtual simulation) 15.1 A WINDOWS Operating computer system is required with at least 4 GB RAM, 800 GB hard drive, network card, suitable graphics card and at least a 48cm flat panel LCD monitor (supporting at least 3 Megapixel resolution) is required. BIDDER'S COMMENTS: 15.2 Provide modern user-interface with high speed-processing and storage facilities as per operator's console. BIDDER'S COMMENTS: 15.3 2D, 3D, surface rendering, volume rendering, image sculpting, Maximum Intensity Projections, Minimum Intensity Projections, MPR and curved MPR and other image processing tools must be supported by system.		
(Two for scanning, one for virtual simulation) 15.1 A WINDOWS Operating computer system is required with at least 4 GB RAM, 800 GB hard drive, network card, suitable graphics card and at least a 48cm flat panel LCD monitor (supporting at least 3 Megapixel resolution) is required. BIDDER'S COMMENTS: 15.2 Provide modern user-interface with high speed-processing and storage facilities as per operator's console. BIDDER'S COMMENTS: 15.3 2D, 3D, surface rendering, volume rendering, image sculpting, Maximum Intensity Projections, Minimum Intensity Projections, MPR and curved MPR and other image processing tools must be supported by system.	14.7	
suitable graphics card and at least a 48cm flat panel LCD monitor (supporting at least 3 Megapixel resolution) is required. BIDDER'S COMMENTS: 15.2 Provide modern user-interface with high speed-processing and storage facilities as per operator's console. BIDDER'S COMMENTS: 15.3 2D, 3D, surface rendering, volume rendering, image sculpting, Maximum Intensity Projections, Minimum Intensity Projections, MPR and curved MPR and other image processing tools must be supported by system.	CLAU	
 15.2 Provide modern user-interface with high speed-processing and storage facilities as per operator's console. BIDDER'S COMMENTS: 2D, 3D, surface rendering, volume rendering, image sculpting, Maximum Intensity Projections, Minimum Intensity Projections, MPR and curved MPR and other image processing tools must be supported by system. 	15.1	suitable graphics card and at least a 48cm flat panel LCD monitor (supporting at least 3 Megapixel resolution) is
15.3 2D, 3D, surface rendering, volume rendering, image sculpting, Maximum Intensity Projections, Minimum Intensity Projections, MPR and curved MPR and other image processing tools must be supported by system.	BIDDE	ER'S COMMENTS:
15.3 2D, 3D, surface rendering, volume rendering, image sculpting, Maximum Intensity Projections, Minimum Intensity Projections, MPR and curved MPR and other image processing tools must be supported by system.		
15.3 2D, 3D, surface rendering, volume rendering, image sculpting, Maximum Intensity Projections, Minimum Intensity Projections, MPR and curved MPR and other image processing tools must be supported by system.	15 2	Provide modern user interface with high speed processing and storage facilities as per enerator's console
 2D, 3D, surface rendering, volume rendering, image sculpting, Maximum Intensity Projections, Minimum Intensity Projections, MPR and curved MPR and other image processing tools must be supported by system. 		
Projections, MPR and curved MPR and other image processing tools must be supported by system.	וטטוט	
BIDDER'S COMMENTS:	15.3	
	BIDDE	ER'S COMMENTS:

15.4	All post-processing software must be available on all workstations.
BIDDE	ER'S COMMENTS:
15.5	Image manipulation and arithmetic must be available
BIDDE	ER'S COMMENTS:
15.6	Archival storage using TIFF, JPEG, DICOM or similar format by means of a DVD must be included, (200 discs must be included).
BIDDE	ER'S COMMENTS:
15.7	Print facility to laser imager must be supported.
וטטום	ER'S COMMENTS:
CLAU	SE 16: DICOM COMPATIBLE LASER LIGHT
	External high quality positioning laser system for Radiotherapy.
	ER'S COMMENTS
וטטום	ER 5 COMMENTS
16.2	Lasers must be green.
BIDDE	ER'S COMMENTS
16.3	3-Way independent lasers must be movable and remote controlled.
BIDDE	ER'COMMENTS

BIDDE	ER'S COMMENTS
16.5 BIDDE	Accuracy of the laser system must be at least 0.5mm. ER'S COMMENTS
16.7 BIDDE	The planes defined by the system must be axial, sagittal, and coronal. ER'S COMMENTS
16.8 BIDDE	Linkage via network to additional workstation and CT Console must be provided. R'S COMMENTS:
16.9 BIDDE	All necessary hardware, software and licenses must be included. R'S COMMENTS:
CLAUS 17.1 BIDDE	SE 17: SOFTWARE PACKAGES AND CLINICAL APPLICATIONS All packages required to perform the clinical applications as listed in SECTIONS 7,10.11 and 12. must be included. ER'S COMMENTS:
17.2 BIDDE	Any clinical application not listed in 6.2. but is offered by the manufacturer, must be quoted separately as an option. ER'S COMMENTS:

16.4

All lasers must intersect at the origin within 1mm.

CLAUSE 18: CT CONTRAST INJECTOR

18.1 A modern dual head microprocessor-controlled contrast injector is required for use with disposable syringes and will be used for all contrast enhanced examinations including cardiac and bolus chasing. (Please supply 200 units of each of the Consumables i.e. Syringes, Connecting Tubes, Y- Adaptor Tube with Dual check valve).

BIDDER'S COMMENTS:

18.2 A touch screen, multi-colored contrast injector monitor must be provided for control and timing of contrast administration.

BIDDER'S COMMENTS:

17.3 The consumables for the injector mentioned above must readily be available in South Africa.

BIDDER'S COMMENTS:

CLAUSE 19: ON-LINE UPS PROTECTION

19.1 Online UPS of 120 kVA of suitable rating should be supplied for the complete system including gantry, computer system, injector system, external laser system, all workstations and treatment with at least 30 minutes back up.

BIDDER'S COMMENTS:

CLAUSE 20: SUPPORTING DOCUMENTATION

- Sales license issued in terms of Hazardous Substance Act must be submitted.
- All supporting documentation as requested in the various sections to be included.
- All original system brochures, product specifications and application notes are to be supplied.
- DICOM 3 conformance statements to be included.

BIDDER'S COMMENTS:

CLAUSE 21: CUSTOMER SUPPORT DETAILS

- **21.1** The application training for Oncologist, Radiologists and Radiotherapists by a fully trained application specialist to include at least the following:
 - CV of trainer to be supplied with the bid.
 - An initial period of at least one week to be included patient scanning of multiple clinical problems.
 - At least 4x 1-week follow-up periods during the first 12 months.
 - Training in the specialized packages to be provided (in addition to the above training). Training schedule to be

BIDDER'S COMMENTS:		

included (to be discussed with End User).

CLAUSE 22: STANDARDS AND SAFETY

CLAUSE	STANDARDS AND SAFETY:	COMPLY	COMMENTS
		(yes/no)	
22.1	Should be FDA or CE approved product.		
22.2	Electrical safety conforms to standards for electrical safety IEC-60601 / IS-13450.		
22.3	The equipment/system must be approved and licensed by the South Africa Health Products Regulator (SAHPRA).		
22.4	A copy of a valid license issued by Radiation Control in terms of the Hazardous Substance Act, Act No 15 of 1973 must be submitted with the tender. Failure to submit such a valid license may result in a tender not being considered.		
22.5	Acceptance testing must be performed as part of the installation (failure to do will lead to failure to pay).		
22.7	Lead glass window to be installed as per room requirement.		
22.8	Safety aspects of radiation dosage leakage should be spelt out.		

CLAUSE 23: TRAINING

- 23.1 On-site training on all new equipment to be included in the pricing schedule.
- 23.2 On-site training must be undertaken to ensure the correct application of the unit. Minimum of 2 x 1-week required. Discuss with end user.
- 23.3 After equipment installation, an application specialist must demonstrate and train all staff on all aspects of the equipment
- 23.4 Follow-up training, with the application specialist, after a specified usage period must be provided at no additional cost. Adequate notification of the scheduled date(s) of this training shall be provided to ensure that all pertinent staff will be able to attend
- 23.5 The successful bidder must provide the Health Technology Service's in house Technicians, a demonstration of the product offered, **first-line** maintenance of the product.

R	וחו	DEF	פיכ	CO	M	ME	N٦	rg.
D	וטו	UEF	vo	vu	πVII	VI C	N	ı ٥.

CLAUSE 23: UPGRADEABILITY

All future hardware upgrades and software updates involving <u>patient safety</u> and removing software viruses from existing software must be supplied at no additional cost.

ANY UPGRADE BEFORE OR AFTER INSTALLATION OF THE EQUIPMENT INVOLVING ADDITIONAL COST MUST BE BROUGHT TO THE ATTENTION OF THE MANAGER, HEALTH TECHNOLOGY SERVICES.

CLAUSE 24: MANUALS AND BROCHURES

The successful bidder must include in their offer at no extra cost to the final bid price:

- i) Complete user Operation / Maintenance Manual x 2 (two) Book / File; CD; DVD copies in English Language
- ii) Complete **ORIGINAL** Service / Repair Manual x 2 (two) Book / File; CD; DVD copies in English Language which MUST include the following information:

The offer submitted must be supported by descriptive literature, colour pamphlets, colour brochures and technical data sheets applicable to the offer. FAILURE TO SUBMIT THE ABOVE WILL RESULT IN THE BID NOT BEING CONSIDERED.

BIDDER'S COMMENTS:		

Clause 25

GUARANTEE/WARRANTY

The bidder must provide a warranty/ guarantee of minimum 24months period.

Clause 26

MAINTENANCE AND SERVICE AGREEMENT

Upon termination of the guarantee / warranty period the bidder must provide a fully - costed FULLY COMPREHENSIVE MAINTENANCE AND SERVICE AGREEMENT for a period of 5 years to commence upon termination of the guarantee / warranty period with an option to enter into a renewable agreement.

FULLY COMPREHENSIVE SERVICE AGREEMENT

- a) The bidder must state the number of services per annum that are required for the equipment offered as per the manufacturer's recommendations.
 b) The bidder must state the cost (inclusive of vat) of each service per unit.
 c) The bidder must include all costs (inclusive of vat) i.e. labour, travelling, mileage, spare parts, service kits, breakdowns and all call outs that is required for the servicing of each unit. (The bidder must attach on a separate annexure detailing the cost of each of the above.)
 d) The bidder must attach proof of the number of services as per the manufacturer's recommendation.

The bidder must submit a draft maintenance and service agreement with their bid.

f) The bidder must complete the schedule below.

e)

A - 1"- -"1-

Activity	Cost of Activity
Bidder:	_
Signature:	Date:

SCHEDULE OF ACCESSORIES

Bidders must quote the price of the accessories listed as well as any other accessories that may be useful to the end users.

The receiving Institutions may purchase individual accessories necessary for their particular Institution.

Cat No	Item	Price including VAT

SCHEDULE OF OPTIONAL ACCESSORIES

Bidders must quote the price of the optional accessories listed as well as any other accessories that may be useful to the end users. The receiving Institutions may purchase individual accessories necessary for their particular Institution.

Cat No	Item	Price including VAT

DETAILED TECHNICAL SPECIFICATION

GENERAL INFORMATION REQUIRED

FAILURE TO COMPLETE THIS PART WILL DISQUALIFY THE BIDDER

Make:				
Model Number / P	art Number for:			
Country of Origin				
Delivery Period				
R S A Import Perm	nit Holder (License	e No)		
Bidder				
Signature			Date	
Address				
Telephone No			_ Fax No	
Contact Person (Please Print)				

BILL OF QUANTITIES: SUPPLY, DELIVERY, DE-COMMISSIONING AND INSTALL 128 SLICE LARGE BORE COMPUTERISED TOMOGRAPHY SCANNER SYSTEM AT ADDINGTON HOSPITAL: ONCE-OFF

CONTRACT No. ZNB9044/2024-H

GENERIC ELECTRICAL/MECHANICAL/STRUCTURAL REQUIREMENTS FOR ALL EQUIPMENT INSTALLATION NOTES TO CONTRACTORS/TENDERERS:-

All items to be priced fully inclusive of all charges e.g labour, scaffolding, materials, profit, etc., but excluding Value Added Tax.

Contractors are warned that the institution is fully functional and occupied.

The Administration reserves the right to negotiate prices in the Bill of Quantities.

All rubble and redundant equipment shall be removed from the institution's property by the contactor as soon as is practically possible.

All equipment and materials used in this contract shall be that which is specified or other approved. Other approved shall mean prior to submission of tender.

The Contractor is advised to examine the drawings and to visit the site prior to tendering to acquaint him/herself with the nature of the work to be done and access to the siting of the existing buildings etc., as no claim will be allowed on the grounds of ignorance of the conditions under which the work will be executed.

Contractors are advised that the mechanical and electrical contractors will run concurrently with the structural contract and are requested to work in consultation with the appointed contractors in this regard.

All work shall be carried out according to the Department of Health's Standard Preambles to all Trade, the OHS Act,

	74
National Building Code of Practices and Regulations a Premises as well as the National Building Regulations. All quantities are provisional and shall be remeasured any invoices for work done by the Contractor.	

Item No	Description	Unit	Quantity	Rate	Total	
	PRELIMINARIES AND GENERAL					
1	Allow for Preliminaries and General items including the appointment of accredited of an accredited Health, Safety and Environment consultant to prepare safety and risk assessment plan including supervision thereof for duration of the contract	Item				
	TOTAL PRELIMINARIES AND GENERAL CARRIED TO SUMMARY PAGE					

1	ELECTRICAL / MECHANICAL WORKS Isolate, disconnect and remove existing light fittings and redundant X-Ray machine DB	Item		
2	Supply and install ILM/SFC/RCE/258 fluorescent fitting complete with electronic ballast and lamps or other approved for ceilings other than suspended ceilings OR see 3 below.	No		
3	Supply and install ILM/DNL/CNV/60W incandescent light fitting complete electronic ballast and lamps or other approved for ceilings where drop-in fitttings are required (suspended ceilings)	No		
4	Supply and install ILM/DNL/CNV/60W incandescent light fitting complete with globes or other approved. Distribution Boards supplied and installed in position	No		
	complete with all switchgear, bus-bar work, terminal wiring, lacing, labelling, warning notices, tested and commissioned as specified and shown on the drawings. At least 30% free space must be allowed for future expansion. The Project Leader shall approve design of all Distribution Boards prior to manufacture.			
5	Supply and install and connect new X-ray DB as per spec. DB shall be positioned behind the protective cubical in the most convenient position that will not impede the staff in any way.	No		
6	Supply and install labelling to conform to DOH Policy Document on the Design of Electrical Installations.	Item		
7	Supply and install Emergency Push Stop/Twist release button on cubicle wall at 1600mm A.F.F.L as indicated on drawing including wiring to new X-Ray DB	Item		
8	Supply and install PVC Trunking 100x100mm with covers. Allow for bends, elbows etc. Grey in colour.	m		
9	Supply and install medium duty (76.2x36mm) galvanised cable tray in ceiling space if required. Cable tray shall be bonded to X-Ray DB board using 4mm PVC earth wire and approved lugs suitably crimped.	m		
10	Supply and install 2 lever plus rotary type dimmer switch for incandescent bulkhead light fittings in X-Ray room including tubing and wiring between light fittings, switches and DB. Incandescent fitting above cubicle shall NOT be dimmable. All toggles on light switches shall be RED in colour.	No		
11	Supply and install Xpelair GX 9 Extractor fan and suitable and approved ducting to exterior of building or other approved for X-Ray room	No		
12	Supply and install a SANS approvedBTU mid wall reverse cycle air-conditioning unit complete with piping etc. in a position approved by the Project Leader.	No		
13	Supply and install 60A DP weatherproof Isolator for A/C unit including tubing and wiring from the existing DB. Allow for the supply and installation of a 30A SP MCB in the existing DB for the supply to the extract fan.	Item		
				

14	Supply and install 60A DP Isolator for extract fan including tubing and wiring from the existing DB. Allow for the supply and installation of a 20A SP MCB in the existing DB for the supply to the extract fan.	Item		
15	Supply and install inside cubicle area surface mounted above counter 1 x duo 16-amp switched socket outlet including tubing and wiring from existing DB. All toggles on switch socket outlets shall be RED in colour.	Item		
16	Replace all existing switched socket outlets with new. All toggles on switch socket outlets shall be RED in colour	No		
17	Replace all existing light switches with new. All toggles on light switches shall be RED in colour	No		
18	Supply and install new viewing box. Type HU-XRU-03	No		
19	Issue of Compliance Certificate in accordance with SANS 10142-1 Code of Practice	Item		
20	3 channel epoxy powder coated trunking above work tops.	m		
	TOTAL ELECTRICAL/MECHANICAL CARRIED TO SUMMARY	PAGE		

	BUILDING WORKS		 	
1	Remove existing floor covering	m2	 	
2	Remove existing xmm splayed cement skirting	m		
3	Remove timber framed lead lined cubicle +/- 1,8m high	Item		
4	Remove and brick up door opening in 230mm thick wall making good surfaces ready for painting.	m2		
5	Remove existing stainless steel sink and glazed wall tiles including cutting and sealing off all water supply and drainage.	Item		
6	Remove existing ceiling and supporting framework.	m2		
7	Remove existing machine floor mounted rail.	m		
8	Brick up opening typing new brickwork to existing with galvanised hoop iron ties every 5th course. Plaster and prepare for painting.	m2		
9	Cut openings into existing 230mm walls size 250 x 150mm and build in new 203mm x 133mm x 25kg Structural steel beams making good exposed and damaged brickwork. No 3 beams required.	m		
10	P1000X 2500mm long white powder coated Unitruts bolted in 3 places to underside of beams with 12mm HT bolts with Nylock nuts.	No		
11	OWAcoustic Brillianto A Premium biologically absorbable mineral wool ceiling tiles, NRC - 0.90, CAC - 30dB, Fire classification A2-s1, d0, weight - 3.5 kg/m², size 1200 x 600x 15mm with Square-edge and white fleece finish, laid on fire rated OWAconstruct S3 exposed demountable T24 suspension system, comprising galvanised main tees and cross tees with main tees suspended by means of galvanised hangers at centres not exceeding 1200mm, and all installed to manufacturer's instructions.	m2		
12	OWAconstruct Shadowline W-trim, plugged and screwed at centres not exceeding 200mm.	m		
13	Cut into existing floor for new cable ducting trench size: 350mm wide x 250mm deep x 1500mm long. Concrete in base and sides.	m2		
14	4mm Pavelite screed to floor.	m2		
15	Polyflor Palettone PUR 2mm thick x 2m wide fully flexible PVC sheet flooring with monolayer and homogeneous construction, non-directional design and PUR reinforcement on screeded floors installed as per the manufacturer's instructions or similar upon approval by DOH only. Colour to be determined on site.	m2		
16	Polyflor MC210C vinyl skirting including PC20 cove fillet	m		
17	Clean floor by washing using a neutral detergent with machine scrubbing, wet vacuumed and well rinsed, once dry apply two coats of a matt sealer Proflor HM (Mepol HM) as per the manufacturer's instructions.	m2		
18	Supply, fit and connect to existing services new "Vaal Sola 510" medical wash hand basin including suitable chrome basin waste and chrome bottle trap (No PVC)	No		
19	Elbow action Medical Mixer tap as "Cobra" 515/055H-21	No		
	7	Ω		

			1	_	
20	300 x300 matt white glazed tiles as splashback with PVC edge trim above basin.	m²			
21	One undercoat and two finishing coats washable acrylic PVA paint on walls with minimum 7 year guarantee including all necessary prepartory work to walls as stated by the manufacturer.	m²			
22	One undercoat and two finishing coats washable acrylic PVA paint on ceilings with minimum 7 year guarantee including all necessary prepartory work to ceilings as stated by the manufacturer.	m2			
23	Sand down existing doors and frames, prime, undercoat and two coats of non-drip enamel paint.	m2			
24	Supply and install new lead lined protective cubicle with counter and cassette storage rack. Fit lead glass 600 x 400 x 2.2mm thick	No			
25	Grade 304 stainless steel sliding door cladding 1200mm high x 1,6mm thick glued to door or with stainless steel counter sunk screws at 500mm c/c. Cladding returned on reveals.	m²			
26	Grade 304 stainless steel wall edge protectors 76mm x76mm x1200mm high x1,6mm thick glued to wall or with stainless steel counter sunk screws at 500mm c/c.	No			
27	Built in new lead lined speech grill sizexmm	No			
28	Supply and fit 1500 x 2200 18mm particle board with 2.2mm thick lead infill sheet including priming and painting with non-drip enamel paint	No			
29	Supply and fit new X-Ray entrance sliding door size 1830 x 2032 x 40mm. Complete with new heavy duty sliding door track, four door stoppers.2mm lead insert between panels. Door to cover 100mm over each side.	No			
30	18mm Melamine faced Bison Dura V313 cupboard with 32mm post formed formica worktop with intermediate shelf. Silicone sealant against wall. Unit size: xmm	No			
31	18MM Melamine faced Bison Dura V313cupboard with 32mm post formed formica worktop with intermediate shelf. Silicone sealant against wall. Unit size 1500 x 600	No			
32	Supply and fit apron hanger to fit 3 aprons.	No			
33	Structural support steel beam and support walls to be certified by registered engineer. Certificate of compliance to be provided.	Item			
34	Sola 510 Medical basin	No			
35	200 x200 matt white glazed tiles as splashback with PVC edge trim above basin.	m²			
36	Cut into existing 230mm wall and build in new window size 1245 x 1511 complete with sill, dpc and making good plaster.	No			
37	Elbow action Medical Mixer tap as "Cobra" 515-21	No			
38	New heavy duty chromed Hasp and Staple locking device 195 x 45x 3,5mm thick screwed in position. Heavy duty Padlock included.	No			
39	Redecorate existing pass through X-Ray hatch.	Item			
40	Sand down existing sliding door and apply two coats of polyurethane varnish.	m2			

41	relimet over sliding door gear 19mm thick x 300mm wide x 3,6mm long timber primed, under coat and two coats of non-drip enamel paint. Allow for brackets and fixing in position.	No		
42	Service existing sliding door gear and adjust where required.	Item		
43	600mm wide x 32mm thick moisture resistant V313 partical board post formed formica worktop with two, 500mm wide 3 drawer non-lockable units constructed out of 18mm thick moisture resistant V313 partical board melamine faced. Allowance to be made for all wall cleats, intermediate supports where required. Silicone sealant against vertical faces.	m		
44	1200mm x 1000mm Parrot pinning board or other approved (by DOH) fixed in position.	No		
45	2mm Lead sheeting per roll.	No		
46	Lead glass window 800 x 500 x 2,2mm	No		
47	50 x 50 x 2mm Aluminium angle plugged and screwed to aperture reveals for lead glass window	m		
48	Out of Bucky Cassete holder wall mounted bracket.	No		
49	Alterations for the relocation of medical gas outlet points and re-numbering accordingly.	No		
50	Sand and varnish slatted timber seats in change cubicle. Apply two coats of polyurethane varnish.	Item		
	TOTAL BUILDING WORKS CARRIED TO SUMMARY PAGE			

SBD 3.2

SECTION M PRICING SCHEDULE

Name of bidder	Bid number: ZNB 9044/2024-H
Closing Time: 11:00	Closing Date: 13/01/2025

OFFER TO BE VALID FOR 180 DAYS FROM THE CLOSING DATE OF BID

DESCRIPTION: BILL OF QUANTITIES: SUPPLY, DELIVERY, INSTALLATION, DEMONSTRATION AND COMMISSIONING 128 SLICE
LARGE BORE COMPLITERISED TOMOGRAPHY SCANNER SYSTEM AT ADDINGTON HOSPITAL ONCE-OFF

ARGE BORE COMPUTERISED TOMOGRAPHY SCANNER SYSTEM AT ADDINGTON HOSPITAL. ONCE-OFF.					
1) UNIT PRICE IN RSA CURRENCY:	R				
 (ALL APPLICABLE TAXES INCLUDED) (ANY STIPULATED COMPULSORY ACCESSORIES) 	AMOUNT IN WORDS				
➤ (INCLUSIVE OF 24 MONTHS WARANTY)					
>					
2) CARRIED OVER FROM MAINTENANCE AGREEMENT:	R				
> (5 YEAR WARRANTY WHICH TAKES EFFECT POST 24 MONTHS WARANTY)	AMOUNT IN WORDS				
(BIDDERS TO SUPPLY A BREAKDOWN OF THE FULLY COMPREHENSIVE MAINTENANCE SERVICE AGREEMENT.					
>					
3) CARRIED OVER FROM BILL OF QUANTITIES	R				
FINAL TOTAL:					
(ALL APPLICABLE TAXES INCLUDED	AMOUNT IN WORDS				
4) TOTAL BID PRICE (X1 UNITS) IN RSA CURRENCY:	R				
> (TOTAL BID PRICE = UNIT PRICE + MAINTENANCE AGREEMENT PRICE i.e. TOTAL OF 1& 2.)	AMOUNT IN WORDS				
> (ALL APPLICABLE TAXES INCLUDED)					

(Signature of Bidder)	 Date	(Signature of Witness)	 Date
Note: All delivery costs mus	t be included in the bid pr	ice, for delivery at prescribed destination	
Failure to comply with the	above shall invalidate t	he offer received.	
Delivery period (on order)			
Brand			
Country of origin			
-At:		HEALTH TECHNOLOGY SERVICES (HTS	S) ADDINGTON HOSPITAL
Required by:		KZN DEPARTMENT OF HEALTH	

ANNEXURE A

THE NATIONAL TREASURY

Republic of South Africa



GOVERNMENT PROCUREMENT: GENERAL CONDITIONS OF CONTRACT

July 2010

NOTES

The purpose of this document is to:

- (i) Draw special attention to certain general conditions applicable to government bids, contracts and orders; and
- (ii) To ensure that clients be familiar with regard to the rights and obligations of all parties involved in doing business with government.

In this document words in the singular also mean in the plural and vice versa and words in the masculine also mean in the feminine and neuter.

- The General Conditions of Contract will form part of all bid documents and may not be amended.
- Special Conditions of Contract (SCC) relevant to a specific bid, should be compiled separately for every bid (if (applicable) and will supplement the General Conditions of Contract. Whenever there is a conflict, the provisions in the SCC shall prevail.

NOTES

The purpose of this document is to:

- i) Draw special attention to certain general conditions applicable to government bids, contracts and orders; and
- ii) To ensure that clients be familiar with regards to the rights and obligations of all parties involved in doing business with government.

In this document words in the singular also mean in the plural and vice versa and words in the masculine also mean in the feminine and neuter.

- o The General Conditions of Contract will form part of all bid documents and may not be amended.
- Special Conditions of Contract (SCC) relevant to a specific bid, should be compiled separately for every bid (if (applicable)
 and will supplement the General Conditions of Contract. Whenever there is a conflict, the provisions in the SCC shall prevail.

84

INITIAL_

TABLE OF CLAUSES

1. Definitions 2. Application 3. General 4. Standards 5. Use of contract documents and information; inspection 6. Patent rights 7. Performance security 8. Inspections, tests and analysis 9. Packing 10. Delivery and documents 11. Insurance 12. Transportation 13. Incidental services 14. Spare parts 15. Warranty 16. Payment 17. Prices 18. Contract amendments 19. Assignment 20. Subcontracts 21. Delays in the supplier's performance 22. Penalties 23. Termination for default 24. Dumping and countervailing duties 25. Force Majeure 26. Termination for insolvency 27. Settlement of disputes

- 28. Limitation of liability
- 29. Governing language
- 30. Applicable law
- 31. Notices
- 32. Taxes and duties
- 33. National Industrial Participation Programme (NIPP)
- 34. Prohibition of restrictive practices

1. DEFINITIONS

The following terms shall be interpreted as indicated:

- 1.1. "Closing time" means the date and hour specified in the bidding documents for the receipt of bids.
- 1.2. **"Contract"** means the written agreement entered into between the purchaser and the supplier, as recorded in the contract form signed by the parties, including all attachments and appendices thereto and all documents incorporated by reference therein.
- 1.3. **"Contract price"** means the price payable to the supplier under the contract for the full and proper performance of his contractual obligations.
- 1.4. **"Corrupt practice"** means the offering, giving, receiving, or soliciting of anything of value to influence the action of a public official in the procurement process or in contract execution.
- 1.5. **"Countervailing duties"** are imposed in cases where an enterprise abroad is subsidized by its government and encouraged to market its products internationally.
- 1.6. **"Country of origin"** means the place where the goods were mined, grown or produced or from which the services are supplied. Goods are produced when, through manufacturing, processing or substantial and major assembly of components, a commercially recognized new product results that is substantially different in basic characteristics or in purpose or utility from its components.
- 1.7. "Day" means calendar day.
- 1.8. "Delivery" means delivery in compliance of the conditions of the contract or order.
- 1.9. "Delivery ex stock" means immediate delivery directly from stock actually on hand.
- 1.10. "Delivery into consignees store or to his site" means delivered and unloaded in the specified store or depot or on the specified site in compliance with the conditions of the contract or order, the supplier bearing all risks and charges involved until the supplies are so delivered and a valid receipt is obtained.
- 1.11. "Dumping" occurs when a private enterprise abroad market its goods on own initiative in the RSA at lower prices than that of the country of origin and which have the potential to harm the local industries in the RSA.
- 1.12. "Force majeure" means an event beyond the control of the supplier and not involving the supplier's fault or negligence and not foreseeable. Such events may include, but is not restricted to, acts of the purchaser in its sovereign capacity, wars or revolutions, fires, floods, epidemics, guarantine restrictions and freight embargoes.
- 1.13. "Fraudulent practice" means a misrepresentation of facts in order to influence a procurement process or the execution of a contract to the detriment of any bidder and includes collusive practice among bidders (prior to or after bid submission) designed to establish bid prices at artificial non-competitive levels and to deprive the bidder of the benefits of free and open competition.
- 1.14. "GCC" means the General Conditions of Contract.
- 1.15. "Goods" means all of the equipment, machinery, and/or other materials that the supplier is required to supply to the purchaser under the contract.
- 1.16. "Imported content" means that portion of the bidding price represented by the cost of components, parts or materials which have been or are still to be imported (whether by the supplier or his subcontractors) and which costs are inclusive of the costs abroad, plus freight and other direct importation costs such as landing costs, dock dues, import duty, sales duty or other similar tax or duty at the South African place of entry as well as transportation and handling charges to the factory in the Republic where the supplies covered by the bid will be manufactured.
- 1.17. "Local content" means that portion of the bidding price which is not included in the imported content provided that local manufacture does take place.
- 1.18. "Manufacture" means the production of products in a factory using labour, materials, components and machinery and includes other related value-adding activities.
- 1.19. "Order" means an official written order issued for the supply of goods or works or the rendering of a service.
- 1.20. "Project site," where applicable, means the place indicated in bidding documents.
- 1.21. "Purchaser" means the organization purchasing the goods.
- 1.22. "Republic" means the Republic of South Africa.
- 1.23. "SCC" means the Special Conditions of Contract.

- 1.24. "Services" means those functional services ancillary to the supply of the goods, such as transportation and any other incidental services, such as installation, commissioning, provision of technical assistance, training, catering, gardening, security, maintenance and other such obligations of the supplier covered under the contract.
- 1.25. "Written" or "in writing" means handwritten in ink or any form of electronic or mechanical writing.

2. Application

- 2.1. These general conditions are applicable to all bids, contracts and orders including bids for functional and professional services, sales, hiring, letting and the granting or acquiring of rights, but excluding immovable property, unless otherwise indicated in the bidding documents.
- 2.2. Where applicable, special conditions of contract are also laid down to cover specific supplies, services or works.
- 2.3. Where such special conditions of contract are in conflict with these general conditions, the special conditions shall apply.

3. General

- 3.1. Unless otherwise indicated in the bidding documents, the purchaser shall not be liable for any expense incurred in the preparation and submission of a bid. Where applicable a non-refundable fee for documents may be charged.
- 3.2. With certain exceptions, invitations to bid are only published in the Government Tender Bulletin. The Government Tender Bulletin may be obtained directly from the Government Printer, Private Bag X85, Pretoria 0001, or accessed electronically from www.treasury.gov.za.

4. Standards

4.1. The goods supplied shall conform to the standards mentioned in the bidding documents and specifications.

5. Use of contract documents and information; inspection.

- 5.1. The supplier shall not, without the purchaser's prior written consent, disclose the contract, or any provision thereof, or any specification, plan, drawing, pattern, sample, or information furnished by or on behalf of the purchaser in connection therewith, to any person other than a person employed by the supplier in the performance of the contract. Disclosure to any such employed person shall be made in confidence and shall extend only so far as may be necessary for purposes of such performance.
- 5.2. The supplier shall not, without the purchaser's prior written consent, make use of any document or information mentioned in GCC clause except for purposes of performing the contract.
- 5.3. Any document, other than the contract itself mentioned in GCC clause shall remain the property of the purchaser and shall be returned (all copies) to the purchaser on completion of the supplier's performance under the contract if so required by the purchaser.
- 5.4. The supplier shall permit the purchaser to inspect the supplier's records relating to the performance of the supplier and to have them audited by auditors appointed by the purchaser, if so required by the purchaser.

6. Patent rights

6.1. The supplier shall indemnify the purchaser against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the goods or any part thereof by the purchaser.

7. Performance security

- 7.1. Within thirty (30) days of receipt of the notification of contract award, the successful bidder shall furnish to the purchaser the performance security of the amount specified in SCC.
- 7.2. The proceeds of the performance security shall be payable to the purchaser as compensation for any loss resulting from the supplier's failure to complete his obligations under the contract.
- 7.3. The performance security shall be denominated in the currency of the contract, or in a freely convertible currency acceptable to the purchaser and shall be in one of the following forms:
 - a) a bank guarantee or an irrevocable letter of credit issued by a reputable bank located in the purchaser's country or abroad, acceptable to the purchaser, in the form provided in the bidding documents or another form acceptable to the purchaser; or
 - b) a cashier's or certified cheque
- 7.4. The performance security will be discharged by the purchaser and returned to the supplier not later than thirty (30) days following the date of completion of the supplier's performance obligations under the contract, including any warranty obligations, unless otherwise specified in SCC.

8. Inspections, tests, and analyses

8.1. All pre-bidding testing will be for the account of the bidder.

- 8.2. If it is a bid condition that supplies to be produced or services to be rendered should at any stage during production or execution or on completion be subject to inspection, the premises of the bidder or contractor shall be open, at all reasonable hours, for inspection by a representative of the Department or an organization acting on behalf of the Department.
- 8.3. If there are no inspection requirements indicated in the bidding documents and no mention is made in the contract, but during the contract period it is decided that inspections shall be carried out, the purchaser shall itself make the necessary arrangements, including payment arrangements with the testing authority concerned.
- 8.4. If the inspections, tests and analyses referred to in clauses 8.2 and 8.3 show the supplies to be in accordance with the contract requirements, the cost of the inspections, tests and analyses shall be defrayed by the purchaser.
- 8.5. Where the supplies or services referred to in clauses 8.2 and 8.3 do not comply with the contract requirements, irrespective of whether such supplies or services are accepted or not, the cost in connection with these inspections, tests or analyses shall be defrayed by the supplier.
- 8.6. Supplies and services which are referred to in clauses 8.2 and 8.3 and which do not comply with the contract requirements may be rejected.
- 8.7. Any contract supplies may on or after delivery be inspected, tested or analysed and may be rejected if found not to comply with the requirements of the contract. Such rejected supplies shall be held at the cost and risk of the supplier who shall, when called upon, remove them immediately at his own cost and forthwith substitute them with supplies which do comply with the requirements of the contract. Failing such removal the rejected supplies shall be returned at the suppliers cost and risk. Should the supplier fail to provide the substitute supplies forthwith, the purchaser may, without giving the supplier further opportunity to substitute the rejected supplies, purchase such supplies as may be necessary at the expense of the supplier.
- 8.8. The provisions of clauses 8.4 to 8.7 shall not prejudice the right of the purchaser to cancel the contract on account of a breach of the conditions thereof, or to act in terms of Clause 23 of GCC.

9. Packing

- 9.1. The supplier shall provide such packing of the goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in the contract. The packing shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packing, case size and weights shall take into consideration, where appropriate, the remoteness of the goods' final destination and the absence of heavy handling facilities at all points in transit.
- 9.2. The packing, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the contract, including additional requirements, if any, specified in SCC, and in any subsequent instructions ordered by the purchaser.

10. Delivery and documents

- 10.1. Delivery of the goods shall be made by the supplier in accordance with the terms specified in the contract. The details of shipping and/or other documents to be furnished by the supplier are specified in SCC.
- 10.2. Documents to be submitted by the supplier are specified in SCC.

11. Insurance

11.1. The goods supplied under the contract shall be fully insured in a freely convertible currency against loss or damage incidental to manufacture or acquisition, transportation, storage and delivery in the manner specified in the SCC.

12. Transportation

12.1. Should a price other than an all-inclusive delivered price be required, this shall be specified in the SCC.

13. Incidental services

- 13.1. The supplier may be required to provide any or all of the following services, including additional services, if any, specified in SCC:
 - a) performance or supervision of on-site assembly and/or commissioning of the supplied goods;
 - b) furnishing of tools required for assembly and/or maintenance of the supplied goods;
 - c) furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied goods;
 - d) performance or supervision or maintenance and/or repair of the supplied goods, for a period of time agreed by the parties, provided that this service shall not relieve the supplier of any warranty obligations under this contract; and
 - e) training of the purchaser's personnel, at the supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied goods.

13.2. Prices charged by the supplier for incidental services, if not included in the contract price for the goods, shall be agreed upon in advance by the parties and shall not exceed the prevailing rates charged to other parties by the supplier for similar services.

14. Spare parts

- 14.1. As specified in SCC, the supplier may be required to provide any or all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the supplier:
 - a) such spare parts as the purchaser may elect to purchase from the supplier, provided that this election shall not relieve the supplier of any warranty obligations under the contract; and
 - b) in the event of termination of production of the spare parts:
 - i) Advance notification to the purchaser of the pending termination, in sufficient time to permit the purchaser to procure needed requirements; and
 - ii) following such termination, furnishing at no cost to the purchaser, the blueprints, drawings, and specifications of the spare parts, if requested.

15. Warranty

- 15.1. The supplier warrants that the goods supplied under the contract are new, unused, of the most recent or current models, and that they incorporate all recent improvements in design and materials unless provided otherwise in the contract. The supplier further warrants that all goods supplied under this contract shall have no defect, arising from design, materials, or workmanship (except when the design and/or material is required by the purchaser's specifications) or from any act or omission of the supplier, that may develop under normal use of the supplied goods in the conditions prevailing in the country of final destination.
- 15.2. This warranty shall remain valid for twelve (12) months after the goods, or any portion thereof as the case may be, have been delivered to and accepted at the final destination indicated in the contract, or for eighteen (18) months after the date of shipment from the port or place of loading in the source country, whichever period concludes earlier, unless specified otherwise in SCC.
- 15.3. The purchaser shall promptly notify the supplier in writing of any claims arising under this warranty.
- 15.4. Upon receipt of such notice, the supplier shall, within the period specified in SCC and with all reasonable speed, repair or replace the defective goods or parts thereof, without costs to the purchaser.
- 15.5. If the supplier, having been notified, fails to remedy the defect(s) within the period specified in SCC, the purchaser may proceed to take such remedial action as may be necessary, at the supplier's risk and expense and without prejudice to any other rights which the purchaser may have against the supplier under the contract.

16. Payment

- 16.1. The method and conditions of payment to be made to the supplier under this contract shall be specified in SCC.
- 16.2. The supplier shall furnish the purchaser with an invoice accompanied by a copy of the delivery note and upon fulfilment of other obligations stipulated in the contract.
- 16.3. Payments shall be made promptly by the purchaser, but in no case later than thirty (30) days after submission of an invoice or claim by the supplier.
- 16.4. Payment will be made in Rand unless otherwise stipulated in SCC.

17. Prices

17.1. Prices charged by the supplier for goods delivered and services performed under the contract shall not vary from the prices quoted by the supplier in his bid, with the exception of any price adjustments authorized in SCC or in the purchaser's request for bid validity extension, as the case may be.

18. Contract amendments

18.1. No variation in or modification of the terms of the contract shall be made except by written amendment signed by the parties concerned.

19. Assignment

19.1. The supplier shall not assign, in whole or in part, its obligations to perform under the contract, except with the purchaser's prior written consent.

20. Subcontracts

20.1. The supplier shall notify the purchaser in writing of all subcontracts awarded under this contract if not already specified in the bid. Such notification, in the original bid or later, shall not relieve the supplier from any liability or obligation under the contract.

21. Delays in the supplier's performance

- 21.1. Delivery of the goods and performance of services shall be made by the supplier in accordance with the time schedule prescribed by the purchaser in the contract.
- 21.2. If at any time during performance of the contract, the supplier or its subcontractor(s) should encounter conditions impeding timely delivery of the goods and performance of services, the supplier shall promptly notify the purchaser in writing of the fact of the delay, its likely duration and its cause(s). As soon as practicable after receipt of the supplier's notice, the purchaser shall evaluate the situation and may at his discretion extend the supplier's time for performance, with or without the imposition of penalties, in which case the extension shall be ratified by the parties by amendment of contract.
- 21.3. No provision in a contract shall be deemed to prohibit the obtaining of supplies or services from a national department, provincial department, or a local authority.
- 21.4. The right is reserved to procure outside of the contract small quantities or to have minor essential services executed if an emergency arises, the supplier's point of supply is not situated at or near the place where the supplies are required, or the supplier's services are not readily available.
- 21.5. Except as provided under GCC Clause 25, a delay by the supplier in the performance of its delivery obligations shall render the supplier liable to the imposition of penalties, pursuant to GCC Clause 22, unless an extension of time is agreed upon pursuant to GCC Clause 21.2 without the application of penalties.
- 21.6. Upon any delay beyond the delivery period in the case of a supplies contract, the purchaser shall, without canceling the contract, be entitled to purchase supplies of a similar quality and up to the same quantity in substitution of the goods not supplied in conformity with the contract and to return any goods delivered later at the supplier's expense and risk, or to cancel the contract and buy such goods as may be required to complete the contract and without prejudice to his other rights, be entitled to claim damages from the supplier.

22. Penalties

22.1. Subject to GCC Clause 25, if the supplier fails to deliver any or all of the goods or to perform the services within the period(s) specified in the contract, the purchaser shall, without prejudice to its other remedies under the contract, deduct from the contract price, as a penalty, a sum calculated on the delivered price of the delayed goods or unperformed services using the current prime interest rate calculated for each day of the delay until actual delivery or performance. The purchaser may also consider termination of the contract pursuant to GCC Clause 23.

23. Termination for default

- 23.1. The purchaser, without prejudice to any other remedy for breach of contract, by written notice of default sent to the supplier, may terminate this contract in whole or in part:
 - a) if the supplier fails to deliver any or all of the goods within the period(s) specified in the contract, or within any extension thereof granted by the purchaser pursuant to GCC Clause 21.2;
 - b) if the Supplier fails to perform any other obligation(s) under the contract; or
 - c) if the supplier, in the judgment of the purchaser, has engaged in corrupt or fraudulent practices in competing for or in executing the contract.
- 23.2. In the event the purchaser terminates the contract in whole or in part, the purchaser may procure, upon such terms and in such manner as it deems appropriate, goods, works or services similar to those undelivered, and the supplier shall be liable to the purchaser for any excess costs for such similar goods, works or services. However, the supplier shall continue performance of the contract to the extent not terminated.
- 23.3. Where the purchaser terminates the contract in whole or in part, the purchaser may decide to impose a restriction penalty on the supplier by prohibiting such supplier from doing business with the public sector for a period not exceeding 10 years.
- 23.4. If a purchaser intends imposing a restriction on a supplier or any person associated with the supplier, the supplier will be allowed a time period of not more than fourteen (14) days to provide reasons why the envisaged restriction should not be imposed. Should the supplier fail to respond within the stipulated fourteen (14) days the purchaser may regard the intended penalty as not objected against and may impose it on the supplier.
- 23.5. Any restriction imposed on any person by the Accounting Officer / Authority will, at the discretion of the Accounting Officer / Authority, also be applicable to any other enterprise or any partner, manager, director or other person who wholly or partly

exercises or exercised or may exercise control over the enterprise of the first-mentioned person, and with which enterprise or person the first-mentioned person, is or was in the opinion of the Accounting Officer / Authority actively associated.

- 23.6. If a restriction is imposed, the purchaser must, within five (5) working days of such imposition, furnish the National Treasury, with the following information:
 - i) the name and address of the supplier and / or person restricted by the purchaser;
 - ii) the date of commencement of the restriction
 - iii) the period of restriction; and
 - iv) the reasons for the restriction.

These details will be loaded in the National Treasury's central database of suppliers or persons prohibited from doing business with the public sector.

23.7. If a court of law convicts a person of an offence as contemplated in sections 12 or 13 of the Prevention and Combating of Corrupt Activities Act, No. 12 of 2004, the court may also rule that such person's name be endorsed on the Register for Tender Defaulters. When a person's name has been endorsed on the Register, the person will be prohibited from doing business with the public sector for a period not less than five years and not more than 10 years. The National Treasury is empowered to determine the period of restriction and each case will be dealt with on its own merits. According to section 32 of the Act the Register must be open to the public. The Register can be perused on the National Treasury website.

24. Anti-dumping and countervailing duties and rights

24.1. When, after the date of bid, provisional payments are required, or antidumping or countervailing duties are imposed, or the amount of a provisional payment or anti-dumping or countervailing right is increased in respect of any dumped or subsidized import, the State is not liable for any amount so required or imposed, or for the amount of any such increase. When, after the said date, such a provisional payment is no longer required or any such anti-dumping or countervailing right is abolished, or where the amount of such provisional payment or any such right is reduced, any such favourable difference shall on demand be paid forthwith by the contractor to the State or the State may deduct such amounts from moneys (if any) which may otherwise be due to the contractor in regard to supplies or services which he delivered or rendered, or is to deliver or render in terms of the contract or any other contract or any other amount which may be due to him

25. Force Majeure

- 25.1. Notwithstanding the provisions of GCC Clauses 22 and 23, the supplier shall not be liable for forfeiture of its performance security, damages, or termination for default if and to the extent that his delay in performance or other failure to perform his obligations under the contract is the result of an event of force majeure.
- 25.2. If a force majeure situation arises, the supplier shall promptly notify the purchaser in writing of such condition and the cause thereof. Unless otherwise directed by the purchaser in writing, the supplier shall continue to perform its obligations under the contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the force majeure event.

26. Termination for insolvency

26.1. The purchaser may at any time terminate the contract by giving written notice to the supplier if the supplier becomes bankrupt or otherwise insolvent. In this event, termination will be without compensation to the supplier, provided that such termination will not prejudice or affect any right of action or remedy which has accrued or will accrue thereafter to the purchaser.

27. Settlement of Disputes

- 27.1. If any dispute or difference of any kind whatsoever arises between the purchaser and the supplier in connection with or arising out of the contract, the parties shall make every effort to resolve amicably such dispute or difference by mutual consultation.
- 27.2. If, after thirty (30) days, the parties have failed to resolve their dispute or difference by such mutual consultation, then either the purchaser or the supplier may give notice to the other party of his intention to commence with mediation. No mediation in respect of this matter may be commenced unless such notice is given to the other party.
- 27.3. Should it not be possible to settle a dispute by means of mediation, it may be settled in a South African court of law.
- 27.4. Mediation proceedings shall be conducted in accordance with the rules of procedure specified in the SCC.
- 27.5. Notwithstanding any reference to mediation and/or court proceedings herein,

- a) the parties shall continue to perform their respective obligations under the contract unless they otherwise agree; and
- b) the purchaser shall pay the supplier any monies due the supplier.

28. Limitation of liability

- 28.1. Except in cases of criminal negligence or wilful misconduct, and in the case of infringement pursuant to Clause 6.:
 - a) the supplier shall not be liable to the purchaser, whether in contract, tort, or otherwise, for any indirect or consequential loss or damage, loss of use, loss of production, or loss of profits or interest costs, provided that this exclusion shall not apply to any obligation of the supplier to pay penalties and/or damages to the purchaser; and
 - b) the aggregate liability of the supplier to the purchaser, whether under the contract, in tort or otherwise, shall not exceed the total contract price, provided that this limitation shall not apply to the cost of repairing or replacing defective equipment.

29. Governing language

29.1. The contract shall be written in English. All correspondence and other documents pertaining to the contract that is exchanged by the parties shall also be written in English.

30. Applicable law

30.1. The contract shall be interpreted in accordance with South African laws, unless otherwise specified in SCC.

31. Notices

- 31.1. Every written acceptance of a bid shall be posted to the supplier concerned by registered or certified mail and any other notice to him shall be posted by ordinary mail to the address furnished in his bid or to the address notified later by him in writing and such posting shall be deemed to be proper service of such notice
- 31.2. The time mentioned in the contract documents for performing any act after such aforesaid notice has been given, shall be reckoned from the date of posting of such notice.

32. Taxes and duties

- 32.1. A foreign supplier shall be entirely responsible for all taxes, stamp duties, license fees, and other such levies imposed outside the purchaser's country.
- 32.2. A local supplier shall be entirely responsible for all taxes, duties, license fees, etc., incurred until delivery of the contracted goods to the purchaser.
- 32.3. No contract shall be concluded with any bidder whose tax matters are not in order. Prior to the award of a bid the Department must be in possession of a tax clearance certificate, submitted by the bidder. This certificate must be an original issued by the South African Revenue Services.

33. National Industrial Participation (NIP) Programme

33.1. The NIP Programme administered by the Department of Trade and Industry shall be applicable to all contracts that are subject to the NIP obligation.

34. Prohibition of Restrictive practices

- 34.1. In terms of section 4 (1) (b) (iii) of the Competition Act No. 89 of 1998, as amended, an agreement between, or concerted practice by, firms, or a decision by an association of firms, is prohibited if it is between parties in a horizontal relationship and if a bidder (s) is / are or a contractor(s) was / were involved in collusive bidding (or bid rigging).
- 34.2. If a bidder(s) or contractor(s), based on reasonable grounds or evidence obtained by the purchaser, has / have engaged in the restrictive practice referred to above, the purchaser may refer the matter to the Competition Commission for investigation and possible imposition of administrative penalties as contemplated in the Competition Act No. 89 of 1998.
- 34.3. If a bidder(s) or contractor(s), has / have been found guilty by the Competition Commission of the restrictive practice referred to above, the purchaser may, in addition and without prejudice to any other remedy provided for, invalidate the bid(s) for such item(s) offered, and / or terminate the contract in whole or part, and / or restrict the bidder(s) or contractor(s) from conducting business with the public sector for a period not exceeding ten (10) years and / or claim damages from the bidder(s) or contractor(s) concerned.

General Conditions of Contract (revised July 2010)

93